



# AEMC INSTRUMENTS JM852 AC Current Probe User Manual

[Home](#) » [AEMC INSTRUMENTS](#) » AEMC INSTRUMENTS JM852 AC Current Probe User Manual 

**AEMC INSTRUMENTS JM852 AC Current Probe**



## Contents

- 1 DESCRIPTION
- 2 INTERNATIONAL ELECTRICAL SYMBOLS
- 3 DEFINITION OF MEASUREMENT CATEGORIES
- 4 RECEIVING YOUR SHIPMENT
- 5 ELECTRICAL SPECIFICATIONS
- 6 MECHANICAL SPECIFICATIONS
- 7 SAFETY SPECIFICATIONS
- 8 ORDERING INFORMATION
- 9 OPERATION
- 10 MAINTENANCE:
- 11 REPAIR AND CALIBRATION
- 12 TECHNICAL AND SALES ASSISTANCE
- 13 LIMITED WARRANTY
- 14 Documents / Resources
  - 14.1 References
- 15 Related Posts

## DESCRIPTION

The JM852 (Catalog #2113.41) is designed for use in industrial environments. The “squared” jaws permit multiple conductor or bus bar positioning. The voltage output makes it the perfect tool for measurement with DMMs, recorders, power and harmonic meters. The Model JM852 is compatible with any voltmeter, multimeter, or other voltage measuring instrument with an input impedance greater than 100K  $\Omega$ . To achieve the stated accuracy, use the JM852 with the DMM having an accuracy of 0.75%.

## WARNING

These safety warnings are provided to ensure the safety of personnel and proper operation of the instrument.

- Read the instruction manual completely and follow all the safety information before attempting to use or service this instrument.
- Use caution on any circuit: Potentially high voltages and currents may be present and may pose a shock hazard.
- Read the Safety Specifications section prior to using the current probe. Never exceed the maximum voltage ratings given.
- Safety is the responsibility of the operator.
- ALWAYS connect the current probe to the display device before clamping the probe onto the sample being tested.
- ALWAYS inspect the instrument, probe, probe cable, and output terminals prior to use. Replace any defective parts immediately.
- NEVER use the current probe on electrical conductors rated above 600V in overvoltage category III (CAT III). Use extreme caution when clamping around bare conductors or bus bars.

## INTERNATIONAL ELECTRICAL SYMBOLS



This symbol signifies that the current probe is protected by double or reinforced insulation. Use only factory specified replacement parts when servicing the instrument.



This symbol signifies CAUTION! and requests that the user refer to the user manual before using the instrument.



This is a type A current sensor. This symbol signifies that application around and removal from HAZARDOUS LIVE conductors is permitted.

## DEFINITION OF MEASUREMENT CATEGORIES

**Cat. I:** For measurements on circuits not directly connected to the AC supply wall outlet such as protected secondaries, signal level, and limited energy circuits.


**Cat. II:** For measurements performed on circuits directly connected to the electrical distribution system. Examples are measurements on household appliances or portable tools.

**Cat. III:** For measurements performed in the building installation at the distribution level such as on hardwired equipment in fixed installation and circuit breakers.

**Cat. IV:** For measurements performed at the primary electrical supply (<1000V) such as on primary overcurrent protection devices, ripple control units, or meters.

## RECEIVING YOUR SHIPMENT

Upon receiving your shipment, make sure that the contents are consistent with the packing list. Notify your distributor of any missing items. If the equipment appears to be damaged, file a claim immediately with the carrier and notify your distributor at once, giving a detailed description of any damage.



## ELECTRICAL SPECIFICATIONS

### Current Range:

1 to 2400A AC, continuous cycle for the full temperature range  
(3000A if temperature < 35°C or 95°F)

### Transformation Ratio:

3000:1

### Output Signal:

0.333mV AC/A AC (1V at 3000A)

### Accuracy and Phase Shift\*:

| Primary current | 150A | 600A  | 3000A |
|-----------------|------|-------|-------|
| Accuracy %      | 1.5% | 0.75% | 0.5%  |
| Phase shift     | 1.5° | 0.75° | 0.5°  |

(\*Reference conditions: 23°C±3°K, 20 to 85% RH, 48 to 65 Hz, external magnetic field < 40 A/m, no DC component, no external current carrying conductor, test sample centered.) Load impedance 1M Ω.

**Overload:** 3600A during 5 mn

**Accuracy:** Per IEC 185-26-27

Class 0.5 from 48 to 1000 Hz

**Frequency Range:** 30 Hz to 5 kHz; current derating above 1 kHz for continuous use

Ampere Second Product: 90 AA

**Load Impedance:** ≥ 100K Ω

**Working Voltage:** 600V AC

**Common Mode Voltage:** 600V AC

**Influence of Adjacent Conductor:** 0.005A/A AC

**Influence of Conductor in Jaw Opening:** 1% ± 0.1A of reading

## MECHANICAL SPECIFICATIONS

**Operating Temperature:** 14° to 122°F (-10° to 50°C)

**Storage Temperature:** -13° to 176°F (-25° to 80°C)

**Influence of Temperature:** < 0.2% per 10°K

**Jaw Opening:** 3.54" (90mm)

**Maximum Conductor Size:** Cable: 2.52" max. (64mm)

**Bus bar:** 1.97 x 5.31" (50 x 135mm) 2.52 x 3.94" (64 x 100mm)

**Envelope Protection:** IP 20 (IEC 529)

**Drop Test:** 500mm (IEC 68-2-32)

**Mechanical Shock:** 100 g (IEC 68-2-27)

**Vibration:** 10/55/10Hz, 0.15 mm (IEC 68-2-6)

**Polycarbonate Material:** Handles: 10% fiberglass charged polycarbonate UL 94 V0

**Jaws:** ABS UL 94 V2

**Dimensions:** 4.72 x 12.40 x 1.89" (120 x 315 x 48 mm)

**Weight:** 2.65 lbs. (1200 g)

**Colors:** Dark gray handles with red jaws

**Output:** 5 ft (1.5m) lead with safety 4mm banana plug

## SAFETY SPECIFICATIONS

### Electrical:

Double insulation or reinforced insulation between primary or secondary and outer case of handle upon IEC 1010-2-32



**Common Mode Voltage:** 600V CAT III, Pollution: 2

## ORDERING INFORMATION

**Current Probe JM852 ..... Cat #2113.41**

### Accessories:

Banana plug adapter (to non-recessed plug)..... **Cat #1017.45**

## OPERATION

Please make sure that you have already read and fully understand the WARNING section on page 1.

## Making Measurements with the AC Current Probe Model JM852

- Connect the black and red terminals to the Voltage AC range of your DMM or voltage measuring instrument. Select the appropriate voltage range (2V AC). Clamp the probe around the conductor to be tested. If the reading is less than 200 mV, select the lower range until you obtain the best resolution. Read the value display on the DMM and multiply it by the probe ratio ( $3000/1 = 3000$ ). (If reading = 0.592V AC, the current flowing through the probe is  $0.592 \times 3000 = 1776\text{A}$ ).
- For best accuracy: carefully center the conductor inside the probe jaw, avoid if possible, the proximity of other conductors which may create noise.

## Tips for Making Precise Measurements

- When using a current probe with a meter, it is important to select the range that provides the best resolution. Failure to do this may result in measurement errors.
- Make sure that probe jaw mating surfaces are free of dust and contamination. Contaminants cause air gaps between the jaws, increasing the phase shift between primary and secondary. It is very critical for power measurement.

## MAINTENANCE:



### Warning

- For maintenance use only original replacement parts.
- To avoid electrical shock, do not attempt to perform any servicing unless you are qualified to do so.
- To avoid electrical shock and/or damage to the instrument, do not get water or other foreign agents into the probe.

## Cleaning

To ensure optimum performance, it is important to keep the probe jaw mating surfaces clean at all times. Failure to do so may result in error in readings. To clean the probe jaws, use a very fine sand paper (fine 600) to avoid scratching the jaw, then gently clean with a soft oiled cloth.

## REPAIR AND CALIBRATION

You must contact our Service Center for a Customer Service Authorization number (CSA#). This will ensure that when your instrument arrives, it will be tracked and processed promptly. Please write the CSA# on the outside of the shipping container.

Chauvin Arnoux®, Inc. d.b.a. AEMC® Instruments

15 Faraday Drive – Dover, NH 03820 USA

(800) 945-2362 (Ext. 360) -(603) 749-6434 (Ext. 360) • [repair@aemc.com](mailto:repair@aemc.com)

(Or contact your authorized distributor)

**NOTE:** All customers must obtain a CSA# before returning any instrument.

## TECHNICAL AND SALES ASSISTANCE

If you are experiencing any technical problems, or require any assistance with the proper use or application of this

instrument, please contact our technical hotline: (800) 343-1391 • (508) 698-2115 • [techsupport@aemc.com](mailto:techsupport@aemc.com)

## LIMITED WARRANTY


The current probe is warranted to the owner for a period of one year from the date of original purchase against defects in manufacture. This limited warranty is given by AEMC® Instruments, not by the distributor from whom it was purchased. This warranty is void if the unit has been tampered with, abused or if the defect is related to service not performed by AEMC® Instruments.

Full warranty coverage and product registration is available on our website at: [www.aemc.com/warranty.html](http://www.aemc.com/warranty.html). Please print the online Warranty Coverage Information for your records.





---

## Documents / Resources

|   |  |
|---|--|
|  | <a href="#">AEMC INSTRUMENTS JM852 AC Current Probe</a> [pdf] User Manual<br>JM852, JM852 AC Current Probe, AC Current Probe, Current Probe, Probe |
|---|--|

## References

-  [AEMC Warranty Registration](#)
-  [Manual-Hub.com - Free PDF manuals!](#)
- [User Manual](#)

[Manuals+](#), [Privacy Policy](#)

This website is an independent publication and is neither affiliated with nor endorsed by any of the trademark owners. The "Bluetooth®" word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. The "Wi-Fi®" word mark and logos are registered trademarks owned by the Wi-Fi Alliance. Any use of these marks on this website does not imply any affiliation with or endorsement.