

## Ameda EVO17080UK

# Ameda Lactaline Personal Double Electric Breastpump Instruction Manual

Model: EVO17080UK

## INTRODUCTION

---

The Ameda Lactaline Personal Double Electric Breastpump is engineered for efficient and comfortable milk expression. This electric pump features a dual collection system, offering convenience for mothers. It is designed for personal use to support breastfeeding by enabling the collection of breast milk.

## Product Overview



Figure 1: Assembled Ameda Lactaline Personal Double Electric Breastpump with bottles and collection kits.

This image displays the complete breast pump unit, including the motor base, two milk collection kits, and two milk storage bottles. The design emphasizes portability and ease of use for daily milk expression.

## Box CONTENTS

Upon opening the packaging, please verify that all components listed below are present and in good condition.



Figure 2: Product packaging displaying the list of included items.

The packaging provides a clear overview of what is included with your breast pump.

- 1 x Lactaline Personal Breast Pump (Motor Unit)
- 1 x Dual HygieniKit Milk Collection System
- 2 x 4 oz. Bottles with Locking Caps
- 1 x AC Power Adaptor
- 1 x Instruction Manual

## SETUP AND ASSEMBLY

Before first use, ensure all parts that come into contact with breast milk are cleaned and sterilized according to the 'Care

and Maintenance' section. Assemble the pump as follows:



Figure 3: Key components of the milk collection kit, including valves and caps.

This image shows the essential parts of the collection kit that require assembly.

- 1. Prepare Collection Kit:** Attach the valve to the breast flange. Ensure it is securely seated.
- 2. Attach Bottle:** Screw the milk storage bottle onto the bottom of the breast flange assembly.
- 3. Connect Tubing:** Insert one end of the silicone tubing into the diaphragm cap and the other end into the pump's tubing port. For dual pumping, repeat for the second collection kit.
- 4. Power Connection:** Connect the AC power adapter to the pump unit and plug it into a wall outlet. Alternatively, the pump can be powered by 6 AA batteries (not included) for portable use.

## OPERATING INSTRUCTIONS

The Ameda Lactaline Personal Double Electric Breastpump offers adjustable settings for comfortable and effective milk expression.

### Pumping Procedure

- 1. Positioning:** Sit comfortably and place the breast flange centrally over your nipple, ensuring a good seal around the breast.
- 2. Power On:** Press the power button on the pump unit to turn it on.
- 3. Adjust Suction and Cycle Speed:** Use the control dials on the pump to adjust the suction level and cycle speed to your comfort. Begin with a gentle suction and gradually increase until milk flow is established and comfortable.
- 4. Pumping Duration:** Pump each breast for approximately 10-15 minutes, or until milk flow significantly slows. For dual pumping, both breasts can be pumped simultaneously.
- 5. Power Off:** Once finished, turn off the pump and carefully remove the breast flanges from your breasts.
- 6. Store Milk:** Remove the bottles from the collection kits, cap them, and store the breast milk according to safe storage

guidelines.

The pump's dual HygeniKit milk collection system is designed to prevent milk from entering the tubing, ensuring a closed system for hygiene.

## CARE AND MAINTENANCE

Proper cleaning and maintenance are crucial for the longevity and hygienic operation of your breast pump.



Figure 4: Milk storage bottles with clear measurement indicators.

These bottles are part of the collection system and require regular cleaning.

### Cleaning Instructions

- Disassemble:** After each use, disassemble all parts of the milk collection kit that come into contact with breast milk (flanges, valves, bottles, diaphragms).
- Rinse:** Rinse all disassembled parts under cool running water to remove any milk residue.
- Wash:** Wash parts in warm, soapy water using a mild dish soap. Use a dedicated brush for pump parts.
- Rinse Thoroughly:** Rinse all parts thoroughly under clean running water.
- Sterilize:** Sterilize parts by boiling them in water for 5-10 minutes, using a steam sterilizer, or a microwave sterilizer bag. Ensure parts do not touch the bottom or sides of the pot if boiling.
- Air Dry:** Place cleaned and sterilized parts on a clean paper towel or drying rack to air dry completely before reassembly.

reassembly or storage. Do not use a cloth towel as it may transfer germs.

**Important:** The pump motor unit and tubing should not be washed or submerged in water. Wipe the motor unit with a clean, damp cloth if needed. The collection kit parts are not dishwasher compatible.

## TROUBLESHOOTING

---

If you encounter issues with your Ameda Lactaline Personal Double Electric Breastpump, refer to the following common problems and solutions:

Problem	Possible Cause	Solution
No suction or weak suction	Improper assembly, damaged valve/diaphragm, kinked tubing, low battery power.	Ensure all parts are assembled correctly and securely. Check valves and diaphragms for tears or wear; replace if necessary. Straighten tubing. If using batteries, replace them.
Pump not turning on	Power adapter not connected, power outlet issue, dead batteries.	Verify power adapter is securely plugged into pump and outlet. Test outlet with another device. Replace batteries if using battery power.
Milk flow is slow or stops	Incorrect flange size, improper positioning, insufficient stimulation.	Ensure breast flange fits correctly. Reposition the flange. Adjust suction and cycle speed settings. Consider breast massage or warm compress.
Unusual noise from pump	Tubing obstruction, motor issue.	Check tubing for blockages or kinks. If noise persists, contact customer support.

If the problem persists after attempting these solutions, please contact Ameda customer support for further assistance.

## SPECIFICATIONS

---

Detailed technical specifications for the Ameda Lactaline Personal Double Electric Breastpump:




Figure 5: Product dimensions for the breast pump unit.

This diagram illustrates the physical size of the pump for reference.

Feature	Detail
Model Number	EVO17080UK
Pump Type	Electric, Double
Power Source	Corded Electric (AC Adapter included), Battery (6 AA batteries required)
Battery Life	Approximately 90 Hours (with alkaline batteries)
Product Dimensions	33.69 x 23.22 x 34.76 cm
Item Weight	997.9 g (2.2 Pounds)
Special Features	Portable, Ergonomic, Dual HygieniKit Milk Collection System
Dishwasher Compatible	No (for pump parts)

## WARRANTY AND SUPPORT

---

Your Ameda Lactaline Personal Double Electric Breastpump is covered by a manufacturer's warranty.

### Warranty Information

This product comes with a **1 Year Local (Singapore) Manufacturer Warranty**. Please retain your proof of purchase for warranty claims.

### Customer Support

For technical assistance, warranty claims, or any questions regarding your Ameda Lactaline Personal Double Electric Breastpump, please contact Ameda customer support. Refer to the contact information provided on the product packaging or the official Ameda website for the most current support details.

© 2024 Ameda. All rights reserved.