Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

Drive Medical 3655D

Devilbiss Pulmo-Aide Compact Nebulizer Instruction Manual

Model: 3655D

Introduction

The Devilbiss Pulmo-Aide Compact Nebulizer (Model 3655D) is designed to provide effective respiratory therapy in a portable and easy-to-use format. Its compact size makes it suitable for both home use and travel, ensuring consistent therapy wherever needed.
Key features include:

- Effective respiratory care delivery.
- Compact and portable design.
- Durable construction for long-term use.

Product Overview



Front view of the Devilbiss Pulmo-Aide Compact Nebulizer, a portable and compact device designed for respiratory therapy. It features a handle for easy carrying, an air outlet, and a power switch.

The Pulmo-Aide Compact Nebulizer is a compressor-based system that converts liquid medication into a fine mist for inhalation. It is designed for ease of use and reliable performance.

Setup Instructions

1. **Unpack Components:** Carefully remove the nebulizer compressor and all accessories from the packaging. Ensure all parts are present: compressor unit, nebulizer cup, mouthpiece or mask, and air tubing.
2. **Placement:** Place the compressor unit on a flat, stable surface, ensuring that the air vents are not obstructed.
3. **Connect Air Tubing:** Firmly attach one end of the air tubing to the air outlet on the compressor unit.
4. **Prepare Nebulizer Cup:** Open the nebulizer cup and pour the prescribed amount of medication into it. Close the cup securely.
5. **Attach Nebulizer Cup:** Connect the other end of the air tubing to the bottom of the nebulizer cup.
6. **Attach Mouthpiece/Mask:** Attach the mouthpiece or mask to the top of the nebulizer cup.

7. **Power Connection:** Plug the power cord into a grounded electrical outlet (115 Volts, 60 Hz).

Operating Instructions

1. **Initiate Treatment:** Turn on the compressor unit using the power switch located on the side. A fine mist should begin to emerge from the mouthpiece or mask.
2. **Inhale Medication:**
    - If using a mouthpiece: Place the mouthpiece between your teeth and seal your lips around it. Breathe in slowly and deeply through your mouth until all medication is consumed.
    - If using a mask: Place the mask over your nose and mouth, ensuring a snug fit. Breathe normally until all medication is consumed.
3. **Duration:** Continue the treatment until the nebulizer cup is empty and no more mist is being produced. This typically takes 5-15 minutes, depending on the medication volume.
4. **End Treatment:** Turn off the compressor unit and unplug it from the electrical outlet.

Important: Always follow your physician's instructions regarding medication dosage and treatment frequency.

Maintenance and Cleaning

Regular cleaning and maintenance are crucial for optimal performance and hygiene.

Daily Cleaning (After Each Use):

1. **Disconnect:** Disconnect the air tubing, nebulizer cup, mouthpiece/mask from the compressor.
2. **Disassemble:** Disassemble the nebulizer cup into its individual parts.
3. **Wash:** Wash all parts (nebulizer cup, mouthpiece/mask) in warm, soapy water. Rinse thoroughly with clean water.
4. **Air Dry:** Allow all parts to air dry completely on a clean paper towel.

Weekly Disinfection:

1. **Prepare Solution:** Prepare a solution of one part white vinegar to three parts distilled water, or use a medical disinfectant solution as recommended by your physician.
2. **Soak:** Soak the disassembled nebulizer parts (excluding tubing and compressor) in the solution for 30 minutes.
3. **Rinse and Dry:** Rinse thoroughly with sterile water and allow to air dry completely.

Compressor Unit Cleaning:

- Wipe the exterior of the compressor unit with a clean, damp cloth. Do not immerse the unit in water.
- Ensure the air filter is clean and free of dust. Replace the filter as recommended by the manufacturer (typically every 6 months or sooner if discolored).

Troubleshooting

Problem	Possible Cause	Solution

Problem	Possible Cause	Solution
No mist produced or weak mist.	Tubing kinked or disconnected. Nebulizer cup clogged or incorrectly assembled. Air filter dirty. Compressor malfunction.	Check and straighten tubing; ensure secure connections. Disassemble, clean, and reassemble nebulizer cup. Check and replace air filter if dirty. Contact customer support if problem persists.
Compressor not turning on.	Power cord not plugged in. Power switch off. No power at outlet.	Ensure power cord is securely plugged into a working outlet. Flip the power switch to the "ON" position. Test the outlet with another device.
Unusual noise from compressor.	Unit on uneven surface. Internal component issue.	Place unit on a flat, stable surface. Contact customer support.

Specifications

- **Model Number:** 3655D
- **Product Dimensions:** 4 x 7 x 3.8 inches
- **Weight:** 3.8 Pounds
- **Sound Level:** 58 dB
- **Electrical Requirements:** 2.5 A, 115 Volts, 60 Hz
- **Maximum Pressure:** 35 PSI
- **Manufacturer:** Global Supply Outlet
- **ASIN:** B019D4L2LY

Warranty and Support

The Devilbiss Pulmo-Aide Compact Nebulizer comes with a **5-Year Limited Warranty**. This warranty covers defects in materials and workmanship under normal use.

For warranty claims, technical support, or replacement parts, please contact Drive Medical customer service. Refer to the contact information provided with your product packaging or visit the official Drive Medical website for the most up-to-date support details.

Note: Statements regarding dietary supplements have not been evaluated by the FDA and are not intended to diagnose, treat, cure, or prevent any disease or health condition. This product is a medical device; consult with a healthcare professional for proper usage and medical advice.

© 2024 Drive Medical. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B019D4L2LY

More documents and user manuals: **https://manuals.plus/**