

[Manuals.plus](#) /

> [Aurynns](#) /

> Aurynns Digital Veterinary Thermometer Instruction Manual

Aurynns 6299010167887

Aurynns Digital Veterinary Thermometer

Model: 6299010167887

1. PRODUCT OVERVIEW

The Aurynns Digital Veterinary Thermometer is designed for fast and accurate temperature measurement in various animals, including dogs, cats, horses, cattle, pigs, birds, and sheep. It features an easy-to-read LCD display, memory function, and automatic shut-off for convenience.



Image 1.1: Aurynns Digital Veterinary Thermometer with lanyard.



Image 1.2: The thermometer is suitable for a wide range of animals.

Key Features:

- Fast temperature reading (approximately 6 seconds for steady temperature).
- LCD display with memory function.
- Switchable between Celsius (°C) and Fahrenheit (°F).
- Automatic shut-off after 2 minutes of non-operation (or 10 minutes without any operation).
- Reliable, sanitary, and safe for veterinary use.

2. PACKAGE CONTENTS

Verify that all items are present in the package:

- 1 x Aurynn's Digital Veterinary Thermometer
- 2 x AAA Batteries (1.5V)
- 1 x Handy Lanyard
- 1 x Instruction Manual



Image 2.1: Included components: thermometer, batteries, lanyard, and manual.

3. SPECIFICATIONS

Feature	Detail
Measurement Range	32.00°C to 43.00°C (89.6°F to 109.4°F)
Accuracy	+/-0.1°C between 35.00°C to 39.00°C +/-0.2°F between 95°F to 102.20°F +/-0.2°C below 35.00°C or over 39.00°C +/-0.4°F below 95.00°F or over 102.20°F
Response Time	Approximately 6 seconds for steady reading
Battery Type	2 x 1.5V AAA DC Batteries
Battery Life	Approx. 5000 measurement readouts
Dimensions	190 x 110 x 35 mm (7.87 x 4.33 x 1.38 inches)
Weight	6.35 ounces

Feature	Detail
Display Type	LCD
Material	Plastic
Automatic Shut-off	After 2 minutes of non-operation; after 10 minutes without any operation.

4. SETUP

4.1 Battery Installation

The thermometer requires two AAA batteries. These are typically included in the package. To install or replace batteries:

1. Locate the battery compartment cover on the back of the thermometer.
2. Slide or unclip the cover to open the compartment.
3. Insert two AAA batteries, ensuring the correct polarity (+/-) as indicated inside the compartment.
4. Close the battery compartment cover securely.

4.2 Unit Conversion (Celsius/Fahrenheit)

To switch the temperature display between Celsius (°C) and Fahrenheit (°F):

1. Ensure the thermometer is turned OFF.
2. Press and hold the power button for a few seconds.
3. Continue holding until you see either °C or °F displayed on the screen.
4. Release the button. The unit is now set to the desired scale.

5. OPERATING INSTRUCTIONS

Follow these steps for accurate temperature measurement:

1. **Turn On:** Press the power button to turn on the thermometer.
2. **Ready for Use:** The thermometer will emit a beep sound. Once the sound stops, the unit is ready for measurement.
3. **Positioning:** Carefully insert the probe tip into the animal's rectum. *Note: This thermometer is designed for rectal measurement only. Do not use in ear or mouth.*
4. **Reading Temperature:** Within approximately 6 seconds, if the temperature is steady, the thermometer will beep again and display the temperature reading on the LCD screen.
5. **Power Off:** The thermometer will automatically shut off after 2 minutes of non-operation. If no buttons are pressed or readings taken, it will shut off after 10 minutes.



Image 5.1: Proper grip for operating the thermometer.

6. CARE AND MAINTENANCE

6.1 Cleaning

To ensure hygiene and proper function, clean the thermometer after each use:

- Wipe the probe tip and the body of the thermometer with a soft cloth dampened with an antiseptic solution or rubbing alcohol.

- Do not immerse the entire thermometer in water or any cleaning solution.
- Allow the thermometer to air dry completely before storage.

6.2 Storage

Store the thermometer in a clean, dry place, away from direct sunlight and extreme temperatures. Keep it out of reach of children and pets.

7. TROUBLESHOOTING

If you encounter issues with your thermometer, consider the following:

- **No Display/Weak Display:** Check if the batteries are correctly installed. Replace with new AAA batteries if the display is dim or absent.
- **Inaccurate Readings:** Ensure the probe tip is properly inserted and held steady during measurement. Verify the unit is set to the desired temperature scale (°C or °F).
- **Slow Reading:** While the thermometer is designed for fast readings, external factors or improper placement can affect speed. Ensure proper contact and a stable environment.

If problems persist, contact customer support for further assistance.

8. WARRANTY AND SUPPORT

Specific warranty information for the Aurynn's Digital Veterinary Thermometer is not provided within this manual. For details regarding warranty coverage, returns, or technical support, please refer to the product packaging or contact Aurynn's customer service directly through their official website or the retailer where the product was purchased.