Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

Braun IRT6515AM

Braun ThermoScan 6 Ear Thermometer User Manual

Model: IRT6515AM

Introduction

The Braun ThermoScan 6 (IRT6515AM) is an advanced ear thermometer designed for accurate and reliable temperature measurement. Its pre-warmed tip technology ensures precise readings, while the color-coded digital display provides clear fever guidance. This thermometer is suitable for all ages, from infants to adults, and is widely recommended by pediatricians for its ease of use and sanitary disposable lens filters.





Image: The Braun ThermoScan 6 Ear Thermometer, displaying a temperature of 98.6°F on its green digital screen. The device is white with blue accents.

Package Contents

Before beginning, please ensure all items are present in the package:

- Braun ThermoScan 6 Ear Thermometer (Model: IRT6515AM)
- Protective Case
- 60 additional disposable lens filters (total of 81 lens filters included)
- 2 AA batteries (pre-installed or included separately)



Image: An illustration showing the complete package contents: the Braun ThermoScan 6 thermometer, its storage case, a stack of disposable lens filters, and two AA batteries.

Setup

1. **Install Batteries:** If not already installed, open the battery compartment and insert the 2 AA batteries, ensuring correct polarity.
2. **Attach a New Lens Filter:** Press the lens filter ejector button to remove any existing filter. Firmly attach a new, clean disposable lens filter to the thermometer tip. The thermometer will not function correctly without a lens filter attached.
3. **Power On:** Press the Power button () to turn on the thermometer. The display will show all segments for a moment, followed by the last measured temperature.



Image: A detailed diagram highlighting the key components of the Braun ThermoScan 6, including the lens filter, temperature button, battery indicator, digital display, and power button.

Operating Instructions

1. **Prepare for Measurement:** Ensure a new, clean disposable lens filter is securely attached to the thermometer tip.
2. **Position the Thermometer:** Gently insert the probe tip into the ear canal. For accurate readings, ensure the probe is positioned correctly. The ExacTemp positioning system will confirm a good position with a light and a beep.
3. **Take Temperature:** Press and release the Temperature button. The ExacTemp light will pulse during measurement and remain steady when the reading is complete. A long beep indicates the measurement is finished.
4. **Read Temperature:** The temperature will be displayed on the screen. The color of the backlight indicates the temperature range: green for normal, yellow for elevated, and red for high fever.
5. **Eject Lens Filter:** Press the ejector button to remove and discard the used lens filter. Always use a new lens filter for each measurement to ensure hygiene and accuracy.
6. **Power Off:** The thermometer will automatically turn off after 60 seconds of inactivity. You can also manually turn it off

by pressing and holding the Power button.



Image: A close-up of the Braun ThermoScan 6 being held by a hand, with the tip positioned near a baby's ear, illustrating how to take a temperature reading.

Maintenance and Care

- **Cleaning the Thermometer:** Use an alcohol swab or cotton swab moistened with alcohol (70% isopropyl) to clean the thermometer casing and the probe tip after each use. Do not immerse the thermometer in water or any other liquid.
- **Storing the Thermometer:** Store the thermometer in its protective case in a dry place, away from direct sunlight and extreme temperatures.
- **Replacing Batteries:** When the low battery indicator appears on the display, replace the 2 AA batteries. Ensure proper disposal of old batteries according to local regulations.
- **Lens Filters:** Always use genuine Braun disposable lens filters to ensure accuracy and hygiene. Replace the lens filter after each use.

Package Contents





Image: A close-up view of the Braun ThermoScan 6's digital display, clearly showing a temperature reading of 98.6°F. This image emphasizes the clear and easy-to-read screen.

Troubleshooting

Problem	Possible Cause	Solution
"LO" or "HI" displayed	Temperature is outside the measurable range (below 93.2°F / 34°C or above 108°F / 42.2°C).	Ensure proper placement in the ear canal. If the issue persists, consult a healthcare professional.
"POS" or "ERR" displayed	Incorrect probe position or dirty lens filter.	Re-position the probe correctly. Replace the lens filter with a new one.
No display / Thermometer won't turn on	Dead batteries or incorrect battery installation.	Replace batteries with new AA batteries, ensuring correct polarity.

Problem	Possible Cause	Solution
Inconsistent readings	Used or dirty lens filter, improper technique, or earwax obstruction.	Always use a new lens filter. Review operating instructions for proper technique. Clean ear if necessary.

Specifications

- **Model:** IRT6515AM
- **Measurement Range:** 93.2°F – 108°F (34°C – 42.2°C)
- **Accuracy:** Clinically proven
- **Display:** Digital, color-coded (green, yellow, red)
- **Power Source:** 2 x 1.5V AA batteries
- **Automatic Shut-off:** Approximately 60 seconds after last measurement
- **Dimensions:** 5.86 x 1.39 x 2.19 inches
- **Weight:** 10.4 ounces (with batteries)
- **Included Components:** Thermometer, 60 extra filters, 2 AA batteries

Warranty and Support

Braun ThermoScan 6 thermometers are designed for reliability and accuracy. For specific warranty information, please refer to the warranty card included with your product or visit the official Braun Healthcare website. For technical support, troubleshooting assistance, or to purchase additional accessories like lens filters, please contact Braun customer service or visit their official support channels.
You can find more information and support at theBraun Store on Amazon.

Protection Plans: Additional protection plans are available for purchase.

- 3-Year Protection Plan: $7.99
- 4-Year Protection Plan: $10.99
- Complete Protect: $16.99/month (covers all eligible past and future purchases)

© 2024 Braun. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B084W9Y7HP

More documents and user manuals: **https://manuals.plus/**