Beurer 60217

Beurer IH 58 Kids Nebulizer User Manual

Model: 60217 | Brand: Beurer

Introduction

The Beurer IH 58 Kids Nebulizer is an inhalation device designed for nebulizing liquids and liquid medications (aerosols) to treat both upper and lower respiratory tracts. It is suitable for managing conditions such as colds, asthma, and other respiratory diseases. Regular use, as prescribed or recommended by a medical professional, can aid in preventing respiratory illnesses, mitigating their symptoms, and accelerating recovery. For specific medical advice regarding its use, consult your doctor or pharmacist.

Product Overview and Components

The Beurer IH 58 Kids Nebulizer is designed for effective and comfortable inhalation, especially for children. It features a compact design and comes with a comprehensive set of accessories.



The Beurer IH 58 Kids Nebulizer, featuring its compact main unit and the nebulizer attachment with the child-friendly giraffe figure.



A wide range of accessories is included: a mouthpiece, silicone masks for children and babies, an elbow connector, the nebulizer chamber, 10 filters, a connecting tube, and a storage bag for convenience.



The device utilizes compressor compressed air technology, ensuring high nebulization capacity (approximately 0.25 ml/min) and a high percentage of respirable particles, leading to short and efficient inhalation times.



The removable clip-on 'giraffe' figure and silicone masks are designed to make the inhalation process more child-friendly and comfortable for children over one month old.

Key Features:

- Designed for nebulizing liquid medications for upper and lower respiratory tract treatment.
- Effective inhalation for children over one month old with a removable giraffe figure and silicone masks.
- Short inhalation time due to high nebulization capacity (approx. 0.25 ml/min).
- Compact and disinfectable medical device, ideal for travel (8.2 x 19 x 16 cm).
- Includes mouthpiece, child/baby silicone masks, elbow, nebulizer, 10 filters, tube, and storage bag.
- Operates via power supply and USB cable; not battery-operated.

Setup

1. **Unpack Components:** Carefully remove all components from the packaging. Ensure all listed accessories are present.
2. **Connect Air Tube:** Connect one end of the air tube to the air outlet on the main compressor unit and the other end to the air inlet on the nebulizer chamber.
3. **Assemble Nebulizer Chamber:** Open the nebulizer chamber and pour the prescribed liquid medication into the designated reservoir. Do not overfill.
4. **Attach Accessories:** Depending on the user's age and preference, attach either the mouthpiece or the appropriate silicone mask (child or baby mask) to the top of the nebulizer chamber. The giraffe figure can be clipped onto the mask for younger users.
5. **Power Connection:** Connect the USB cable to the nebulizer unit and plug the power adapter into a suitable electrical outlet. The device is not battery-operated and requires a continuous power supply during use.

Operating Instructions

Always consult your doctor or pharmacist regarding the type, dosage, and frequency of medication for inhalation.

1. **Preparation:** Ensure the device is properly assembled and the medication is loaded into the nebulizer chamber.
2. **Positioning:** Have the user sit upright and comfortably. If using a mask, ensure it fits snugly over the nose and mouth. If using a mouthpiece, instruct the user to place it between their teeth and seal their lips around it.
3. **Start Inhalation:** Turn on the nebulizer unit. The device will begin to produce a fine mist (aerosol).
4. **Breathing Technique:** Instruct the user to breathe calmly and deeply through the mouthpiece or mask. Encourage slow, steady breaths to maximize medication delivery to the lungs.
5. **Duration:** Continue inhalation until the medication chamber is empty or as directed by your medical professional. The device is designed for short inhalation times.
6. **Completion:** Once the medication is fully nebulized, turn off the device and disconnect it from the power supply.

Maintenance and Cleaning

Proper cleaning and maintenance are crucial for the hygiene, performance, and longevity of your nebulizer.

Daily Cleaning:

- After each use, disassemble the nebulizer chamber, mouthpiece/mask, and elbow connector.
- Wash these components thoroughly with warm water and a mild detergent. Rinse them well to remove any soap residue.
- **Important:** Disinfect accessories such as masks and tubes only with warm water, not boiling water, as boiling water may deform the silicone mask.
- Allow all parts to air dry completely on a clean surface before reassembling or storing.

Filter Replacement:

- The air filter on the compressor unit should be checked regularly and replaced when it appears discolored or every 10-12 uses, or as indicated in the full product manual.
- Refer to the specific instructions in the detailed manual for the exact location and procedure for replacing the air filter.

Storage:

- Store the nebulizer and its accessories in the provided storage bag in a clean, dry, and dust-free environment.
- Keep out of reach of children.

Troubleshooting

If you encounter issues with your Beurer IH 58 Kids Nebulizer, refer to the following common problems and solutions:

Problem	Possible Cause	Solution
No mist produced or very little mist.	• Nebulizer chamber incorrectly assembled. • Air tube kinked or disconnected. • Medication level too low. • Air filter clogged.	• Check assembly of nebulizer chamber. • Ensure air tube is securely connected and not kinked. • Add more medication if needed. • Replace the air filter.
Device is noisy.	Normal operation noise for a compressor nebulizer.	Some noise is expected from compressor technology. Ensure the device is on a stable, flat surface to minimize vibration.
Mask does not fit properly.	Incorrect mask size or improper positioning.	Ensure you are using the correct silicone mask (child or baby) for the user's face. Adjust positioning for a snug fit. If deformed, replace the mask.
Device stopped working after a few months.	Internal malfunction or power issue.	Ensure the power cable and USB connection are secure and undamaged. Try a different power outlet. If the issue persists, contact customer support.

If the problem persists after attempting these solutions, please contact Beurer customer support or your retailer for further assistance.

Specifications

Attribute	Value

Attribute	Value
Product Name	Beurer IH58 Inhaler for child
Model Number	60217
Brand	Beurer
Dimensions (L x W x H)	8.2 x 19 x 16 cm
Weight	160 grams
Nebulization Capacity	Approx. 0.25 ml/min
Power Source	Mains powered (via USB cable and adapter)
Target Audience	Unisex-Babies, Unisex-Children
Specific Uses	Cold, Asthma
Special Feature	Clinically proven
Country of Origin	China

Safety Information

As a medical device, the Beurer IH 58 Kids Nebulizer must be used with care and according to instructions to ensure safety and effectiveness.

- Always use the device under adult supervision, especially when used by children.
- Do not use the device if it or any of its components are damaged.
- Only use medications prescribed or recommended by a medical professional. Do not use essential oils or other substances not intended for nebulization.
- Ensure proper ventilation during use.
- Keep the device and its accessories clean and disinfected as per maintenance instructions.
- Do not immerse the main compressor unit in water or any other liquid.
- Disconnect the device from the power supply before cleaning or if not in use for an extended period.

Warranty and Support

For warranty information, product support, or to purchase replacement parts, please refer to the official Beurer website or contact your local retailer. Keep your proof of purchase for warranty claims.