Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

Carejoy Teeth Whitening Lamp 3 Mode

Carejoy Teeth Whitening LED Lamp 3 Mode User Manual

Model: Teeth Whitening Lamp 3 Mode

Introduction

This manual provides detailed instructions for the safe and effective use of the Carejoy Teeth Whitening LED Lamp 3 Mode. This device is designed for professional dental teeth whitening, featuring three light modes (blue, purple, red) and an accelerator for use while holding on a dental chair. Please read this manual thoroughly before operation.

Safety Information

Important Safety Precautions:

- For external use only.
- Not for use by children under 18 years of age.
- May cause temporary sensitivity or irritation to teeth and gums.
- If pain or discomfort occurs, discontinue use and consult a dental professional.
- Keep out of reach of children.

Legal Disclaimer: Statements regarding dietary supplements have not been evaluated by the FDA and are not intended to diagnose, treat, cure, or prevent any disease or health condition.

Product Overview

The Carejoy Teeth Whitening LED Lamp features a flexible hose, control panel, heat dissipation holes, and a silicone mouthpiece. It offers three distinct light modes for various whitening and comfort needs.

Figure 1: Overview of the LED lamp components, including the hose, control panel, heat dissipation hole, and silicone mouthpiece. It also illustrates the three light modes: blue, red, and purple.

Figure 2: The complete Carejoy Teeth Whitening LED Lamp system, including the main unit, power adapter, and mounting accessories.

Setup Instructions

1. **Mounting the Lamp:** Securely attach the lamp's adjustable clamp to a dental chair or suitable stable surface.

Ensure the clamp is tightened to prevent movement during operation.



Figure 3: Detail of the adjustable clamp mechanism used to secure the lamp to a dental chair or other fixture.

2. **Adjusting Direction:** The flexible hose allows for adjustment of the lamp's direction. Position the lamp head to comfortably illuminate the patient's mouth.

Your browser does not support the video tag.

Video 1: Demonstration of adjusting the direction of the dental teeth whitening lamp when held on a dental chair. This video illustrates the flexibility of the lamp's arm for optimal positioning.

3. **Power Connection:** Connect the power adapter to the lamp and then plug it into a suitable electrical outlet.
4. **Mouthpiece Preparation:** Ensure the silicone mouthpiece is clean and properly positioned on the lamp head.

Operating Instructions

Follow these steps for effective use of the teeth whitening LED lamp:

1. **Prepare the Patient:** Ensure the patient is comfortable and has any necessary protective eyewear in place. Apply a teeth whitening gel to the patient's teeth as per the gel manufacturer's instructions.
2. **Position the Mouth Guard:** Carefully place the mouth guard (if applicable, or position the lamp head with its silicone mouthpiece) in the patient's mouth, ensuring the light will evenly cover the teeth.



Figure 4: A patient wearing protective eyewear while undergoing a teeth whitening session with the LED lamp.

3. **Turn On and Select Mode:** Press the **ON/OFF** button on the control panel to power on the device. Select the desired light mode (Blue, Red, or Purple) using the corresponding buttons.



Figure 5: The LED screen displaying the timer and indicators for Blue, Red, and Purple light modes, along with control buttons.

4. **Set the Timer:** Use the **TIME UP** and **TIME DOWN** buttons to set the desired duration for the whitening session (0-30 minutes).



Figure 6: Detailed view of the control panel, highlighting the ON/OFF, mode selection (B, R, BR), and timer adjustment buttons.

5. **Post-Whitening Care (Red Light Therapy):** After the whitening session, if the patient experiences any pain or sensitivity, activate the **Red Light** mode for 10-15 minutes to aid in discomfort relief. If no pain is felt, further irradiation is not necessary.
6. **Turn Off and Clean:** Once the session is complete, press the **ON/OFF** button to turn off the LED lamp. Proceed to clean the mouthpiece and lamp head as described in the Maintenance section.

Maintenance

Proper maintenance ensures the longevity and hygiene of your device.

- **Cleaning the Mouthpiece:** The silicone mouthpiece is removable for alcohol disinfection. After each use, detach the mouthpiece and clean it thoroughly with an appropriate disinfectant, such as isopropyl alcohol. Allow it to air dry completely before reattaching.
- **Cleaning the Lamp Head:** Wipe the exterior of the LED lamp head with a soft, damp cloth. For disinfection, use a cloth lightly dampened with a mild, non-abrasive disinfectant. Avoid direct spraying of liquids onto the device.
- **Storage:** Store the device in a clean, dry place away from direct sunlight and extreme temperatures.

- **Inspection:** Regularly inspect the power cord and connections for any signs of damage. Do not use the device if any damage is observed.

Troubleshooting

Problem	Possible Cause	Solution
Device does not power on.	Power cord not properly connected; power outlet malfunction.	Check all power connections. Try plugging into a different outlet.
LED lights are dim or not working.	Device malfunction; insufficient power.	Ensure the device is receiving full power. If problem persists, contact customer support.
Timer not functioning correctly.	Control panel issue.	Restart the device. If the issue continues, contact customer support.
Patient experiences discomfort.	Sensitivity to whitening process.	Discontinue use immediately. Apply red light therapy for 10-15 minutes. Consult a dental professional if discomfort persists.

Specifications

Feature	Detail
Model Number	Teeth Whitening Lamp 3 Mode
Brand	Carejoy
Light Modes	Blue, Red, Purple
Timer Range	0-30 minutes
Product Benefits	Effective teeth whitening with convenient and safe design
Material Features	Durable, Safe, Non-Toxic, Professional Grade
Package Dimensions	14.53 x 9.29 x 3.86 inches
Item Weight	3.66 Pounds
Age Range Description	Adult
UPC	192240923864

Warranty and Support

Warranty Information: This product comes with a standard manufacturer's warranty. Please refer to the product packaging or contact your retailer for specific warranty terms and conditions.
Customer Support: For technical assistance, troubleshooting, or warranty claims, please contact Carejoy customer service. Contact details can typically be found on the product packaging or the official Carejoy website.

© 2023 Carejoy. All rights reserved. Information in this manual is subject to change without notice.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B0936ZR2T6

More documents and user manuals: **https://manuals.plus/**