Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

Mobiclinic Venecia

Mobiclinic Venecia Folding Wheelchair Instruction Manual

Model: Venecia (QA-00039/35-TQ)

Safety Information Setup & Assembly Operation Maintenance Troubleshooting Specifications Warranty & Support

Introduction

This manual provides essential information for the safe and effective use, maintenance, and adjustment of your Mobiclinic Venecia folding wheelchair. Please read it thoroughly before initial use and keep it for future reference.

Safety Information

Your safety is paramount. Always adhere to the following guidelines:

- Ensure all components are securely fastened before each use.
- Always engage the brakes when transferring in or out of the wheelchair.
- Use the safety belt provided to secure the user during transit.
- The anti-tip wheels are designed to prevent backward tipping; ensure they are correctly positioned.
- Do not exceed the maximum weight capacity of 120 kg.
- Avoid using the wheelchair on uneven or excessively steep terrain without assistance.
- Regularly inspect the wheels, brakes, and frame for any signs of wear or damage.



Figure 1: Safety Features. The image highlights the safety belt and hand brake system, crucial for user security.



Figure 2: Anti-Tip Wheels. This image illustrates the anti-tip wheels, which provide maximum safety by preventing the wheelchair from falling backward.

Setup & Assembly

The Mobiclinic Venecia wheelchair is designed for easy setup. Follow these steps for initial assembly and adjustments:

1. Unfolding the Wheelchair

1. Place the folded wheelchair on a flat, stable surface.
2. Gently pull the armrests outwards to begin unfolding the frame.
3. Push down firmly on the seat rails until the chair is fully open and locked into position.

2. Attaching the Cushion

- Place the thick cushion onto the seat for enhanced comfort.



Figure 3: Thick Cushion. The image shows the comfortable, thick cushion placed on the wheelchair seat.

3. Adjusting Armrests

- The armrests are adjustable in height. Locate the adjustment mechanism on the side of the armrest.
- Press the button or lever and slide the armrest to the desired height. Ensure it clicks into place securely.



Figure 4: Adjustable Armrests and Footrests. This image demonstrates the adjustable armrests and the foldable footrests for user convenience.

4. Adjusting Footrests

- The footrests can be swung inwards or outwards and are detachable.
- To adjust, lift the footrest and pivot it to the desired position. To detach, lift and pull it away from the frame.
- Ensure the foot straps are secured around the user's feet to prevent slipping.

5. Adjusting Front Fork Height

- The front forks are adjustable in height to three positions.
- To adjust, loosen the bolt on the front fork, select the desired height, and re-tighten the bolt securely.

Operation

1. Moving the Wheelchair

- To move forward, push the handrims forward. To turn, push one handrim more than the other.
- For assisted movement, the attendant can push the handles at the back of the wheelchair.

2. Engaging Brakes

- To engage the brakes, push the brake levers forward until they lock the wheels.
- Ensure both brakes are engaged when the wheelchair is stationary, especially during transfers.

3. Folding for Transport and Storage

1. Remove the seat cushion.
2. Pull up on the center of the seat fabric to begin folding the chair.
3. Fold the backrest down (split backrest feature).
4. Remove the footrests by lifting and pulling them away.
5. **Quick Release Wheels:** Press the button on the center of the rear wheel hub and pull the wheel outwards to detach it. This significantly reduces the width for transport.
6. Secure any loose parts before lifting.



Figure 5: Folding Process. This image illustrates the quick three-step process to minimize the wheelchair for transport: quick-release wheels, folding the backrest, and removing footrests.



Figure 6: Easy Transport. The folded wheelchair, with wheels removed, fits easily into a car trunk, demonstrating its portability.

▶ **Watch video online**

Video 1: Mobiclinic Venecia Wheelchair Overview. This video demonstrates the key features of the Mobiclinic Venecia wheelchair, including its folding mechanism, adjustable components, and ease of use.

Maintenance

Regular maintenance ensures the longevity and safe operation of your wheelchair.

- **Cleaning:** The seat and backrest are made of nylon and are tensioned with sponge. The covers are removable and washable for easy cleaning. Wipe the aluminum frame with a damp cloth.
- **Wheels:** Inspect polyurethane tires regularly for wear and tear. Ensure quick-release mechanisms are functioning smoothly.
- **Brakes:** Check brake functionality before each use. Adjust if necessary to ensure effective stopping power.
- **Fasteners:** Periodically check all bolts, screws, and fasteners to ensure they are tight.
- **Storage:** Store the wheelchair in a dry, clean place away from direct sunlight and extreme temperatures.

Troubleshooting

Problem	Possible Cause	Solution
Wheelchair does not roll smoothly.	Debris in wheels, loose axle, worn tires.	Clean wheels, check axle tightness, inspect and replace worn tires.
Brakes are not holding.	Brake pads worn, brake cable loose, improper adjustment.	Adjust brake mechanism, replace worn brake pads.
Wheelchair feels unstable.	Loose frame components, anti-tip wheels not engaged.	Check and tighten all frame fasteners. Ensure anti-tip wheels are correctly deployed.

If you encounter issues not listed here or require further assistance, please contact Mobiclinic customer support.

Specifications



Figure 7: Technical Specifications. This diagram provides a visual overview of the wheelchair's dimensions and key features.

Feature	Detail
Overall Length	100 cm (39.4 inches)
Overall Width	70 cm (27.6 inches)
Overall Height	92.5 cm (36.4 inches)
Folded Width	33.5 cm (13.2 inches)
Seat Width	43 cm (16.9 inches)
Seat Length	39.5 cm (15.6 inches)
Backrest Height	43.5 cm (17.1 inches)
Seat Height from Ground	46 cm (18.1 inches)
Maximum Supported Weight	120 kg (264.5 lbs)
Net Weight	16.8 kg (37 lbs)
Frame Material	Aluminum
Upholstery Material	Nylon with sponge tensioning
Front Wheel Size	6 inches
Rear Wheel Size	24 inches
Tire Type	Polyurethane (solid)
Brake Type	Push-to-lock
Special Features	Quick-release rear wheels, split backrest, adjustable height armrests, anti-tip wheels, safety belt, thick cushion.

Warranty & Support

Mobiclinic is committed to providing high-quality products. For specific warranty details, please refer to the warranty card included with your purchase or visit the official Mobiclinic website.

For technical support, spare parts, or any inquiries regarding your Mobiclinic Venecia wheelchair, please contact Mobiclinic customer service through their official channels. Keep your purchase receipt and model number (QA-00039/35-TQ) handy when contacting support.

Mobiclinic S.L. has been a leading manufacturer of clinical and hospital furniture, technical aids, and orthopedics since 1985, ensuring quality and reliability for its customers.

© 2023 Mobiclinic. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B0BHDXX7V8

More documents and user manuals: **https://manuals.plus/**