Viconor Red Light Therapy Device

Viconor Red Light Therapy Device User Manual

Model: Red Light Therapy Device

Introduction

Thank you for choosing the Viconor Red Light Therapy Device. This device is designed to provide red light (660nm) and near-infrared light (850nm) therapy, offering a comprehensive wavelength combination for various applications. This manual provides essential information for the safe and effective use of your device. Please read it thoroughly before operation and retain it for future reference.

Safety Information

Before using the Viconor Red Light Therapy Device, please observe the following safety precautions to prevent injury or damage to the device:

- **Eye Protection:** Always wear the provided protective eyewear during operation to shield your eyes from intense light. Do not look directly into the LED lights.
- **Heat Warning:** The device can reach temperatures up to 122°F (50°C). Avoid prolonged direct skin contact with the device surface during or immediately after use to prevent burns.
- **Medical Conditions:** Consult with a healthcare professional before use if you have any medical conditions, are pregnant, or are taking medications that may increase light sensitivity.
- **Children and Pets:** Keep the device out of reach of children and pets.
- **Water Exposure:** Do not use the device near water or in wet environments. This device is not waterproof.
- **Power Supply:** Use only the original power adapter supplied with the device.



The Viconor Red Light Therapy Device features a curved panel with multiple LED lights, designed for comprehensive light exposure.

Package Contents

Please check the box for the following items:

- Viconor Red Light Therapy Panel
- Adjustable Holder Stand
- Power Adapter (1.8m power cord)
- Protective Eyewear
- Hanging Kit (hooks and chains)
- User Manual



The device includes an adjustable holder stand, allowing the panel to be positioned vertically or horizontally for versatile use. It features anti-slip silicone feet for stability.

Setup

Follow these steps to set up your Viconor Red Light Therapy Device:

1. **Unpack:** Carefully remove all components from the packaging.
2. **Assemble Stand:** Attach the adjustable holder stand to the back of the therapy panel. Ensure it is securely fastened. The stand allows for vertical or horizontal positioning.
3. **Position Device:** Place the device on a stable, flat surface using the stand, or use the hanging kit to suspend it from a secure fixture. Ensure there is adequate ventilation around the device.
4. **Connect Power:** Plug the power adapter into the device's power input port, then plug the adapter into a standard electrical outlet.

Operating Instructions

To use your Viconor Red Light Therapy Device:

1. **Wear Eye Protection:** Always put on the protective eyewear before turning on the device.
2. **Power On:** Press the power button on the device or power cord to turn it on. The red and near-infrared lights will illuminate.
3. **Position for Use:** Position the device approximately 6 to 12 inches (15-30 cm) from the area of your body you wish to treat. The curved design helps focus the light.
4. **Session Duration:** Use the device for 15-30 minutes per session.
5. **Frequency:** For optimal results, use the device daily.
6. **Power Off:** After your session, press the power button to turn off the device and unplug it from the electrical outlet.

Maintenance

Proper care ensures the longevity and performance of your device:

- **Cleaning:** Ensure the device is unplugged and cool before cleaning. Wipe the surface with a soft, dry cloth. Do not use abrasive cleaners or immerse the device in water.
- **Storage:** Store the device in a cool, dry place away from direct sunlight and moisture when not in use.
- **Inspection:** Periodically inspect the power cord and device for any signs of damage. If damage is found, discontinue use and contact customer support.

Troubleshooting

Problem	Possible Cause	Solution
Device does not turn on.	Power cord not properly connected or no power from outlet.	Ensure the power cord is securely plugged into the device and the wall outlet. Test the outlet with another device.
Lights flicker or turn off randomly.	Loose connection or power supply issue.	Check all power connections. If the issue persists, contact Viconor customer support.
Device feels too hot.	Normal operation heat or insufficient ventilation.	The device can reach 122°F (50°C). Ensure adequate airflow around the device. If it feels excessively hot or emits smoke/odor, unplug immediately and contact support.

Specifications

- **Wavelengths:** 660nm (Red Light) & 850nm (Near-Infrared Light)
- **Dimensions:** Approximately 16.54 x 9.45 x 3.23 inches
- **Weight:** Approximately 2.34 Pounds
- **Manufacturer:** Viconor
- **Power Cord Length:** 1.8m

Warranty and Support

Viconor provides a limited warranty for this product against defects in materials and workmanship under normal use. For specific warranty terms and conditions, please refer to the documentation included with your purchase or visit the official Viconor website.

If you encounter any issues or have questions regarding your Viconor Red Light Therapy Device, please contact Viconor customer support for assistance. Our team is dedicated to providing prompt and helpful solutions.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question