Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

Nuliie S32

Nuliie Hands-Free Breast Pump S32 User Manual

MODEL: S32

Product Overview

The Nuliie Hands-Free Breast Pump S32 is an electric wearable breast pump designed for convenience and efficiency. It offers a discreet and comfortable pumping experience, allowing users to multitask while expressing milk. This manual provides detailed instructions for the proper use, care, and maintenance of your S32 breast pump.

Key Features:

- **Hands-Free Design:** Allows for discreet and convenient pumping on the go.
- **Multiple Modes & Levels:** Features 2 modes and 9 adjustable suction levels for personalized comfort and efficiency.
- **Comfortable Flange:** Equipped with a 24MM comfortable flange, with options for other sizes.
- **Memory Function:** Remembers your last used settings for quick startup.
- **Quiet Operation:** Designed for ultra-quiet use, minimizing disturbance.
- **Easy Assembly & Cleaning:** Simple structure for quick setup and hygienic maintenance.



Image: The Nuliie Hands-Free Breast Pump S32, including the pump units, charging cables, and milk storage bags, presented with its packaging.

Components and Assembly

Package Contents:

Please check the package to ensure all components are present:

- Breast Pump Motor Unit (x2)
- Milk Collector Cup (x2)
- Silicone Flange (24mm, x2)
- Silicone Diaphragm (x2)
- Duckbill Valve (x2)

- USB-C Charging Cable (x2)
- User Manual
- Additional Flange Inserts (18mm, 21mm, x2 each)
- Breast Milk Storage Bags

Assembly Instructions:

1. Ensure all parts are clean and dry before assembly.
2. Insert the duckbill valve firmly into the bottom of the milk collector cup. Ensure it is securely seated to prevent leaks and maintain suction.
3. Place the silicone diaphragm onto the top of the milk collector cup, ensuring it forms a tight seal.
4. Attach the silicone flange to the milk collector cup. Press firmly around the edges to ensure a secure fit.
5. Align the assembled milk collector cup with the breast pump motor unit. Press them together until they click into place.



Image: An exploded diagram illustrating the individual components of the Nuliie breast pump and their assembly order, highlighting its

BPA-free, food-grade silicone, skin-friendly, soft-touch, and easy-to-assemble features.

Getting Started

Charging the Pump:

Before first use, fully charge the breast pump motor unit. Use the provided USB-C charging cable and connect it to a suitable power adapter (not included). The display will indicate charging status. A full charge typically takes approximately 1.5 hours.

Flange Size Selection:

Choosing the correct flange size is crucial for comfortable and efficient pumping. The S32 comes with a 24mm flange and additional 18mm and 21mm inserts. Refer to the guide below to determine your optimal size:



Image: A detailed flange size guide showing a nipple measurement chart (nipple size before/after pumping vs. flange size) and visual examples of a nipple that is too small, correctly fitted, and too large within a flange.

To find your size, measure the diameter of your nipple (not the areola) after a pumping session. Select the flange size that is 2-4mm larger than your nipple diameter. A correct fit ensures the nipple moves freely within the tunnel without excessive rubbing or discomfort.

Cleaning and Sterilization Before First Use:

Before using your Nuliie breast pump for the first time, disassemble all parts that come into contact with breast milk (milk collector cup, silicone flange, duckbill valve, silicone diaphragm). Wash them thoroughly with warm soapy water, rinse, and then sterilize them. Sterilization methods include boiling for 5 minutes, steam sterilization, or using a sterilizer designed for baby products.

Operation Guide

Wearing the Pump:

Once assembled, place the breast pump inside your nursing bra, ensuring the flange is centered over your nipple. Adjust your bra for a secure and comfortable fit. The pump should feel snug but not overly tight.

Powering On/Off:

- **To Power On:** Press and hold the power button on the motor unit for 2 seconds. The display will light up.
- **To Power Off:** Press and hold the power button for 2 seconds. The pump will automatically shut off after 30 minutes of continuous use.

Mode Selection:

The Nuliie S32 offers 2 modes: Massage Mode and Expression Mode. Press the 'M' button to switch between modes.

- **Massage Mode:** Indicated by a specific icon on the display. This mode uses rapid, shallow suction to stimulate milk let-down. It is recommended to start with this mode for 2-3 minutes.
- **Expression Mode:** Indicated by a different icon. This mode uses slower, deeper suction to efficiently express milk once let-down has occurred.

Suction Level Adjustment:

Each mode has 9 adjustable suction levels. Use the '+' and '-' buttons on the motor unit to increase or decrease the suction intensity. Start with a low level and gradually increase to a comfortable yet effective suction that does not cause pain.

Pumping Session:

A typical pumping session lasts between 15-30 minutes. The pump has a built-in timer displayed on the screen. The pump will automatically shut off after 30 minutes.



Image: A woman holding a baby while discreetly using the Nuliie breast pump, illustrating its quiet operation (45-55 dB) compared to other sound levels.



Image: A diagram illustrating the efficient and comfortable suction mechanism of the Nuliie breast pump, designed to mimic a baby's natural sucking motion with a pressure range of 10-45 Kpa.



Image: A woman pouring expressed milk from the self-standing Nuliie breast pump into a bottle, demonstrating its spill-free design.

Care and Maintenance

Daily Cleaning:

After each use, disassemble all parts that come into contact with breast milk. Wash them immediately with warm water and mild dish soap. Use a soft brush to clean hard-to-reach areas. Rinse thoroughly with clean water.

Sterilization:

It is recommended to sterilize all milk-contacting parts at least once a day. You can boil them for 5 minutes, use a steam sterilizer, or a sterilizer specifically designed for breast pump parts.

Drying and Storage:

After cleaning and sterilization, allow all parts to air dry completely on a clean drying rack. Do not use cloth towels as they may leave lint. Once dry, store the assembled pump or individual parts in a clean, dry place away from direct sunlight and dust.
Note: Do not wash or immerse the breast pump motor unit in water. Wipe it clean with a damp cloth if necessary.

Troubleshooting

If you encounter any issues with your Nuliie S32 breast pump, please refer to the table below for common problems and their solutions.

Problem	Possible Cause	Solution
No suction or weak suction	Parts not assembled correctly. Duckbill valve damaged or not sealed. Silicone diaphragm not sealed. Flange not properly sealed to breast. Battery low.	Disassemble and reassemble all parts, ensuring a tight fit. Check duckbill valve for tears or damage; replace if necessary. Ensure it is firmly seated. Ensure silicone diaphragm is flat and securely placed. Reposition the pump on your breast to create a good seal. Charge the pump fully.
Pump not turning on	Battery depleted. Power button not pressed correctly.	Charge the pump fully. Press and hold the power button for 2 seconds.
Milk leakage	Milk collector cup overfilled. Parts not assembled correctly. Pump tilted excessively.	Monitor milk volume and empty the cup before it reaches maximum capacity. Reassemble all parts, ensuring tight seals. Maintain an upright position while pumping.
Uncomfortable pumping experience	Incorrect flange size. Suction level too high.	Refer to the flange size guide and try a different insert size. Reduce the suction level to a comfortable setting.

Product Specifications

Specification	Value

Specification	Value
Model	S32
Product Dimensions	4.45"L x 2.72"W x 4.96"H
Weight	1.85 Pounds (approx. for 2 packs)
Material	Silicone
Input	5V-1A
Operation Temperature	5-40°C / 41-104°F
Storage Temperature	-20-60°C / -4-140°F
Operation/Storage Humidity	15-90%
Operation/Storage Pressure	90-106KPa

Warranty and Support

Nuliie products are manufactured with high-quality materials and undergo strict quality control. This product is covered by a manufacturer's warranty against defects in materials and workmanship under normal use. Please refer to the warranty card included in your package for specific terms and conditions.
For any questions, concerns, or technical support, please contact Nuliie customer service:

- **Email:** support@nuliie.com
- **Website:** www.nuliie.com

Please have your model number (S32) and purchase information ready when contacting support.

© 2024 Nuliie. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B0CGNHHK6X

More documents and user manuals: **https://manuals.plus/**