Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

Vulpes Goods PRO (Model 1)

Vulpes Goods BabyCare Handsfree Double Electric Breast Pump PRO User Manual

Model: PRO (Model 1)

Introduction

Thank you for choosing the Vulpes Goods BabyCare Handsfree Double Electric Breast Pump PRO. This manual provides essential information for the safe and effective use of your breast pump. Please read all instructions carefully before first use and keep this manual for future reference.
The Vulpes Goods BabyCare breast pump is designed to provide comfortable and efficient milk expression, offering a hands-free and wireless experience. It features multiple modes and suction levels, an anti-backflow system, and is made from BPA-free, food-grade materials.

Important Safety Information

Please observe the following safety precautions to prevent injury or damage to the product:

- Always inspect all pump components for damage before each use. Discard and replace any damaged parts.
- Do not immerse the main pump unit in water or other liquids. It is not waterproof.
- Keep the device out of reach of children.
- Use only the original USB charging cable provided.
- Do not use the pump while sleeping or drowsy.
- If you experience pain or discomfort during pumping, stop immediately and consult a healthcare professional.
- Ensure all parts are clean and dry before assembly and use to maintain hygiene and prevent contamination.
- This product is for personal use only. Do not share with others.

Product Components

The Vulpes Goods BabyCare Handsfree Double Electric Breast Pump PRO package includes the following items:

- Main Pump Unit PRO (2 units for double pump)
- Bra Adjustment Straps (2 units)
- USB Charging Cable
- Silicone Inserts (19 mm, 21 mm, 24 mm)
- Silicone Breast Shield (28 mm)

- Silicone Valve
- Milk Collection Cup
- Dust Cap
- Breast Milk Storage Bags (60 units)
- Exclusive Travel Bag
- Informative E-Book: "All about breastfeeding" (digital)
- Digital User Manual (via email)



Image: Two main pump units of the Vulpes Goods BabyCare Handsfree Breast Pump PRO.

Setup and Assembly

1. Charging the Pump

Before first use, fully charge the pump unit. Connect the USB charging cable to the charging port on the main pump unit and plug the other end into a compatible USB power source (e.g., laptop, power bank, car charger). A full charge takes approximately 3.5 hours and provides up to 2.5 hours of pumping time.



Image: Illustration of the breast pump's battery capacity and charging capability.

2. Assembly Instructions

Ensure all parts are clean and dry before assembly. Refer to the digital manual for detailed visual instructions.

1. Insert the silicone valve firmly into the bottom of the milk collection cup.
2. Place the silicone breast shield (28 mm) onto the collection cup, ensuring a tight seal.
3. If needed, insert the appropriate size silicone insert (19 mm, 21 mm, or 24 mm) into the 28 mm breast shield. Choose the size that best fits your nipple for optimal comfort and efficiency.

4. Attach the main pump unit securely to the top of the assembled collection cup and breast shield, ensuring it clicks into place.
5. Place the dust cap over the breast shield when not in use for hygiene.

Operating Instructions



Image: Comparison highlighting the hands-free advantage of the Vulpes Goods breast pump over traditional models.

1. Wearing the Pump

Once assembled, place the breast pump inside your bra, ensuring the breast shield is centered over your nipple. The hands-free design allows you to move freely during pumping. Use the bra adjustment straps if needed for a secure and comfortable fit.



Image: A woman demonstrating the hands-free use of the breast pump while engaged in other activities.



Image: Visuals demonstrating the light and portable design of the breast pump, suitable for use anywhere.

2. Powering On and Adjusting Settings

Press and hold the power button on the main pump unit to turn it on. The LED display will illuminate, showing the current mode and suction level. Use the mode button to cycle through the four available pumping modes and the '+' / '-' buttons to adjust the suction intensity.



Image: Visual representation of the four smart modes and adjustable levels of the breast pump.

3. Pumping Modes

- **2-in-1 Mode (1-6 Strengths):** This mode combines gentle massage with powerful pumping cycles to stimulate milk flow and prevent discomfort.
- **Stimulation Mode (1-5 Strengths):** Features weak suction and a fast frequency to comfortably initiate milk let-down.
- **Expression Mode (1-6 Strengths):** Provides strong suction at a slower pace to efficiently express milk.
- **Double Frequency Mode (1-4 Strengths):** Cycles between weak, pause, and strong suction to help correct inverted nipples and facilitate effective pumping.

4. Anti-Backflow System

The pump is equipped with an anti-backflow system to prevent milk from flowing back into the pump unit or tubing, ensuring hygiene and maintaining milk quality.

5. Quiet Operation

The breast pump operates quietly, producing less than 40dB of sound, allowing for discreet pumping without disturbing your baby or others.



Image: Visual representation of the quiet operation of the breast pump, comparing its noise level to common sounds.

Maintenance and Cleaning

Proper cleaning and maintenance are crucial for hygiene and the longevity of your breast pump. All parts that come into contact with breast milk are made from BPA-free, food-grade materials and are easy to clean.

1. Disassembly

After each use, carefully disassemble all parts that came into contact with breast milk: the breast shield, silicone insert, silicone valve, and milk collection cup. The main pump unit should not be disassembled or washed.

2. Cleaning

- Rinse all disassembled parts immediately after use with cold water to remove milk residue.
- Wash parts in warm, soapy water using a mild dish soap. Use a dedicated brush for breast pump parts.
- Rinse thoroughly under running water to remove all soap residue.
- Air dry on a clean drying rack or use a clean cloth. Ensure all parts are completely dry before reassembly or storage to prevent mold or bacterial growth.
- The main pump unit can be wiped clean with a damp cloth. Do not submerge it in water.

3. Sterilization (Optional)

For additional sterilization, parts can be boiled for 5 minutes or sterilized using a steam sterilizer or microwave sterilizer bag. Always follow the manufacturer's instructions for your sterilizer.

4. Storage

Store clean and dry pump components in a clean, dry place away from direct sunlight and dust. Use the provided dust cap for the breast shield and the travel bag for convenient and hygienic storage when on the go.



Image: Features of the breast pump highlighting its food-grade, BPA-free, skin-friendly, soft silicone, and easy-to-clean materials.

Troubleshooting

If you encounter any issues with your breast pump, please refer to the following common problems and solutions:

Problem	Possible Cause	Solution
No suction or weak suction.	◦ Improper assembly. ◦ Damaged silicone valve or breast shield. ◦ Low battery.	◦ Check all connections for a tight seal. ◦ Inspect valve and shield for tears or damage; replace if necessary. ◦ Charge the pump fully.
Pump not turning on.	◦ Low or depleted battery. ◦ Power button not pressed correctly.	◦ Charge the pump. ◦ Press and hold the power button for a few seconds.
Milk leaking from the pump.	◦ Improper assembly. ◦ Collection cup overfilled. ◦ Damaged anti-backflow components.	◦ Ensure all parts are securely assembled. ◦ Do not exceed the maximum fill line on the collection cup. ◦ Inspect anti-backflow components for damage.
Pain or discomfort during pumping.	◦ Incorrect breast shield size. ◦ Suction level too high. ◦ Improper nipple alignment.	◦ Ensure you are using the correct breast shield and insert size. ◦ Reduce the suction level. ◦ Adjust the breast shield to ensure your nipple is centered. ◦ If pain persists, consult a healthcare professional.

Specifications

Feature	Detail
Model	PRO (Model 1)
Battery Capacity	1500mAh
Charging Time	Approx. 3.5 hours
Usage Time (Full Charge)	Up to 2.5 hours
Pumping Modes	4 (2-in-1, Stimulation, Expression, Double Frequency)
Suction Levels	Up to 6 levels per mode (varies by mode)
Noise Level	Less than 40dB
Material	BPA-free, Food-grade Silicone
Breast Shield Size	28 mm (with 19, 21, 24 mm inserts)
Anti-Backflow System	Yes
Charging Interface	USB
Approx. Weight (per unit)	160g (based on package weight, individual unit weight may vary)

Warranty and Support

1. Satisfaction Guarantee

Vulpes Goods offers a 100% satisfaction guarantee. If you are not completely satisfied with your Vulpes Goods BabyCare Handsfree Double Electric Breast Pump PRO, you may return the product within 30 days of purchase.

2. Customer Support

For any questions, concerns, or assistance with your product, please contact Vulpes Goods customer support. You can typically find contact information on the official Vulpes Goods website or through your purchase platform.
Please have your purchase details and model information ready when contacting support.

© 2024 Vulpes Goods. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B0D1KRF81P

More documents and user manuals: **https://manuals.plus/**