SY308

Portable Mesh Nebulizer User Manual

Model: SY308 | Brand: Generic

1. Introduction

Thank you for choosing the Portable Mesh Nebulizer. This device is designed to provide efficient and quiet respiratory treatment for both adults and children. Its compact and portable design makes it ideal for use at home or on the go. Please read this manual carefully before use to ensure proper operation and maintenance.



This image displays the portable mesh nebulizer alongside its product packaging. The packaging highlights key features such as a

nebulization rate of 0.2ml/min, particle size of 5μm±25%, work noise below 50dB, and a bare machine weight of 85g. The model number SY308 is prominently displayed. The device itself is white with a clear mask attachment, showing its compact and user-friendly design.

2. Package Contents

Please check the box for the following items:

- 1 × Portable Mesh Nebulizer
- 1 × Child Mask
- 1 × Adult Mask
- 1 × Nebulizer Mouthpiece
- 1 × Type-C Charging Cable
- 1 × Instruction Manual
- 1 × Color Box (Packaging)

3. Technical Specifications

Parameter	Value
Model Number	SY308
Material	Medical-grade PVC
Mist Modes	3 types (Light, Medium, Strong)
Mist Output	0.2-0.3 ml/minute
Working Noise	< 50dB
Operating Temperature	10 - 40 Degrees Celsius
Power Supply	Type-C USB or 2 AA batteries (not included)
Product Dimensions	12 x 6.5 x 4.3 cm
Net Weight	85 g

4. Setup and Assembly

4.1 Charging the Device

Before first use, fully charge the nebulizer using the provided Type-C USB cable. Connect the cable to the device's

Type-C port and the other end to a compatible USB power adapter (not included) or a computer USB port. The indicator light will show charging status.

4.2 Battery Installation (Optional)

If using AA batteries (not included), open the battery compartment cover and insert 2 AA batteries, ensuring correct polarity (+/-). Close the cover securely.

4.3 Adding Medication

1. Ensure the device is clean and dry.
2. Open the medication cup cover.
3. Carefully pour the prescribed liquid medication into the medication cup. Do not overfill.
4. Close the medication cup cover securely.

4.4 Attaching Accessories

Select the appropriate accessory (adult mask, child mask, or mouthpiece) and firmly attach it to the nebulizer's outlet port.

5. Operating Instructions

1. After setup, press the power button to turn on the nebulizer. The device will start nebulizing.
2. Place the mask over your nose and mouth, or hold the mouthpiece between your lips.
3. Breathe normally and deeply to inhale the medication mist.
4. To switch between the three nebulization modes (light, medium, strong), briefly press the power button while the device is operating.
5. The device will automatically shut off when the medication runs out or after a set time. You can also press and hold the power button to turn it off manually.

Always consult your doctor or pharmacist for proper medication dosage and usage instructions.

6. Maintenance and Cleaning

Proper cleaning and maintenance are crucial for the longevity and hygiene of your nebulizer.

6.1 After Each Use

1. Turn off the device and disconnect any power source.
2. Remove the medication cup and empty any remaining liquid.
3. Rinse the medication cup, mask, and mouthpiece with clean, distilled water.
4. Wipe the main unit with a soft, damp cloth. Do not immerse the main unit in water.
5. Allow all components to air dry completely before reassembling or storing.

6.2 Storage

Store the nebulizer and its accessories in a clean, dry place, away from direct sunlight and extreme temperatures. Keep out of reach of children.

7. Troubleshooting

If you encounter any issues with your nebulizer, please refer to the following common problems and solutions:

Problem	Possible Cause	Solution
Device does not turn on.	Low battery or no power.	Charge the device using the Type-C cable or replace AA batteries.
Weak mist output or no mist.	Insufficient medication, clogged mesh, or incorrect assembly.	Ensure enough medication is in the cup. Clean the mesh thoroughly. Reassemble components correctly.
Unusual noise during operation.	Improper assembly or internal issue.	Check if all parts are securely attached. If noise persists, contact customer support.
Device stops working unexpectedly.	Battery depleted or internal malfunction.	Recharge or replace batteries. If the issue continues, contact customer support.

If the problem persists after trying these solutions, please contact customer support for further assistance.

8. Warranty and Support

This product comes with a standard manufacturer's warranty. For warranty claims, technical support, or any questions regarding the use and maintenance of your Portable Mesh Nebulizer, please refer to the contact information provided on the product packaging or contact the seller directly.

© 2024 Generic. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Answered questions about this manual

What does the blue light mean on the SY308 Portable Mesh Nebulizer, and how can I tell when it is fully charged?

Based on the manual for the SY308 Portable Mesh Nebulizer, here is how to interpret the indicator lights and charging status:Charging Status: When you connect the Type-C USB cable, the indicator light will turn...

Read full answer