VOCIC AY03

VOCIC AY03 Sit to Stand Patient Lift Instruction Manual

Model: AY03 | Brand: VOCIC

1. Introduction and Product Overview

The VOCIC AY03 Sit to Stand Patient Lift is a mobility aid designed to assist caregivers in safely transferring individuals with limited mobility. This unit facilitates sitting, standing, and transfers between various surfaces such as beds, wheelchairs, and toilets, promoting independence and reducing the risk of falls and caregiver injuries.

Key Features:

- **User-Friendly Design:** Foldable seat for easy operation, strategically placed support points at the patella and tibia for patient security.
- **Ergonomic Comfort:** Padded knee and seat support, foam-wrapped handles for secure and comfortable grip.
- **Strong Capacity:** Supports up to 400 lbs, suitable for users with some upper and lower limb mobility.
- **Optimized for Accessibility:** Narrow build and turning radius allow navigation through most doorways and spaces.

2. Safety Information

Always prioritize safety during operation. Read and understand all instructions before using the device. Failure to follow safety guidelines may result in injury to the user or caregiver.

- Ensure all casters are locked before assisting a patient to stand or sit.
- Verify the patient can bear at least 80% of their body weight and has some upper limb strength to grasp handles.
- Always maintain a secure grip on the unit's handles during transfers.
- Do not exceed the maximum weight capacity of 400 lbs.
- Regularly inspect the unit for any signs of wear, damage, or loose components.
- Use a gait belt in conjunction with the lift for added caregiver and patient security, if appropriate.

3. Assembly Instructions

The VOCIC AY03 Sit to Stand Lift is designed for quick and straightforward assembly, typically taking approximately 15 minutes. Follow these steps:

1. **Tighten the Handlebars:** Securely attach the handlebars to the main frame.
2. **Attach the Kneepads:** Install the cushioned kneepads, ensuring they are firmly in place.
3. **Install the Base:** Connect the base unit to the main frame, ensuring all connections are secure.



Figure 3.1: Visual guide for the three-step assembly process.

Detailed Description: This image illustrates the three primary steps for assembling the VOCIC AY03 Sit to Stand Lift. Step 1 shows tightening the handlebars, Step 2 depicts attaching the kneepads, and Step 3 demonstrates installing the base. The image emphasizes the ease of installation.

4. Operating Instructions

Follow these steps for safe and effective patient transfers using the VOCIC AY03 Sit to Stand Lift.

4.1. Preparing for Transfer

- Position the lift close to the patient.

- Lock the casters by pressing down on the brake levers on the rear wheels.
- Ensure the patient's feet are placed firmly on the anti-slip base.
- Adjust the kneepads to comfortably support the patient's shins.
- Ensure the patient can grasp the foam handles securely.



Figure 4.1: Caregiver assisting a patient from a bed using the lift.

Detailed Description: This image shows a caregiver positioning the VOCIC AY03 Sit to Stand Lift next to a patient sitting on a bed. The patient's feet are on the footplate, and their knees are against the knee pads, preparing for a stand-up transfer. The caregiver holds the handles, ready to assist.

4.2. Assisting to Stand

- Instruct the patient to grasp the inner handles and pull themselves up.
- The caregiver provides support and guidance from behind the unit.
- Once the patient is in a standing position, pull the ring to flip the seat pads behind the user.



Figure 4.2: Patient standing upright with support from the lift.

Detailed Description: This image depicts a patient in a standing position, supported by the VOCIC AY03 Sit to Stand Lift.

The patient is holding the handles, and the knee pads provide stability. A caregiver is positioned behind the patient, ready to assist with movement.

4.3. Transferring the Patient

- Unlock the casters by lifting the brake levers.
- Carefully maneuver the lift to the desired destination (e.g., wheelchair, toilet).
- The narrow design allows the unit to fit through most doorways (ensure door width is >25 inches).



Figure 4.3: Dimensions of the lift and its ability to pass through doorways.

Detailed Description: This image highlights the compact dimensions of the VOCIC AY03 Sit to Stand Lift, showing it can fit through doorways wider than 25 inches. Key specifications like weight capacity (400 lbs), patient height range (5'-6'4"),

and maximum base width (25") are displayed.

4.4. Assisting to Sit

- Position the lift in front of the target seating surface.
- Lock the casters.
- Pull the ring to lower the seat pads, allowing the patient to slowly sit down.
- Ensure the patient is fully seated before moving the lift away.



Figure 4.4: The lift's applicability in multiple scenarios, including bed, wheelchair, and toilet transfers.

Detailed Description: This image showcases the versatility of the VOCIC AY03 Sit to Stand Lift, illustrating its use in various home care settings. It demonstrates how the unit can be used for transfers to and from a bed, a wheelchair, and a

toilet, emphasizing its multi-scenario applicability.

4.5. Video Demonstration

Your browser does not support the video tag.

Video 4.5: Official demonstration of the VOCIC AY03 Sit to Stand Lift in use.

Detailed Description: This official product video provides a comprehensive demonstration of the VOCIC AY03 Sit to Stand Lift. It illustrates the proper techniques for positioning the unit, assisting a patient to stand, transferring them, and assisting them to sit, highlighting the ease of use and safety features.

5. Maintenance

Regular maintenance ensures the longevity and safe operation of your VOCIC AY03 Sit to Stand Lift.

- **Cleaning:** Wipe down all surfaces with a damp cloth and mild detergent. Avoid abrasive cleaners.
- **Inspection:** Periodically check all bolts, screws, and connections to ensure they are tight. Inspect casters, kneepads, and seat pads for wear or damage.
- **Storage:** Store the unit in a dry, clean environment when not in use.

6. Troubleshooting

If you encounter any issues with your VOCIC AY03 Sit to Stand Lift, refer to the following common solutions:

- **Difficulty Moving:** Ensure casters are unlocked. Check for any obstructions around the wheels.
- **Unstable During Transfer:** Confirm all casters are locked before initiating a stand or sit. Ensure patient's feet are centered on the base.
- **Pads Not Secure:** Verify that kneepads and seat pads are correctly attached and adjusted.

For issues not listed here, please contact customer support.

7. Specifications

- **Weight Capacity:** Up to 400 lbs
- **Patient Height Range:** 5' - 6'4"
- **Maximum Base Width:** 25 inches
- **Package Dimensions:** 33 x 17 x 11 inches
- **Product Weight:** Approximately 75 Pounds
- **Manufacturer:** VOCIC
- **ASIN:** B0DQ17D5JY

8. Warranty and Support

For warranty information, please refer to the documentation included with your product packaging. For customer support, inquiries, or assistance, please contact VOCIC US directly. Information regarding returns is typically available for an extended period, such as until January 31, 2026, for eligible purchases.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question