

Manuals+

[Q & A](#) | [Deep Search](#) | [Upload](#)

[manuals.plus](#) /

› [McKesson](#) /

› [McKesson Stainless Steel Surgical Instrument Tray B0F3RBVYYG User Manual](#)

McKesson B0F3RBVYYG (8.5 x 3 x 1.5 inch)

McKesson Stainless Steel Surgical Instrument Tray User Manual

Model: B0F3RBVYYG

1. PRODUCT OVERVIEW

The McKesson Stainless Steel Surgical Instrument Tray is designed for the safe and organized handling of medical and surgical instruments. Constructed from durable stainless steel, this tray is autoclavable, ensuring it can withstand high-temperature sterilization processes required in medical environments. It features a fitted lid to protect instruments from environmental contaminants and a strap handle for secure transportation. This product is supplied non-sterile and requires proper sterilization before its first use and subsequent uses.



Figure 1: McKesson Stainless Steel Surgical Instrument Tray with its fitted lid.

2. SETUP AND INITIAL PREPARATION

- Unpacking:** Carefully remove the surgical tray and its lid from the packaging. Inspect for any visible damage during transit.

- Initial Cleaning:** Before first use, thoroughly clean the tray and lid using a mild, non-abrasive detergent and warm water. Rinse completely to remove all detergent residue.
- Sterilization:** As the tray is supplied non-sterile, it must be sterilized prior to its first use. Follow standard medical sterilization protocols, such as autoclaving, to ensure the tray is ready for use with medical instruments.

3. OPERATING INSTRUCTIONS

The McKesson Surgical Instrument Tray is designed for straightforward use in medical settings.

- Placing Instruments:** After sterilization, carefully place cleaned and sterilized instruments into the tray. Ensure instruments are arranged in a manner that prevents damage or contamination.
- Securing the Lid:** Place the fitted lid securely onto the tray to protect the contents from dust, debris, and other environmental factors.
- Transportation:** Utilize the integrated strap handle for safe and stable transportation of the tray and its contents. Always ensure the lid is properly secured before moving the tray.



Figure 2: The surgical tray with instruments, demonstrating its capacity and the open lid.

4. MAINTENANCE AND CARE

Proper maintenance ensures the longevity and hygienic integrity of your McKesson Surgical Instrument Tray.

- **Cleaning After Use:** Immediately after each use, remove all instruments and thoroughly clean the tray and lid. Use appropriate cleaning agents and brushes to remove any biological material or residues.
- **Sterilization:** The tray is autoclavable. Follow your facility's established protocols for steam sterilization. Ensure the tray is completely dry before sterilization to prevent water spots or corrosion.
- **Inspection:** Regularly inspect the tray and lid for signs of wear, damage, or corrosion. Pay close attention to seams and corners. Discontinue use if significant damage is observed.
- **Storage:** Store the clean and sterilized tray in a dry, clean environment, away from direct sunlight and extreme temperatures, to maintain its condition.

5. TROUBLESHOOTING

- **Issue:** Water spots or discoloration after cleaning/sterilization.
Solution: Ensure the tray is thoroughly rinsed and completely dry before sterilization. Use distilled or deionized water for final rinse if tap water quality is an issue.
- **Issue:** Lid does not fit securely.
Solution: Check for any debris or foreign objects obstructing the fit. Inspect the lid and tray edges for deformation or damage. If damaged, the lid or tray may need replacement.
- **Issue:** Visible corrosion.
Solution: Stainless steel is resistant to corrosion, but not immune. Ensure proper cleaning agents are used (avoid harsh chemicals or steel wool). Ensure thorough drying after washing and sterilization. If corrosion persists, the tray may need replacement.

6. SPECIFICATIONS

Feature	Detail
Brand	McKesson
Model Number	B0F3RBVYYG
Material	Stainless Steel
Color	Stainless Steel
Product Dimensions (L x W x H)	8.5 inches x 3 inches x 1.5 inches
Autoclavable	Yes
Sterility	Non-Sterile (requires sterilization before use)
Features	Fitted Lid, Strap Handle





7. WARRANTY AND SUPPORT

For specific warranty information regarding your McKesson Surgical Instrument Tray, please refer to the

documentation provided at the time of purchase or contact your supplier. For product support or inquiries, please reach out to McKesson customer service or your authorized McKesson distributor.

© 2023 McKesson. All rights reserved.

Related Documents - B0F3RBVYYG (8.5 x 3 x 1.5 inch)

	<p>McKesson Disposable Vaginal Specula: Instructions for Use and Product Information</p> <p>Comprehensive guide to McKesson's disposable vaginal specula, covering usage instructions, cautions, indications, sterilization, and product specifications for medical professionals.</p>
	<p>McKesson Refrigerator/Freezer Thermometer TRIPLE DISPLAY SINGLE PROBE MCK821RFV</p> <p>Detailed information and operating instructions for the McKesson Refrigerator/Freezer Thermometer, Triple Display, Single Probe (Model MCK821RFV). Features include high accuracy, min/max memory, alarm settings, and NIST traceability.</p>
	<p>McKesson ErgoCart Cleaning and Disinfection Instructions</p> <p>Detailed cleaning and disinfection instructions for McKesson ErgoCart storage carts and mobile workstations, including recommended cleaning agents and processes.</p>
	<p>McKesson Wireless Refrigerator/Freezer Thermometer MCK895WRFT - Triple Display, Single Probe</p> <p>User manual and specifications for the McKesson Wireless Refrigerator/Freezer Thermometer, model MCK895WRFT. Details features like triple display, single probe, large digit, wireless connectivity, alarm settings, and error codes.</p>
	<p>McKesson Refrigerator/Freezer Thermometer TRIPLE DISPLAY DUAL PROBE MCK821RFV2 User Manual</p> <p>User manual and specifications for the McKesson Refrigerator/Freezer Thermometer TRIPLE DISPLAY DUAL PROBE (Model MCK821RFV2). Features include high accuracy, min/max memory, low/high alarm settings, and NIST traceability.</p>
	<p>McKesson Consult Fentanyl Urine Test Cassette: Instructions and Information</p> <p>Comprehensive guide to the McKesson Consult Fentanyl Urine Test Cassette, including intended use, test procedure, result interpretation, warnings, and mailing instructions for confirmation testing.</p>

