Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

loktch 10Ah Wheelchair Replacement Battery

loktch 10Ah Wheelchair Replacement Battery User Manual

Comprehensive instructions for the safe and effective use of your loktch 10Ah Wheelchair Replacement Battery.

1. Product Overview

The loktch 10Ah Wheelchair Replacement Battery is designed to provide reliable power for compatible electric wheelchairs. This lithium-ion battery offers a 10 Ampere-hour capacity, ensuring extended operation and consistent performance. It is engineered for straightforward replacement and integration into your mobility device.



Figure 1: The loktch 10Ah Wheelchair Replacement Battery (left) shown alongside a compatible powered wheelchair (right). The battery is a black rectangular unit with a red connector.

2. Setup and Installation

Before installing the battery, ensure the wheelchair is powered off and disconnected from any charging source. Refer to your wheelchair's specific manual for battery compartment access and removal procedures.

1. **Safety First:** Always wear appropriate personal protective equipment, such as gloves and eye protection, when handling batteries.
2. **Remove Old Battery:** Carefully disconnect the old battery from the wheelchair's power connectors. Note the orientation of the connectors to ensure correct installation of the new battery.
3. **Insert New Battery:** Place the loktch 10Ah Replacement Battery into the designated battery compartment of your wheelchair. Ensure it fits securely and does not move freely.

4. **Connect Power:** Connect the battery's red connector to the corresponding port on the wheelchair. Ensure a firm and secure connection.
5. **Secure Compartment:** Close and secure the battery compartment according to your wheelchair's instructions.
6. **Initial Charge:** It is recommended to fully charge the battery before its first use. Refer to the "Operating" section for charging instructions.

Important: *If you are unsure about any step of the installation process, consult a qualified technician or your wheelchair's manufacturer.*

3. Operating Instructions

3.1 Charging the Battery

Use only the charger specifically designed for your electric wheelchair and compatible with lithium-ion batteries. Using an incompatible charger can damage the battery and pose safety risks.

- Connect the charger to a standard electrical outlet.
- Connect the charger's output plug to the charging port on your wheelchair.
- Monitor the charger's indicator lights (if available) to confirm charging status. Charging typically takes several hours.
- Disconnect the charger once the battery is fully charged. Avoid overcharging for extended periods.

3.2 Discharging and Usage

The battery will discharge as the wheelchair is used. The wheelchair's control panel typically displays the current battery level. It is advisable to recharge the battery before it is completely depleted to prolong its lifespan.

4. Maintenance

- **Regular Charging:** For optimal battery life, avoid letting the battery remain fully discharged for long periods. Charge it regularly, even if the wheelchair is not in frequent use.
- **Storage:** If storing the battery for an extended period (more than one month), charge it to approximately 50-70% capacity and store it in a cool, dry place away from direct sunlight and extreme temperatures. Recharge it every 3-6 months to maintain its health.
- **Cleaning:** Keep the battery and its connectors clean and free from dust and debris. Use a dry, soft cloth for cleaning. Do not use liquid cleaners directly on the battery or connectors.
- **Temperature:** Operate and store the battery within the recommended temperature range specified by your wheelchair manufacturer. Extreme temperatures can negatively impact battery performance and lifespan.
- **Inspection:** Periodically inspect the battery for any signs of damage, swelling, or leakage. If any issues are observed, discontinue use immediately and contact support.

5. Troubleshooting

Problem	Possible Cause	Solution
Battery not charging	Charger not connected properly; faulty charger; battery issue.	Ensure charger is securely plugged into outlet and wheelchair. Check charger indicator lights. Try a different outlet. If problem persists, contact support.

Problem	Possible Cause	Solution
Wheelchair not powering on	Battery fully discharged; battery not connected; battery fault.	Charge the battery. Verify battery is securely connected. If fully charged and connected, and still no power, contact support.
Reduced operating time	Battery degradation; improper charging habits; environmental factors.	Ensure regular and full charging cycles. Review maintenance guidelines. If significant reduction persists, battery may need replacement.

For issues not listed above or if solutions do not resolve the problem, please contact customer support.

6. Specifications

- **Brand:** loktch
- **Model:** 10Ah Wheelchair Replacement Battery
- **Battery Capacity:** 10 Amperes Hours
- **Battery Composition:** Lithium-ion
- **Product Weight:** 2 Kilograms
- **Product Dimensions:** 17.2 x 6 x 26 cm
- **Number of Units:** 1
- **ASIN:** B0FLVC883G
- **UPC:** 701347269944

7. Warranty and Support

This loktch product is covered by a standard manufacturer's warranty. Please retain your proof of purchase for warranty claims. Specific warranty terms and conditions may vary and are typically provided at the time of purchase or can be found on the manufacturer's official website.
For technical assistance, troubleshooting, or warranty inquiries, please contact loktch customer support through their official channels. Provide your product model and purchase details when contacting support to expedite assistance.

© 2025 loktch. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page’s manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B0FLVC883G

More documents and user manuals: **https://manuals.plus/**