Saved from manuals.plus.
Open the live page for links to other documents and user manuals.

VEVOR SC7060B

VEVOR Raised Toilet Seat with Arms

Model: SC7060B - User Manual

Introduction Safety Information Package Contents Setup & Installation Operation Maintenance Troubleshooting Specifications Support

1. Introduction

The VEVOR Raised Toilet Seat with Arms (Model SC7060B) is designed to provide a 5-inch (127 mm) lift to your toilet seat, enhancing comfort and ease of use for individuals who may experience difficulty sitting down or standing up from a standard toilet. This includes pregnant women, seniors, patients recovering from surgery, and individuals with disabilities. The elevated design helps reduce strain on hips, knees, legs, and Achilles tendons.

2. Safety Information

Please read all instructions carefully before installation and use. Failure to follow these instructions may result in injury or damage to the product.

- **Weight Capacity:** Do not exceed the maximum load capacity of 300 lbs (136.1 kg).
- **Secure Installation:** Always ensure the raised toilet seat is securely installed and stable before each use. Verify that the knob-lock mechanism is fully tightened.
- **Compatibility:** This product is designed for most standard round and elongated toilets. It is**not** compatible with groove-style waterfall toilets or toilets with sloped seat rings. Refer to the "Setup & Installation" section for compatibility checks.
- **Cleaning:** Use only mild, non-abrasive cleaners. Avoid harsh chemicals that may damage the product's surface or materials.
- **Inspection:** Regularly inspect the seat and armrests for any signs of wear, damage, or loose components. Discontinue use if any damage is found and contact support.
- **Supervision:** Individuals with severe mobility issues or cognitive impairments should be supervised during use.

3. Package Contents

Verify that all components are present and undamaged upon opening the package.

- 1 x Raised Toilet Seat
- 2 x Armrest Tubes (with padded handles)

- 1 x User Manual (this document)



Image: All components included in the VEVOR Raised Toilet Seat with Arms package.

4. Setup & Installation

The VEVOR Raised Toilet Seat with Arms features a knob-lock installation system, requiring no tools or drilling. Follow these steps for secure setup:

4.1. Toilet Compatibility Check

Before installation, ensure your toilet is compatible. This seat fits most standard round and elongated toilets. It is **not** suitable for groove-style waterfall toilets or toilets with sloped seat rings.

- **Toilet Bowl Opening Dimensions:**
  - Inner Width: ≥ 8.2 in (208 mm)
  - Inner Length: ≥ 9.3 in (237 mm)
- **Toilet Bowl Outer Dimensions:**
  - Length: ≥ 20.4 in (519 mm)
  - Width: ≤ 17.2 in (438 mm)



Image: Visual guide for toilet compatibility and the knob-lock mechanism.

4.2. Installation Steps

1. **Prepare the Toilet:** Ensure your toilet bowl is clean and dry before installation. Lift your existing toilet seat and lid.
2. **Position the Raised Seat:** Carefully place the VEVOR Raised Toilet Seat directly onto the ceramic rim of your toilet bowl. Ensure it is centered and aligned properly.
3. **Secure with Knob-Lock:** Locate the central knob on the front of the raised toilet seat. Turn the knob clockwise to tighten the internal locking mechanism. Continue turning until the seat is firmly secured to the toilet bowl and does

not wobble. Do not overtighten.

4. **Install Armrests:** Insert the two armrest tubes into the designated slots on either side of the raised toilet seat. Ensure they click into place or are securely fastened according to the design. The padded handles should face upwards for comfortable gripping.
5. **Final Check:** Gently test the stability of the installed seat and armrests before initial use. Ensure there is no movement or looseness.



Image: The raised toilet seat installed, illustrating the 5-inch lift height.

5. Operation

The VEVOR Raised Toilet Seat is designed for straightforward use to assist with sitting and standing.

- **Sitting:** Approach the toilet facing forward. Use the padded armrests for support as you slowly lower yourself onto the raised seat.
- **Standing:** To stand up, place your hands firmly on the padded armrests and push down to assist yourself in rising. The armrests provide a secure, non-slip grip.



Image: User demonstrating safe sitting and standing with the aid of the armrests.



Image: The wide seat space (21.1 inches / 535 mm) between the armrests.

6. Maintenance and Cleaning

Regular cleaning and maintenance will ensure the longevity and hygiene of your raised toilet seat.

- **Daily Cleaning:** Wipe down the seat and armrests with a damp cloth and a mild, non-abrasive household cleaner. The waterproof and dampness-resistant surface allows for effortless cleaning.
- **Avoid Harsh Chemicals:** Do not use abrasive cleaners, solvents, or strong chemical disinfectants, as these can damage the HDPE material and aluminum alloy.
- **Drying:** Ensure the seat is dry after cleaning to prevent water spots and maintain hygiene.
- **Inspection:** Periodically check all connections, especially the knob-lock mechanism and armrest attachments, to ensure they remain secure.



Image: Demonstrating the easy cleaning process of the raised toilet seat.

7. Troubleshooting

If you encounter any issues with your VEVOR Raised Toilet Seat, refer to the following common problems and solutions:

Problem	Possible Cause	Solution
Seat is unstable or wobbles.	Knob-lock not fully tightened; incompatible toilet shape; improper placement.	Ensure the central knob is turned clockwise until the seat is firmly secured. Verify your toilet meets the compatibility requirements (round/elongated, correct inner and outer dimensions). Re-position the seat to ensure it is centered on the toilet bowl.
Armrests feel loose.	Armrest tubes not fully inserted or secured.	Remove and re-insert the armrest tubes, ensuring they click firmly into their designated slots.

Problem	Possible Cause	Solution
Seat does not fit my toilet.	Toilet is not compatible with the raised seat design.	Review the "Toilet Compatibility Check" section. This seat is not compatible with groove-style waterfall toilets or those with sloped seat rings. Ensure your toilet's inner and outer dimensions meet the specified requirements.

8. Specifications

Feature	Detail
Item Model Number	SC7060B
Lift Height	5 in (127 mm)
Load Capacity	300 lbs (136.1 kg)
Product Dimensions (L x W x H)	21.1 x 17.7 x 13.0 in (535 x 450 x 330 mm)
Installation Size Range (Toilet Bowl Dimensions)	*Bowl Opening:* Length ≥ 9.3 in (237 mm), Width ≥ 8.2 in (208 mm) *Overall Bowl:* Length ≥ 20.4 in (519 mm), Width ≤ 17.2 in (438 mm)
Material	HDPE Seat & PP Knob & Aluminum Alloy Armrest Tubes
Net Weight	5.7 lbs (2.6 kg)



Image: Detailed product dimensions and specifications.

9. Warranty and Support

For warranty information, technical support, or any questions regarding your VEVOR Raised Toilet Seat with Arms, please contact VEVOR customer service. Refer to the contact information provided with your purchase or visit the official VEVOR website for assistance.

Website: www.vevor.com

© 2026 VEVOR. All rights reserved.

Ask a question about this manual

Ask about setup, troubleshooting, compatibility, parts, safety, or missing instructions. Manuals+ will review the question and use this page's manual context to help answer it.

Name

Anonymous

Email

you@example.com

After a successful submission, we securely hash the supplied account or form email before sending it to Microsoft Advertising for conversion attribution. Privacy details.

Question

Example: How do I reset this device or fix this error?

Details

Model number, symptoms, what you tried, or the section of the manual you are using.

Submit question

Open the live page

https://manuals.plus/asin/B0FYN9J5DB

More documents and user manuals: **https://manuals.plus/**