



B K PRECISION 4011A 5 MHz Function Generator with Digital Display Instructions

[Home](#) » [B&K Precision](#) » B K PRECISION 4011A 5 MHz Function Generator with Digital Display Instructions 

B K PRECISION 4011A 5 MHz Function Generator with Digital



Contents

1 TEST INSTRUMENT SAFETY
1.1 WARNING
2 INTRODUCTION
3 SPECIFICATIONS
3.1 FREQUENCY CHARACTERISTICS
3.2 OUTPUT CHARACTERISTICS
3.3 SINE WAVE
3.4 SQUARE WAVE
3.5 TRIANGLE WAVE
3.6 TTL OUTPUT
3.7 CMOS OUTPUT
3.8 VCG (Voltage Controlled Generator) INPUT
3.9 FREQUENCY COUNTER
3.10 AC INPUT
3.11 DIMENSIONS (H x W x D)
3.12 WEIGHT
3.13 ACCESSORIES
4 CONTROLS AND INDICATORS
4.1 FRONT PANEL (Refer to Fig. 1)
5 OPERATING INSTRUCTIONS
6 FREQUENCY AND WAVEFORM SELECTION
6.1 CONSIDERATIONS
6.2 DUTY CYCLE CONTROL
7 TTL/CMOS OUTPUT
7.1 VOLTAGE CONTROLLED FREQUENCY OPERATION
7.2 OUTPUT PROTECTION CONSIDERATIONS
7.3 FUNCTION GENERATOR APPLICATIONS
GUIDEBOOK
8 MAINTENANCE
8.1 WARNING
8.2 FUSE REPLACEMENT
8.3 INSTRUMENT REPAIR SERVICE
8.4 AC LINE SELECTION
9 CUSTOMER SUPPORT
9.1 Service Information
9.2 Limited Two-Year Warranty
9.3 TEST INSTRUMENT SAFETY
10 Documents / Resources
10.1 References
11 Related Posts

TEST INSTRUMENT SAFETY

WARNING

Normal use of test equipment exposes you to a certain amount of danger from electrical shock because testing must sometimes be performed where exposed voltage is present. An electrical shock causing 10 milliamps of current to pass through the heart will stop most human heartbeats. Voltage as low as 35 volts dc or ac rms should be considered dangerous and hazardous since it can produce a lethal current under certain conditions. Higher voltages pose an even greater threat because such voltage can more easily produce a lethal current. Your normal work habits should include all accepted practices to prevent contact with exposed high voltage, and to steer current away from your heart in case of accidental contact with a high voltage. You will significantly reduce the risk factor if you know and observe the following safety precautions:

1. Don't expose high voltage needlessly. Remove housings and covers only when necessary. Turn off equipment

- while making test connections in high-voltage circuits. Discharge high-voltage capacitors after removing power.
2. If possible, familiarize yourself with the equipment being tested and the location of its high voltage points.
However, remember that high voltage may appear at unexpected points in defective equipment.
 3. Use an insulated floor material or a large, insulated floor mat to stand on, and an insulated work surface on which to place equipment; and make certain such surfaces are not damp or wet.
 4. Use the time proven "one hand in the pocket" technique while handling an instrument probe. Be particularly careful to avoid contacting a nearby metal object that could provide a good ground return path.
 5. When testing ac powered equipment, remember that ac line voltage is usually present on some power input circuits such as the onoff switch, fuses, power transformer, etc. any time the equipment is connected to an ac outlet, even if the equipment is turned off.

(continued on inside back cover)

INTRODUCTION

The B+K Precision Model 4011A Function Generator is a versatile signal source which combines several functions into one unit. Additionally, the instrument provides the added convenience of a built-in frequency counter. This permits more accurate determination of output frequency than is possible with a simple calibrated dial. Coarse and fine tuning controls permit precision settability of the output frequency. High stability assures that the output frequency does not drift.

With this versatility, the unit has a vast number of applications in both analog and digital electronics in the engineering, manufacturing, servicing, educational, and hobbyist fields.

The heart of the function generator is a VCG (voltagecontrolled generator) that produces precision sine, square, or triangle waves over the 0.5Hz to 5MHz range. This encompasses subaudible, audio, ultrasonic, and RF applications. A continuously variable dc offset allows the output to be injected directly into circuits at the correct bias level.

Variable symmetry of the output waveform converts the instrument to a pulse generator capable of generating rectangular waves or pulses, ramp or sawtooth waves, and slewed sine waves.

In addition to the above features, an external signal may be used to sweep the output frequency or control operating frequency. This is useful in situations where an externally controlled frequency is desirable.

SPECIFICATIONS

FREQUENCY CHARACTERISTICS

Waveforms

Sine, Square, Triangle, \pm Pulse, \pm Ramp

Range

0.5Hz to 5MHz in 7 ranges

Resolution

4 digits

Tuning Range

Coarse: 10:1, Fine: $\pm 5\%$ of Coarse Setting

Variable Duty Cycle

15:85:15 Continuously Variable

Operating Modes

Normal, VCG (Voltage Controlled Generator)

Frequency Stability

Output frequency will not change more than 0.09% in 15 minutes after 1 hour warmup.

OUTPUT CHARACTERISTICS**Impedance**

50 Ω \pm 10%

Level

20V p-p Open-circuit, 10V p-p into 50 Ω

Amplitude Control

Variable, 20dB range typical

Attenuation

-20 dB \pm 1 dB

DC Offset

Preset: \pm 0.1V typical

Variable: \pm 10V open-circuit, \pm 5V into 50 Ω

SINE WAVE**Distortion**

3% typical at 1kHz

Flatness

\pm 5% (.45 dB)

SQUARE WAVE**Symmetry**

0.5Hz to 100kHz \leq 2%

Rise Time

\leq 20nS

TRIANGLE WAVE**Linearity**

\geq 98% to 100kHz

TTL OUTPUT

Level	0.8V to 2.4V
Rise Time	$\leq 20\text{nS}$
Duty Cycle	50% typical

CMOS OUTPUT

Max Frequency	2MHz
Level	4V to 14V $\pm 0.5\text{V}$ p-p, Continuously variable
Rise Time	$\leq 120\text{nS}$

VCG (Voltage Controlled Generator) INPUT

Input Voltage

0 – 10V $\pm 1\text{V}$ causes a 100:1 frequency change (Typical)

Impedance

10 k Ω $\pm 5\%$

FREQUENCY COUNTER

Accuracy

Time Base Accuracy ± 1 count

Time Base Accuracy

± 10 PPM (23 °C ± 5 °C)

Display

4 digit LED

AC INPUT

120 / 230VAC $\pm 10\%$, 50 / 60Hz, internal jumper selectable

DIMENSIONS (H x W x D)

10-3/8" x 3-3/8" x 11-7/16" (26.4cm x 8.6cm x 29.1cm)

WEIGHT

4 lb. (1.8 kg.)

ACCESSORIES

Instruction Manual

Output Cable, BNC to Alligator Clips


NOTE: Specifications and information are subject to change without notice. Please visit www.bkprecision.com for the most current product information.

CONTROLS AND INDICATORS

FRONT PANEL (Refer to Fig. 1)

1. **POWER Switch.** Turns power on and off.
2. **RANGE Switch.** Selects output frequency range. Seven ranges from 5Hz to 5MHz. Switch indicates maximum frequency of range and is adjusted with COARSE FREQUENCY control to 0.1 times the maximum. For example, if the 500kHz range is selected, the output frequency can be adjusted from 50kHz to 500kHz.
3. **FUNCTION Switch.** Selects sine, square, or triangle waveform at OUTPUT jack.
4. **OUTPUT LEVEL Control.** Controls the amplitude of the signal at the OUTPUT jack. Output level can be decreased by approximately 20 dB with this control.
5. **DC OFFSET Control.** Enabled by the DC OFFSET Switch (12). Clockwise rotation from center changes the DC offset in a positive direction while counterclockwise rotation from center changes the DC offset in a negative direction.
6. **OUTPUT Jack.** Waveform selected by FUNCTION switch as well as the superimposed DC OFFSET voltage is available at this jack.
7. **TTL/CMOS Jack.** TTL or CMOS square wave, depending on the position of the CMOS LEVEL switch (13) is output at this jack. This output is independent of the OUTPUT LEVEL and DC OFFSET controls.
8. **CMOS LEVEL Control.** Rotating this control clockwise increases the amplitude of the CMOS square wave at the TTL/CMOS jack.
9. **VCG Jack.** Voltage Controlled Generator input. Permits external control of generator output frequency by a DC voltage input at this jack. A positive voltage will decrease frequency.
10. **DUTY CYCLE Control.** Enabled by the DUTY CYCLE Switch (14). Rotation from center position adjusts the duty cycle of the main OUTPUT signal.
11. **-20dB Switch.** When engaged, the signal at the OUTPUT jack is attenuated by 20dB.
12. **DC OFFSET Switch.** When engaged, enables operation of the DC OFFSET control (5).

Figure 1. Model 4011A Controls and Indicators.



13. **CMOS LEVEL Switch.** When engaged, changes the TTL signal to CMOS signal at the TTL/CMOS jack.
14. **DUTY CYCLE Switch.** When engaged, enables operation of DUTY CYCLE control (10).

15. **FINE FREQUENCY Control.** Vernier adjustment of the output frequency for ease of setting frequency.
16. **COARSE FREQUENCY Control.** Coarse adjustment of the output frequency from 0.1 to 1 times the selected range.
17. **COUNTER DISPLAY.** Displays frequency of internally generated waveform.
18. **GATE LED.** Indicates when the frequency counter display is updated. When the 50K through 5M ranges are selected, the LED will flash 10 times per second (every 0.1 seconds). When the 50 through 5K ranges are selected, the LED will flash once every second and when the 5 range is selected, the LED will flash every 10 seconds. As the LED turns off, the display is updated.
19. **Hz and KHz LED.** Indicates whether the counter is reading in Hz or kHz.
20. **Inverter Switch**

OPERATING INSTRUCTIONS

The B+K Precision Model 4011A Function Generator is a versatile instrument, capable of producing a variety of output waveforms over a broad range of frequencies. To gain a working familiarity with the unit, it is recommended that it be connected initially to an oscilloscope, so that the effects of the various controls on the output waveforms can be observed. Use this manual as required for reference until becoming accustomed to the operating procedures.

FREQUENCY AND WAVEFORM SELECTION

1. Initially, verify that the DUTY CYCLE (14), CMOS LEVEL (13), DC OFFSET (12), -20dB (11) switches are in the OUT position (released). This will produce a symmetrical waveform unaffected by the other controls.
2. Plug the unit into an appropriate power source and turn it on by engaging the POWER switch (1).
3. Select the desired waveform (SINE, SQUARE, or TRIANGLE) by engaging one of the FUNCTION switches (3). Phase relationships of the waveforms are shown in Fig. 2.
4. Select the frequency of the waveform by engaging one of the RANGE switches (2). The output frequency is displayed, along with the appropriate measurement units, kHz or Hz (19), on the LED display.
5. Rotate the COARSE (16) frequency control to quickly set the output frequency to the approximate desired value. The FINE (15) frequency control can then be used to easily set the output to the specific desired value. The frequency selected is available at the OUTPUT jack (6). In addition, a digital signal, either TTL or CMOS is available at the TTL/CMOS jack (7) (refer to the "TTL/CMOS OUTPUT" section of this manual).





Figure 2. Output Waveform and Phase Relationship

6. Adjust the amplitude of the output as desired using the OUTPUT LEVEL control (4). Rotation of this control varies the amplitude from maximum to 20dB below maximum. An additional attenuation of -20dB is available by pushing in the -20dB switch (11). The attenuation factors can be combined for a total of -40dB. The maximum signal level is 10V p-p (into 50 Ω).
7. A superimposed DC component can be added to the output signal by engaging the DC OFFSET switch (12) to enable operation of the DC OFFSET control (5). Rotation of this control adds a positive or negative DC component to the output signal. The DC component introduced is independent of the OUTPUT LEVEL control and can be varied by ± 10 volts open circuited or ± 5 volts into 50 Ω . The DC Offset does not affect the TTL/CMOS output jack. The effect of DC OFFSET is shown in **Fig. 3**.

CONSIDERATIONS

1. Counterclockwise rotation of the COARSE frequency control decreases the output frequency to approximately one-tenth of the maximum for the range selected (10:1). For example, if the 50K range is selected and the COARSE frequency control is set to full counterclockwise, the output frequency is approximately 5kHz.
2. It is advisable to set the FINE frequency control to the approximate center of its rotation before setting the COARSE frequency control. This assures that the FINE control will not reach its limit while trying to finalize the frequency setting.
3. The FINE frequency control provides approximately $\pm 5\%$ frequency deviation from the COARSE control setting. This provides vernier adjustment to easily set the frequency to a precise value.
4. When the 5Hz range is selected, the gate time is 10 seconds and the display is updated once every 10 seconds. The result of a frequency change will not be displayed until 10 seconds later. Adjust the frequency in progressively smaller steps, waiting for the display to update until the desired frequency is obtained.
5. When outputting square waves or when using the TTL output, terminate the cable into 50 Ω to minimize ringing. Also, keep cables as short as possible.

Figure 3. Use of DC OFFSET Control



6. Remember that the output signal swing of the generator is limited to ± 10 volts open circuited or ± 5 volts into 50Ω , and applies to the combined peak-to-peak signal and DC offset. Clipping occurs slightly above these levels. **Fig. 3** illustrates the various operating conditions encountered when using the DC offset. If the desired output signal is large or if a large DC offset is required, an oscilloscope should be used to make sure that the desired signal is obtained without undesirable clipping.

DUTY CYCLE CONTROL

The DUTY CYCLE control can be used to alter the symmetry of the output waveform, to produce waveshapes such as those shown in **Fig. 4**. For a square wave, symmetry variation amounts to changing the duty cycle (ratio of “high” to “low” time), effectively converting the instrument into a pulse generator. For a triangle wave, the result is a ramp, and with a sine wave, a distorted waveshape called a slewed sine is produced. The Model 4011A provides for symmetry variation from 15% to 85%.

1. Select the waveform desired either SINE, SQUARE or TRIANGLE.
2. Engage the DUTY CYCLE switch (14) and adjust the DUTY CYCLE control (10) for the desired waveshape. Clockwise rotation from center results in an increase in square wave duty cycle, and changes the sine and triangle waves as shown in the top waveform of each pair of **Fig. 4**. Counterclockwise rotation results in the bottom waveform in each pair.
3. Varying the duty cycle setting results in a slight change in frequency. Adjust the COARSE and FINE frequency controls as required.




Figure 4. Effects of Symmetry Variation.

TTL/CMOS OUTPUT

The TTL/CMOS output jack provides a fast rise time square wave output. Either a fixed TTL or a variable CMOS output level is available. The output is positive with respect to ground and can be used as an external sync pulse for oscilloscopes or as a variable frequency signal source for exercising logic circuits. Because of the fast rise time of this output, cable length should be minimized to limit ringing and overshoot.

1. Select the desired frequency range and adjust the frequency controls as required. The OUTPUT LEVEL and DC OFFSET controls have no effect on the signal at the TTL/CMOS jack.
2. When the CMOS LEVEL switch (13) is OFF, a TTL signal is output at the TTL/CMOS jack. Select a CMOS signal by engaging the CMOS LEVEL switch and adjust the level of the signal by rotating the CMOS LEVEL control (8).

VOLTAGE CONTROLLED FREQUENCY OPERATION

The Model 4011A can be operated as a voltage-controlled generator by using an external control voltage applied to the VCG INPUT jack. The externally applied voltage will vary the frequency which is preselected by the range switches and the frequency controls. Applying approximately +10 V with the COARSE control at full clockwise decreases the output frequency by about 100 times (a 100:1 ratio).

1. Select the desired frequency range and waveform.
2. Set the starting frequency with the COARSE control. Apply a positive DC voltage to the VCG INPUT jack (9) to decrease the frequency. A voltage from 0 to +10V will cause the frequency to decrease by a factor of 100 if the COARSE frequency control is set at maximum CW rotation. For example, if the starting frequency is 500kHz, applying +10V will change the output frequency to 5kHz.
3. To operate the function generator as a sweep generator, apply a positive going ramp signal to the VCG INPUT jack. As the ramp voltage increases, the frequency decreases. The rate of sweep can be adjusted by varying the frequency of the ramp signal.

- Specific frequencies can be selected by applying a fixed dc voltage to the VCG INPUT jack or the frequencies can be stepped by applying a stepped dc voltage.
- Do not apply more than ± 15 volts (dc or dc + ac peak) to the VCG INPUT jack. Inputs of more than 15 volts will not cause any further shift in the frequency and could cause damage to the generator.


OUTPUT PROTECTION CONSIDERATIONS

Use care when connecting the function generator output to a signal injection point. Excessive voltage at the point of signal injection of the function generator can cause internal damage. Under normal operation, the generator output should never be connected to an external voltage other than low dc values that can be matched with the DC OFFSET control. The Model 4011A is overload protected so that shorting the output, even continuously, will not cause damage. A fuse has been added in series with the OUTPUT jack to help protect the instrument from damage by connection to excessive external voltage.

Damage of this type usually occurs by accidentally connecting the output of the function generator to a voltage in the equipment under test. The following protective measures are strongly recommended:

- The user should understand the equipment under test well enough to identify valid signal injection points (i.e., the base of a transistor, a logic input of a gate, etc.). The voltage at valid signal injection points is rarely high enough to damage the instrument.
- If in doubt about the safety of a signal injection point, measure the voltage present at the intended point of signal injection before connecting the function generator output to that point.
- When applying the main output of the function generator to a circuit point containing a dc level, adjust the DC OFFSET control so that the dc level at the main output matches the circuit voltage.
- Connect the TTL output only to TTL-level circuits. Connect the CMOS output only to CMOS circuits. Measure the Vcc of the circuit under test and adjust the CMOS LEVEL control as instructed in the manual.
- When the function generator is used by students or other inexperienced users, the circuit of **Fig. 5** could be added into your TTL output probe or test clip set. It will protect the TTL output of the generator against external voltages up to ± 20 volts.

Figure 5. Circuit for Protection of TTL Output.



B+K Precision offers a “Guidebook to Function Generators” which describes numerous applications for this instrument, including hook-up details. It also includes a glossary of function generator terminology and an explanation of function generator circuit operation. It may be downloaded off our website at www.bkprecision.com.

MAINTENANCE

WARNING

The following instructions are for use by qualified service personnel only. To avoid electrical shock, do not perform servicing other than contained in the operating instructions unless you are qualified to do so.

Remember that ac line voltage is present on line voltage input circuits any time the instrument is plugged into an ac outlet, even if turned off. Always unplug the function generator before performing service procedures.

FUSE REPLACEMENT

1. Locate the fuse holder on the input line receptacle.
2. Remove the fuse holder and replace the fuse with an equal value fuse

INSTRUMENT REPAIR SERVICE

Because of the specialized skills and test equipment required for instrument repair and calibration, many customers prefer to rely upon B+K PRECISION for this service. We maintain a network of B+K PRECISION authorized service agencies for this purpose. To use this service, even if the instrument is no longer under warranty, follow the instructions given in the WARRANTY SERVICE INSTRUCTIONS portion of this manual. There is a nominal charge for instruments out of warranty.

AC LINE SELECTION

This instrument can operate on 120 or 230 VAC source at 50 or 60 Hz. The internal jumper allows you to select the line voltage. Before connecting the power plug to an AC line outlet, be sure to check that voltage selector plug is set in the correct position corresponding to the line voltage in your location and the fuse rating shown on the rear panel of the unit.

CUSTOMER SUPPORT

1-800-462-9832

B+K Precision offers courteous, professional technical support before and after the sale of their test instruments. The following services are typical of those available from our toll-free telephone number:

- Technical advice on the use of your instrument.
- Technical advice on special applications of your instrument.
- Technical advice on selecting the best instrument for a given task.
- Information on optional accessories for your instrument.
- Information on instrument repair and recalibration services.
- Replacement parts ordering.
- Availability of service publications.

- Information on other B+K Precision instruments.
- Requests for a new B+K Precision catalog.
- The name of your nearest B+K Precision distributor.

Call toll-free 1-800-462-9832

Monday through Thursday, 8:00 A.M. to 5:00 P.M., Friday 8:00 A.M. TO 11:30 A.M.

Pacific Standard Time

(Pacific Daylight Time summer)

Service Information

Warranty Service: Please return the product in the original packaging with proof of purchase to the address below. Clearly state in writing the performance problem and return any leads, probes, connectors and accessories that you are using with the device.

Non-Warranty Service: Return the product in the original packaging to the address below. Clearly state in writing the performance problem and return any leads, probes, connectors and accessories that you are using with the device. Customers not on open account must include payment in the form of a money order or credit card. For the most current repair charges please visit www.bkprecision.com and click on "service/repair".

Return all merchandise to B&K Precision Corp. with pre-paid shipping. The flat-rate repair charge for Non-Warranty Service does not include return shipping. Return shipping to locations in North American is included for Warranty Service. For overnight shipments and non-North American shipping fees please contact B&K Precision Corp.

B&K Precision Corp.

22820 Savi Ranch Parkway Yorba Linda, CA 92887

www.bkprecision.com

714-921-9095

Include with the returned instrument your complete return shipping address, contact name, phone number and description of problem.

Limited Two-Year Warranty

B&K Precision Corp. warrants to the original purchaser that its products and the component parts thereof, will be free from defects in workmanship and materials for a period of two years from date of purchase.

B&K Precision Corp. will, without charge, repair or replace, at its option, defective product or component parts. Returned product must be accompanied by proof of the purchase date in the form of a sales receipt.

To obtain warranty coverage in the U.S.A., this product must be registered by completing a warranty registration form on www.bkprecision.com within fifteen (15) days of purchase.

Exclusions: This warranty does not apply in the event of misuse or abuse of the product or as a result of unauthorized alterations or repairs. The warranty is void if the serial number is altered, defaced or removed.

B&K Precision Corp. shall not be liable for any consequential damages, including without limitation damages resulting from loss of use. Some states do not allow limitations of incidental or consequential damages. So the above limitation or exclusion may not apply to you.

This warranty gives you specific rights and you may have other rights, which vary from state-to-state.

MODEL 4011A

DATE PURCHASED

TEST INSTRUMENT SAFETY

(continued from inside front cover)

6. Some equipment with a two-wire ac power cord, including some with polarized power plugs, is the “hot chassis” type. This includes most recent television receivers and audio equipment. A plastic or wooden cabinet insulates the chassis to protect the customer. When the cabinet is removed for servicing, a serious shock hazard exists if the chassis is touched. Not only does this present a dangerous shock hazard, but damage to test instruments or the equipment under test may result from connecting the ground lead of most test instruments to a “hot chassis”. To test “hot chassis” equipment, always connect an isolation transformer between the ac outlet and the equipment under test. The B&K Precision Model TR-110 or 1604 Isolation Transformer, or Model 1653 or 1655 AC Power Supply is suitable for most applications. To be on the safe side, treat all two-wire ac equipment as “hot chassis” unless you are sure it has an isolated chassis or an earth ground chassis.
7. On test instruments or any equipment with a 3-wire ac power plug, use only a 3-wire outlet. This is a safety feature to keep the housing or other exposed elements at earth ground.
8. Never work alone. Someone should be nearby to render aid if necessary. Training in CPR (cardio -pulmonary resuscitation) first aid is highly recommended.



according to EEC directives and NF EN 45014 norm

Responsible Party		Alternate Manufacturing Site
Manufacturer's Name:	B&K Precision Corporation	B&K Taiwan 0574
Manufacture's Address:	22820 Savi Ranch Pkwy. Yorba Linda, CA 92887-4610 USA	

Declares that the below mentioned product

Product Name:	Function Generator
Part Numbers:	4010A, 4011A, 4012A, 4040A, 4017A

complies with the essential requirements of the following applicable European Directives:

Low Voltage Directive 73/23/EEC (19.02.73) amended by 93/68/EEC (22.07.93)
Electromagnetic Compatibility (EMC) 89/336/EEC (03.05.88) amended by 92/68/EEC (22.07.93)


and conforms with the following product standards:

Safety	EN 61010-1:2001
EMC	EN 61326:1997 + A1:1998 + A2:2001 EN 50081-1 EN 50081-2

This Declaration of Conformity applies to above listed products place on the EU market after:

Date
February 4, 2005

Victor Tolan President



22820 Savi Ranch Parkway • Yorba Linda, CA 92887

© 2023 B+K Precision


v032323

Printed in U.S.A

22820 Savi Ranch Parkway • Yorba Linda, CA 92887-4610



Documents / Resources

	B K PRECISION 4011A 5 MHz Function Generator with Digital Display [pdf] Instructions 4011A 5 MHz Function Generator with Digital Display, 4011A, 5 MHz Function Generator with Digital Display, Digital Display, Display, 5 MHz Function Generator, Function Generator, Genera tor
---	---

References

-  [B&K Precision Corporation](#)