


# BAKER INSTRUMENTS MM2P Max-Min Thermometer Instructions

[Home](#) » [BAKER INSTRUMENTS](#) » BAKER INSTRUMENTS MM2P Max-Min Thermometer Instructions 

BAKER INSTRUMENTS MM2P Max-Min Thermometer



## Contents

- [1 Features](#)
- [2 Specifications](#)
- [3 Instructions](#)
- [4 Customer Support](#)
- [5 Documents / Resources](#)
  - [5.1 References](#)
- [6 Related Posts](#)

## Features

- Pushbutton reset releases limit markers
- Thermometer registers highest and lowest readings from last setting
- Permacolor mercury-free thermometer

A Max/Min thermometer records highest and lowest temperatures. A “U” shaped tube holds red spirit. As the temperature increases, the liquid expands, forcing the red spirit up the maximum scale. When the temperature falls, the liquid contracts, and the red spirit follows it back up the minimum scale. Two black markers record changes in temperature.

## Specifications

- **Temperature Range:** -40 to 120°F (-40 to 50°C)
- **Dimensions:** 8 x 2.19 x 0.75" (203.2 x 55.6 x 19.1mm)

## Instructions

The markers are held in position by a magnet in the back of the case.

The Max/Min temperatures are read at the top point of the markers.

The current temperature can be read at the bottom of each red spirit column.

**Note:** The minimum scale is inverted with the lower temperatures above the high ones. Press the center button to reset the Max/Min readings.

Bubbles in the capillary will disappear if the thermometer is put on a heater.

Should the red spirit column separate, the thermometer must be shaken to unite the red spirit. If one of the black Max/Min markers disappear, press the button during the shaking process, and it will appear again.

Do not mount or place the thermometer in direct sunlight.

For service, repairs or calibration as well as product info on this or any other Baker product, contact Baker Instruments.

## Customer Support




[www.calcert.com](http://www.calcert.com)

[sales@calcert.com](mailto:sales@calcert.com)



## Documents / Resources

 <p>BAKER INSTRUMENTS 1 YEAR WARRANTY</p> <p>Model MM2P Max-Min Thermometer</p> <p>Instruction Manual</p> <p>2008-010-7004 <a href="http://www.calcert.com">www.calcert.com</a> <a href="mailto:sales@calcert.com">sales@calcert.com</a></p>	<p><a href="#">BAKER INSTRUMENTS MM2P Max-Min Thermometer</a> [pdf] Instructions</p> <p>MM2P Max-Min Thermometer, MM2P, Max-Min Thermometer, Min Thermometer, Thermometer</p>
---	---

## References

-  [Calcert](http://www.calcert.com)