



CPAP MAN BYOND ResPlus B-30P Bi-Level PAP with Humidifier and Mask User Manual

[Home](#) » [CPAP MAN](#) » CPAP MAN BYOND ResPlus B-30P Bi-Level PAP with Humidifier and Mask User Manual 

Contents

1	CPAP MAN BYOND ResPlus B-30P Bi-Level PAP with Humidifier and Mask
2	Introduction
2.1	Intended use
2.2	Warnings
2.3	Cautions
2.4	Contraindications
3	Model
4	Package Contents
5	Device Components
6	Device Symbols
7	Device Operation
7.1	First Use
7.2	Daily Use
8	Humidifier
8.1	Humidifier Composition
8.2	Humidifier Connection & Separation with the Host Unit
9	Using the SpO2 Kit
10	Parameter settings
10.1	Main Interface Screens
10.2	PF Curve interface
10.3	Parameter Setting Interface
10.4	System setting interface
10.5	Information setting interface
10.6	Alarm Interface
11	Prompt
12	Cleaning, Disinfection and Maintenances
13	Troubleshooting
14	Specifications
15	Traveling with the Device
16	Service
17	Technical Support
18	Disposal
19	Warranty
20	EMC Requirements
21	Manufacturer:
22	Documents / Resources
23	Related Posts



CPAP MAN BYOND ResPlus B-30P Bi-Level PAP with Humidifier and Mask



Introduction

Intended use

The BEYOND ResPlus B-30P Bi-Level PAP is designed for delivery of positive airway pressure to provide non-invasive ventilation for adult patients with respiratory insufficiency or obstructive sleep apnea (OSA) in home or hospital environment. Rx only: Federal law restricts this device to sale by or on the order of a licensed healthcare practitioner. Please consult with a licensed health care professional prior to using this device. BEYOND Medical manufactures several accessories that are available to make your OSA treatment with this device as convenient and comfortable as possible. To ensure that you receive the safe and effective therapy, only use accessories manufactured by BEYOND Medical with this device.

Warnings

These warnings are meant to caution the user or operator of this device that there is a substantial risk of injury if these warnings are not followed.

These instructions are for reference only. They cannot replace the expert medical guidance from a licensed healthcare professional for the proper use of this device.

This device is not intended to be used to provide life support. The proper operation of this device may be influenced or disturbed if any of the following conditions exist nearby:

- Electromagnetic fields exceeding 3V/m under the EN60601-1-2 test conditions.
- The operation of a high frequency device (diathermy).
- Electric shocks from a defibrillator or a short-wave therapeutic device.
- Radiation (such as, x-ray or CT).
- Electromagnetic field (such as, MRI).

Use only BEYOND Medical circuit accessories. Do not wear the mask for longer than a few minutes while the device is not in operation. Keep the device dry, ensure the tubing isn't tangled. If you notice any damage to the device or experience any unusual performance (such as, harsh sounds or unfamiliar odors), quickly disconnect the power supply, empty the water tank, and stop using the device. Once the device is turned off, contact your BEYOND representative or local distributor. Devices have the potential to allow rebreathing of exhaled air. To reduce the chance of this possibility occurring, please observe the following precautions: – Only use BEYOND Medical circuit accessories. – Do not wear the mask for longer than a few minutes while the device is not in operation. – Do not block or seal the vent holes in the exhalation port. This device is not recommended to be used in combination with oxygen therapy. Sources of medical oxygen must be located more than 1 meter from this

device to avoid the risk of fire. Do not operate this device in the presence of flammable anesthetic mixtures, especially in combination with oxygen. Do not operate this device in the presence of nitrous oxide and oxygen. Keep away from toxic or hazardous steam. Do not use this device if the room temperature is higher than 35(95°F). If the ambient temperature of the room is higher than 35(95°F), the air flow from the device may rise to exceed 40(105°F), which may cause damage to the user's airway. Do not use this device in direct sunlight or near heating equipment. Doing so may increase the air temperature of the output to unsafe levels. To avoid electrical shock, disconnect the power cord before cleaning. Do not immerse the device in fluid of any kind. Contact your healthcare provider if symptoms of sleep apnea recur. Regularly check the power cable and accessories for any damage or wear. Disconnect the power supply before checking the device. When using this device, make sure the mask is positioned above the height of the host unit. Failure to do so may cause any condensed water in the tubing to flow into the user's nose, creating the risk of possible suffocation. Stop using this device if the humidifier is damaged. Do not touch the heater plate after turning the device off. Let the heater plate cool off before touching. The plate will remain hot for several minutes after the power supply has been disconnected. Do not add water hotter than 35(95°F). Do not splash any water into the device when refilling the water tank. Be sure that the device is in a place that cannot be accessed or touched by children, who could be entangled, injured, or strangled by the tube. Appliance coupler is used as isolation means for the device, please do not position the device so that it is difficult to operate the coupler. Do not attempt to modify this device without written authorization from the manufacture. Do not place the device next to any curtains or similar obstruction. Curtains can block the airflow to the device, which may cause this device to overheat. Do not block the air Intake Port. Doing so will interfere with the therapy. The proper placement and positioning of the mask on the face is critical to the consistent operation of this device. The local distributor of this device is available to provide technical support. Do not perform any maintenance on the device while patient is receiving the treatment. All materials that come in contact with the human body have been tested for biocompatibility and meet the requirements of biological compatibility Approved adapters are specified as part of the device.

Cautions

The following precautionary statements indicate conditions or actions that might damage the device or injure the user. Please read this section carefully.

Do not begin wearing the mask until the device has started to run normally. Do not operate the device if the ambient temperature is outside of the operating temperature range (listed under Section 13). If the device has been exposed to temperatures that are above the operating temperature range, return the device to room temperature before use. Do not immerse the device in liquid of any kind. Do not let any liquid enter into the device or into the filter near the air inlet.

The condensate water can damage this device. Make sure that the device returns to room temperature before use.

Be sure that the filter is seated properly in the device to ensure normal operation. Any tar from smoking cigarettes near the device will stop the device from working properly. If any liquid splashes onto the heater plate, disconnect the power supply and do not use it until the plate is completely dried. Take all possible preventive measures to avoid water damage to the device. Only use distilled water in the water tank. If any other type of liquid is put in the water tank, it may damage the humidifier and/or device, possibly endangering the health of the user. Do not exceed the maximum water level mark on the water tank. Do not splash the water into the device when filling the water tank. Do not tilt the device. Keep the device level in order to avoid water flowing back into the device. If water does get inside the device, disconnect the power immediately and stop using it.

Contraindications

If the patient has severe respiratory failure without spontaneous breathing, do not use the device. If any of the following conditions apply, consult the healthcare provider before using the device.

Absolute contraindications: Pneumothorax or mediastinal emphysema; cerebrospinal fluid leak, traumatic brain injury, or pneumocephalus; shock caused by a variety of conditions before treatment; active epistaxis; upper gastrointestinal bleeding before treatment; coma or impaired consciousness that makes using the mask during therapy impossible; or giant vocal fold polyp, etc.

Relative contraindications: Severe coronary heart disease complicated with left ventricular failure, acute otitis media, excessive respiratory secretions and weak cough, weak spontaneous breathing, nasal or oral tracheal intubation and tracheotomy, severe nasal congestion caused by a variety of conditions, lung bullae, and allergies

to the breathing masks, etc.

The following side effects may occur during treatment:

- Dryness of the mouth, nose and throat
- Abdominal bloating
- Ear or sinus discomfort
- Eye irritation
- Skin irritation due to the use of a mask
- Chest discomfort

If you have any symptoms of discomfort, please contact your healthcare provider immediately.


Model

Model	Type	Main components	Working Mode	Max Pressure(cm H2O)
B-30P	Bi-Level	Host Humidifier H20 SpO2 Kit S10 (optional)	CPAP, S, T, S/T, APC V	30

Package Contents

Item	Articles	Quantity	Remark
1	Host	1	Standard
2	Humidifier	1	Standard
3	Tubing	1	Standard
4	Mask	1	Standard
5	Adapter	1	Standard
6	User Manual	1	Standard
7	Air Filter	2	Standard
8	SpO2 Kit	1	Optional
9	Carrying Case	1	Standard
10	TF Card	1	Standard

Device Components



Item	Articles	Quantity	Remark
1	Host	1	Standard
2	Humidifier	1	Standard
3	Tubing	1	Standard
4	Mask	1	Standard
5	Adapter	1	Standard
6	User Manual	1	Standard
7	Air Filter	2	Standard
8	SpO2 Kit	1	Optional
9	Carrying Case	1	Standard
10	TF Card	1	Standard

Do not remove the Dial cap. The metal rod connected to the Dial may come in contact with the external current, creating the risk of an electric shock. If the Dial cap falls off, contact the local distributor to purchase a replacement.

In case of the failure of any of the buttons, contact the local distributor.

Device Symbols

These symbols are provided in the labeling papers to give users safety instructions, please read through their definitions carefully before using.

Symbol	Definition
	Warnings or Caution
	Type BF applied part
	Serial number of the product
	Date of manufacture
	Follow instructions for use
	The authorized EU-representative
	Manufacturer's Information
	CE Mark
IP22	Degree of Protection Against Ingress of Water and Particulates
	Conforms to the Waste Electrical and Electronic Equipment /the Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment
	Heat Warning
Rx Only	Caution: Federal law restricts this device to sale by or on the order of a licensed healthcare practitioner.

Device Operation

First Use

- Place the device on a steady flat table where the settings are easy to reach and the information on the display can be clearly seen by the user. Maintain at least 2 inches between the device and the wall to ensure that the air inlet is not obstructed. When using the device, ensure that the surrounding air flows freely. Keep the device away from any heating and cooling devices (air conditioning vents, radiators, etc)
- Place the air filter in the slot at the Air Inlet port.
- Connect one end of the tubing into the Air Outlet.
- Connect the mask into the other end of the tubing.
- Connect the power adapter into the Power Port on the rear of the device.

- Connect the device to mains power supply. The device will display the Standby Interface 6
- Connect the power supply. Press the On/Off Button to start the device.

The Connections should appear as shown in in Figure 6.1 below:




Figure 6.1

Put the device on a solid flat surface where the approach is not blocked, the displays are easy to see, and the controls are convenient to access. Be sure to place the device where it cannot fall easily. The device needs to be lower than the height of the bed. Do not place the device in or on a container where water may gather. Be sure to take preventive measures in order to avoid water damage to the device. When water is in the humidifier, do not move the device, so as to avoid water entering into the device.

Daily Use

Assembling the Device

Connect the power supply and the tubing as listed for the first use (see above). Connect the mask and the headgear according to the Mask User Manual (provided separately).

Adjust the Tubing

Adjust the tubing to make sure the tubing can move freely when the user is in a deep sleep. Adjust the mask and the headgear to make the user feel as comfortable as possible and prevent air from blowing into the user's eyes.

Start the Device Running

Press the On/Off Button. The device will begin to run and therapy data (such as, air pressure and temperature) will be shown on the display screen.

Humidifier Heating

If using the humidifier, it is recommended that the unit be used with the tubing which provided by BEYOND Medical to prevent leakage. If the humidifier is attached to the device but the patient does not want to use the humidifier, turn the humidifier off.

Ramp Function

When using the Ramp Function, the device will begin at a low initial pressure and increase to the desired pressure setting in a steady rise over the time period selected by the user. After this function is selected, press the On/Off Button once, the device will start to run in accordance with the Ramp Function's settings. If the On/Off Button is pressed a second time, the device will cancel the Ramp Function and run at the selected pressure.

Turn Off the Device

When the mask is taken off, the device will stop running. The user then needs to turn off the device by unplugging the power supply.

Humidifier

In drier regions, prolonged use of the device may dry out the user's nasal passages, causing discomfort. This device is equipped with a humidifier to address this issue. When water is added to the humidifier and it is turned on, moisture will be added to the inhaled air to moisten the user's nasal passages.

Humidifier Composition




Fig.7.1

1	Air Outlet
2	Humidifier Cap Key
3	Water Checking Window
4	Air Inlet
5	Humidifier Connector
6	Humidifier Separation Key

Humidifier Connection & Separation with the Host Unit

Connect with the host unit




Fig.7.2.1a Before the connection

Fig.7.2.1b After the connection

Push the host unit and humidifier together to connect them to each other. A “click” sound will be heard when they are properly connected.

Push the humidifier to connect with the host completely.

Separate from the host unit





Fig.7.2.2

Press the Humidifier Separation Button while simultaneously pulling the humidifier apart from the host unit.

Adding water into the water tank

1. Take out the water tank: move the humidifier cover by sliding the Humidifier Cap Key to the right side as shown in Figure 7.2.3a below. Then, open the humidifier cover to take out the water tank by grasping the tank with thumb and index finger as pictured in Figure 7.2.3b.



Humidifier Cap Key

Fig.7.2.3a




Fig.7.2.3b

2. Add water through the Air Inlet port on top of the water tank. Make sure the water level does not exceed the Max line.




Fig.7.2.3c In winter months, be sure to add warm water, but no hotter than 35(95°F). Do not fill the water tank above the Max line. Turn off the humidifier when the water tank is empty. Turn the device off before refilling water. Do not add water when the humidifier is in operation. Only add water when the humidifier has cooled down. Do not let any water enter the host unit.

Using the SpO2 Kit

The SpO2 Kit consists with a SpO2 Probe, Adapter, and Connector. The SpO2 Kit is intended to be used for continuous, non-invasive functional arterial oxygen saturation (SpO2) and pulse rate monitoring for adults weighing more than 40 kg (90 pounds). The SpO2 Kit is immediately ready to use when it is connected to the main device via the Communications Port. Attach the Oxygen Probe sensor to the patient's index finger (although other fingers can also be used). The sampling frequency of the SpO2 signal is 50Hz, and the update frequency of

the device is 1Hz. The value of SpO2 is calculated by the average of the previous eight pulse waveforms. If the SpO2 Kit does not connect or is not performing correctly, the value of SpO2 will not be displayed on the Main Interface.

Parameter settings

Main Interface Screens

Connect the device properly to mains power supply with power cord and power adapter. The screen displays the Main Interface screen will appear as shown in Fig. 9.1 below.




Fig.9.1: B-30P Main Interface

PF Curve interface

Under main interface, rotate the “Dial” button to move the cursor onto the “PFCurve” indicator and press “Dial” button; the PFCurve Interface screen will appear on the display screen, as shown in Fig. 9.2 below.



Fig.9.2: B-30P PFCurve Interface Screen

Users can set the desired humidifier levels and the Ramp time in the PFCurve Interface; the selectable values ranges are preset on the parameter interface.

When the humidifier is not connected to the device, the Humidifier setting icon is grayed and cannot be accessed. SpO2 and Pulse Rate are only displayed when the SpO2 Kit (optional parts) is correctly attached.

Icon		Meaning
1		The TF card is inserted.
2		The device is working.
3		The Ramp function has been set.
4		The humidifier level has been se.
5		Indicates that the parameters setting interface is locked. When the device is connected to the power, its locked by default; Unlock the parameter interface by pressing the Dial button and holding for 5 seconds while the cursor is located on "Parameter" option. After unlocking successfully, the icon will change to .

Parameter Setting Interface

Under the main interface, rotate the "Dial" button to move the cursor onto the "Parameters" icon, and then press the " Dial " button for 5 seconds to unlock. This will bring you to Parameter Setting Interface screen, as shown in Fig.9.3a below. Next, press the " Dial " button again and the selected font will turn yellow, indicating that the device's parameters can now be set.

Note: as shown in Fig.9.3a through Fig. Fig.9.3d below, indicates more values are presented in the next page; indicates more values are presented in the previous page. When both appear on the same page, it means the user can go either direction to see more values.



Fig.9.3a :B-30P Parameter Setting Interface Screen 1



Parameter

AutoOFF	OFF	
AutoON	OFF	
Mode	ST	
IPAP	10.0	cmH2O
EPAP	4.0	cmH2O
StartPress	4.0	cmH2O
ISens	Auto	
		



Back

Fig.9.3b : Selected Status Screen

Parameter		
ESens	Auto	
BPM	3	
InspTime	0.5	sec
ISlop	1	
MaxInspTime	0.5	sec
MinInspTime	0.5	sec
Ramp	60	min
▼▲		
⋮ Back		

Fig.9.3c : B-30P Parameter Setting Interface Screen 2

Parameter		
Humidifier	5	Lv
VAFSetting	ON	
VT	50	ml
MaxIPAP	30.0	cmH2O
MinIPAP	6.0	cmH2O
▲		
⋮ Back		

Fig.9.3d : B-30P Parameter Setting Interface Screen 3

Parameter	Range	Description	Mode
Auto ON	ON/OFF	<p>When the Auto ON is set to “ON”, the user can wear the mask and take 3 breaths in device's Standby State. The device will then automatically enter into the Working State.</p>	CPAP; S; T; S/T; APCV
Auto OFF	ON/OFF	<p>When the Auto OFF is set to “OFF”, the user needs to remove the mask while the device is in the Working State. Within 15 seconds the device will automatically enter into the Standby State. When this function has been turned ON, the device will automatically enter into the Standby State whenever the user's mask falls off or the tubing becomes disconnected while the user is asleep.</p>	CPAP; S; T; S/T; APCV

* Mode	CPAP; S; T; S/t; APCV	<p>CPAP: provides a constant level of pressure throughout the breathing cycle.</p> <p>S: a Bi-Level mode that responds to both inhalation and exhalation. The pressure increases as the user inhales and decreases as the user exhales. There is no inspiratory pressure if user does not inhale. IPAP (Inspiratory Positive Airway Pressure) and EPAP (Expiratory Positive Airway Pressure) are pre-determined by user's healthcare provider.</p> <p>T: a Bi-Level mode that the device controls the timing of inhalation and exhalation according to the preset parameter.</p> <p>S/T: a Bi-Level mode that responds to both inhalation and exhalation. The pressure increases as the user inhales and decreases as the user exhales. If no inhalation is detected within a set time, the device automatically starts inhalation. When the device starts inhalation, it controls the time of inhalation and automatically decreases the pressure for exhalation within a set time.</p> <p>APCV: a Bi-Level mode based on the S/T – the device controls the time of inhalation and automatically decreases the pressure for exhalation within a set time.</p>	
Press	4-20 cmH ₂ O	This setting sets the output pressure for the device. It can be set in increments of ± 0.5 cmH ₂ O.	CPAP

Start Press	4.0-25.0 cmH2O	<p>CPAP Mode: The initial air pressure for the Ramp Function is equal to or less than the standard air pressure selected by the user.</p> <p>S; T; S/T; APCV Mode: The initial air pressure for the Ramp Function is equal to or less than the EPAP.</p> <p>This parameter can be set in increments of ± 0.5 cmH2O.</p>	S; T; S/T; A PCV
Belex	0-3 level	<p>Setting the Belex function while the device is in CPAP Mode will enable the device to detect the respiratory rhythm of the user. This function will lower the pressure in the mask during exhalations to make the user feel more comfortable.</p>	CPAP
IPAP	4.0-25.0 cmH2O	<p>Inspiratory Positive Airway Pressure.</p> <p>This parameter can be set in increments of ± 0.5 cmH2O.</p>	S; T; S/T; APCV
EPAP	4.0-25.0 cmH2O	<p>Expiratory Positive Airway Pressure.</p> <p>This parameter can be set in increments of ± 0.5 cmH2O.</p>	S; T; S/T; A PCV

ISens	Auto; 1-6 Levels	<p>Sensitivity to inhalation. The device senses the transition when the user enters inhaling phase. It allows the device to synchronize with user's inhalation so that the user can breathe more comfortably.</p> <p>Smaller level indicates higher sensitivity. Auto means the device adjusts its sensitivity according to user's breathing condition.</p>	S; S/T; APC V
ESens	Auto; 1-6 Levels	<p>Sensitivity to exhalation. The device senses the transition when the user enters exhaling phase. It allows the device to synchronize with user's exhalation so that the user can breathe more comfortably.</p> <p>Smaller level indicates higher sensitivity. Auto means the device adjusts its sensitivity according to user's breathing condition.</p>	S; S/T; APC V
ISlop	1-6 Levels	<p>Time that it takes for the pressure to rise from exhaling phase to inhaling phase. The corresponding times to six levels are 100ms, 200ms, 300ms, 400ms, 500ms, and 600ms. Smaller level indicates shorter time.</p>	S; T; S/T; APCV

BPM	3-40	The device outputs an airflow with a set respiratory rate when the user is unable to self- breathe.	T; S/T
Insp Time	0.5-4.0s	The device outputs an airflow with a set respiratory time when the user is unable to self- breathe.	T; S/T
Max InspTime	0.5-4.0s	When the user starts to inhale, the device outputs an airflow with the maximum respiratory time set for IPAP	S; S/T; APC V
Min Insp Time	0.5-4.0s	When the user starts to inhale, the device outputs an airflow with the minimum respiratory time set for IPAP	S; S/T; APC V

Max IPAP	4.0-30.0 cmH ₂ O	<p>Maximum level of pressure applied during the inspiration phase.</p> <p>This parameter can be set in increments of ± 0.5 cmH₂O.</p>	S; S/T; APC V
Ramp	0-60 min	This parameter sets the time it takes for the air pressure to ramp up to the final pressure selected by the user (Max Press)	S; T; S/T; A PCV
Humidifier	0 (OFF) to 5 (Maximum)	<p>This parameter sets the output air humidity level for the device. The higher the value, the higher the humidity.</p> <p>“0” indicates that the humidifier is off.</p> <p>The default value before dispatched from the manufacturer is “0”.</p>	S; T; S/T; A PCV
VAF Settings	ON/OFF	The Volume Assured Function. Allows the user to set target tidal volume, the maximum inspiratory pressure, and the minimum inspiratory pressure.	S; T; S/T; A PCV

VT	50- 2000mL	The device through the automatic pressure adjustment to tidal volume achieves the set value.	S; T; S/T; APCV
Tips	indicates that this parameter can only be adjusted while the parameter interface is unlocked. To unlock the parameter interface, position the cursor over the “Parameter” icon and then press and hold the Dial Button for 5 seconds to unlock.		

System setting interface

Rotate the “Dial” button so that the cursor on the Main Interface screen is pointing at “System”, then press the “Dial” button to enter into the System Setup Interface screen, as shown in Fig.9.4a . .

Fig.9.4a : System Setting Interface Screen 1




Fig.9.4a : System Setting Interface Screen 1





Fig.9.4b : System Setting Interface Screen 2

Fig.9.4b : System Setting Interface Screen 2


Fig.9.4c : System Setting Interface Screen (with Reset Dialogue Box active)



Parameter	Setup range	Description
Time	h-m-s	This parameter sets the device's internal clock to track the time. This time will then be used to record the application information for users. This setting needs to be checked frequently to ensure continued accuracy.
Date	yyyy-mm-dd	This parameter sets the device's internal calendar to track the date. This date will then be used to record the application information for users. This setting needs to be checked frequently to ensure continued accuracy.
Back Light	30-600 sec	This parameter sets the length of time that the LCD display will be back lit. When the user sets a value, the backlight will turn off after the set time elapses.

Brightness	1-5 level	This parameter has a five-level range. The higher the level, the greater the screen brightness.
Language	Chinese- English	Can be switched from English to Chinese. Chinese language User Manuals can be requested from the manufacturer.
Press Unit	cmH2O-hPa	Displayed values can be switched between“cmH2O” and “hPa”, as desired by the user.
Reset	—	System Restore will return all system parameters back to the factory default values.

Information setting interface



Information		
UseCycle	1	Day
UseTime	24.5	hr
AvgPress	10.0	cmH2O
TiPressure	9.5	cmH2O
AvgLeakage	102.5	L/min
AHI	12.5	
UseDays	0	Day
Back		

Fig.9.5a B-30P Information Setting Interface 1



 Back

Fig.9.5b : B-30P Information Setting Interface 2

Parameter	Setup range	Description
Use Cycle	1/7/30/90/180 /365	The time period over which the following information items are calculated. Unit is “day”.
Use Time	—	The length of time that the device has been connected to the power supply. Unit is “hour”.
Avg Press	—	The average output pressure value of the device in the running state within the life cycle. Unit is “cmH ₂ O”.

TiPressure	—	The 95% output pressure value of the device in the running state within the life cycle, measured by usage time. Unit is “cmH2O”
Avg Leakage	—	The average leakage value of the device in the running state within the lifecycle. Unit is “L/min”.
Run Hrs	—	The duration of running time after the device is shipped from factory. This value cannot be cleared.
AHI	—	Apnea-hypopnea index of the device in the using cycle. Calculation method as Apnea- hypopnea frequency per hour
Use Days	—	The device is connected to the power supply and continuously running for 4 hours or more it is counted as 1 day.

Alarm Interface

Rotate the “Dial” button so that the cursor on the Main Interface screen is pointing at the “Alarm”, then press the “Dial” button to enter into the Alarm Interface screen, as shown in Fig9.6.




Fig.9.6 Alarm Interface

Parameter	Setup range	Description
High Press	OFF/8.0-24.0 cmH2O	It sets the upper limit pressure for the device. Increment: ± 0.5 cmH2O.
Low Press	OFF/3.0-7.0 cmH2O	It sets the lower limit pressure for the device. Increment: ± 0.5 cmH2O.
High Leakage	OFF/ON	It turns on or off the leakage prompt function for the device.
LowMV	OFF/1-30L	It sets the lower limit for volume per minute value. Increment: ± 1 L.
Replace Filter	ON/OFF	It turns on or off the filter replacement prompt.

Prompt

Prompt Message	Description
Power Failure	<p>If the power supply is disconnected while the device in the Working State, the device will provide a voice prompt via the buzzer for approximately 30 seconds. Pressing the "On/Off " Button or re-powering the device will stop the voice prompts. After power is returned to the device, it will return to the normal state.</p>
HIGH PRESS!!!	<p>When the device's output pressure exceeds the High Press setting value, the device will provide a voice prompt via the buzzer and "HIGH PRESS!!!" will appear on the display. The warning sounds can be canceled by pressing the Mute button.</p>
LOW PRESS!!	<p>When the device's output pressure is below the Low Press setting value, the device will provide a voice prompt via the buzzer and "LOW PRESS!!" will appear on the display. The warning sounds can be canceled by pressing the Mute button.</p>
LOW MV!!	<p>When the device detects that the MV is below the LowMV setting value, it will provide a voice prompt via the buzzer and "LOW MV!!" will appear on the display. The warning sounds can be canceled by pressing the Mute button.</p>
LOW VOLTAGE!!	<p>When the device detects that the adapter output voltage is below acceptable limits, the device will provide a voice prompt via the buzzer and "LOW VOLTAGE!!" will appear on the display. The warning sounds can be cancelled by pressing the Mute Button.</p>

AIRWAY BLOCK!!	When the air flow of the device is blocked, the device will provide a sound prompt via the buzzer, and "AIRWAY BLOCK!!" will appear on the display. The warning sound can be cancelled by pressing the Mute Button.
HIGH LEAKAGE!!	When there is a leak in the device's breathing tube, the device will provide a sound prompt via the buzzer and "HIGH LEAKAGE!!" will appear on the display. The warning sound can be cancelled by the Mute Button.
HUMI FAILURE!	When the device is connected to the humidifier, the device is powered on, and the humidifier has been heating for 2 minutes, but the humidifier is not working, the device will provide a sound prompt via the buzzer and "HUMI FAILURE!" will appear on the display. The warning sound can be cancelled by pressing the Mute Button.
TF CARD FULL!	When the TF card that has been inserted into the device reaches its capacity limit, the device will provide a sound prompt via the buzzer and "TF CARD FULL!" will appear on the display. The audio prompt for this warning is a single tweet, it will only tweet once.

Cleaning, Disinfection and Maintenances

Timing between Cleanings

Based on hygienic reasons, under normal usage, it is recommended that the operator or user service and / or replace the parts as shown below:

- Clean the device, tube and filter before the device is used for the first time.
- Empty the water tank and clean it daily to prevent bacterial growth.
- Clean the air filter at least once every 2 weeks.
- Replace the air filter with a new filter at least every 3 months.
- Replace the mask with a new one every 3 months.

Cleaning

Cleaning the host and tubing

To avoid the possibility of electrical shocks and damage to the electrical system, unplug the power cord from the host unit before cleaning the device. Clean the front panel and the outside of the case with a soft cloth that has been moistened with warm water or mild detergent. Before plugging in the power cord, make sure that the device is completely dry.

If the device is used by multiple users, the tubing and mask must be replaced between each use by a different

user.

When cleaning the tubing and mask, please refer to the Mask User Manual.

Cleaning the water tank

Remove the water tank, open the upper lid, and clean the internal walls of the humidifier unit. Clean and rinse the water tank thoroughly. Use mild liquid detergent to clean the unit and tank, then rinse with clean water. After cleaning, wipe it clean and allow to dry naturally. Please ensure that the unit seals tightly after each cleaning.

Clean and Replace the air filter

Clean the air filter thoroughly with warm water and mild detergent, then rinse off all detergent residues thoroughly with clean water. Before reinstalling the air filter, let it air dry completely. If air filter is damaged in any way, please replace it with new filter provided by authorized dealer.

1. Disconnect the device from the power source.
2. Remove the air filter enclosure from the device and remove the air filter.
3. Examine the air filter closely to ensure that it is clean and undamaged.
4. Wash the air filter in warm water with mild detergent as instructed above.
5. Reinstall the air filter

Caution:

Never install a wet air filter into the device. Users should alternate between the two air filters included with the device. This will ensure the air filter has the chance to sufficiently dry after each cleaning and before it is used in the device.

Disinfection

If you have followed the cleaning instruction correctly, you should not have to sterilize the device or its components. However, if the water tank or any other component is contaminated or if the device is used in a clinic setting, the device should be sterilized with any sanitizer gel available from a pharmacist. Please note that the use of a sanitizing agent will damage the device's surface material and shorten its service life. Therefore, you should follow the sanitizer manufacturer's instruction for the type of material being cleaned. Always use clean water to completely rinse the device components that come into direct contact with the patient's skin, including the mask, headgear and tubes. Cleaning the material will help to prevent skin and respiratory tract infection that may be caused by any residual sanitizing solution.

Transferring to another patient

If device is to be used by another patient, for sanitation purpose, the device components which come into close contact with this patient's skin, such as the mask, headgear, tube and air filter should be replaced with new ones. Alternatively, users can also follow the procedures listed in Section 11.3 "Disinfection".

Troubleshooting

If the problem can't be eliminated immediately by the above suggestions, please contact your device supplier for repair.

To avoid more injuries or faults, please don't use the device when any of these problems persist.

Opening this device voids the warranty. Only professionals certified by BEYOND Medical are authorized to open the device.

Phenomenon	Possible Causes	Troubleshooting
Nothing displays on the screen or the Main Interface screen does not appear after turning the device on.	The power supply is not connected properly.	Disconnect the power plug and re-connect it.
The device beeps after turning on.	The power supply is not connected properly.	Disconnect the power plug and re-connect it.
The device fails to stop automatically after removing the mask.	“Auto OFF” function is off.	Set the “Auto OFF” function to “ON”.

<p>Nose and/or throat are dry and/or irritated when using this device.</p>	<p>The air is dry in the room.</p>	<p>Turn up the humidifier level or consult your doctor.</p>
<p>Dry mouth and throat.</p>	<p>Likely because the patient sleeps with his or her mouth opened so that the pressurized air goes through the mouth, which leads to dryness of the nasal passages and throat.</p>	<p>The setting pressure may be too low.</p> <p>You may also want to consult your doctor .</p>
<p>Eye irritation or dryness.</p>	<p>The model or size of the mask may not be suitable. The mask may also be positioned incorrectly causing air to blow into the patient's eyes.</p>	<p>Adjust the mask position and the headgear's tightness. Consult your doctor to replace the mask.</p> <p>If the mask is old or broken, replace immediately.</p> <p>Try another mask model.</p>

Face reddening or inflamed.	The headgear is too tight. The mask model or size may not be suitable. The patient may be allergic to the materials of the mask.	Loosen the headgear and insure proper fit. Consult your doctor.
Water in the mask	Room temperature is too low or (if the humidifier is being	Lower the humidifier setting, increase the room temperature, or
	used) output air can condense in the tube and collect in the mask.	place a towel or blanket over the tube to maintain the temperature of the output air flow.
Nasal, sinus, or ear pain	Inflammation in sinus or middle ear.	Stop using the device immediately and contact your doctor.

Obstructive Sleep Apnea- Hypopnea Syndrome recrudescence (e.g.: daytime sleepiness)	The required treatment pressure may have changed due to your weight, nasal obstruction, drinking or some other reasons.	Consult your doctor.
Air output is too hot.	The air filter is either clogged with dirt or the air inlet is otherwise blocked. The device might also be too close to wall, curtains or other obstructions to the air flow around the device.	Check and clean the air inlet, replace if necessary. Reposition the device to a place that is at least 2 inches away from the wall, curtain or other obstructions.
No airflow output	Device malfunction	Contact your device supplier.

<p>The air flow output is too low</p>	<p>If the “Ramp” function is on, it takes time for air flow to rise from the initial pressure to the treatment pressure. The air inlet may also be blocked.</p>	<p>Turn off or change the settings of the Ramp feature.</p> <p>Check and clean the air inlet or replace the air filter.</p>
<p>The blower is always at an abnormally high rotation rate.</p>	<p>The device is leaking air.</p>	<p>Contact your device supplier.</p>
<p>The device doesn’t work when it is turned on.</p>	<p>Device malfunction.</p>	<p>Contact your device supplier</p>
<p>The device works but the pressure in the mask is obviously different from the setting pressure.</p>	<p>Tubes are either leaking or the device is malfunctioning.</p>	<p>Check to see if the tubing is connected properly. If problem persists, contact your device supplier.</p>

The device can only output air at low pressure.	The air filter or air inlet is blocked, or the treatment pressure has been set too low. If the Ramp function is on, it takes time for air flow to rise from the initial	Replace the air filter and clean the air inlet. You may also need to turn off or change the settings of the Ramp feature.
	pressure to the treatment pressure.	
The device is excessively loud.	The tube isn't connected properly. The mask or tubing may also be leaking.	Reconnect the tubing and ensure there are no leaks in the tubing.
The pressure can't be set.	The Ramp feature is on.	Turn off the Ramp feature, then re-set the pressure.

If the problem can't be eliminated immediately by the above suggestions, please contact your device supplier for repair.

To avoid more injuries or faults, please don't use the device when any of these problems persist.

Opening this device voids the warranty. Only professionals certified by BEYOND Medical are authorized to open the device.

Specifications

Environmental

	Operating	Storage
Temperature	5°C to 35°C (41°F to 95°F)	-20°C to 55°C (-13°F to 158°F)
Humidity	15% to 93% (no condensation)	15% to 93% (no condensation)
Atmospheric Pressure	700 to 1060 hPa	700 to 1060 hPa

Physical

	Dimensions	165mm*200mm*120mm, 280mm*200mm*140mm (with humidifier)
	Weight	1.2Kg, 2.2Kg (with humidifier)

Electrical

Power adapter	Input: AC100-240V, 50/60Hz, 1.8Amax Output: DC24V, 3.75A
Type of protection Against Electric Shock	Class II Equipment
Degree of protection Against Electric Shock	Type BF applied part
Degree of protection Against Ingress of Water	IP22
Mode of operation	Continuous
Applied part	SpO2 Probe, Tubing, Mask

Noise level

The A-weighted sound pressure level does not exceed 30dBA, the A-weighted sound power level does not exceed 38dBA, when the device is working at the pressure of 10 cmH₂O.

Pressure Accuracy

According to the pressure accuracy of ISO 80601-2-70:2015 standard.

Pressure Range:

CPAP: 4 to 20 cmH₂O (in 0.5 cmH₂O increments), 30 cmH₂O under single fault conditions. Bi-Level PAP: IPAP: 4 to 30 cmH₂O (in 0.5 cmH₂O increments) EPAP: 4 to 25 cmH₂O (in 0.5 cmH₂O increments). The maximum value 40 cmH₂O under single fault conditions.

Pressure display accuracy:

± (0.5 cmH₂O +4% of the actual reading)

Pressure stability:**CPAP**

Static	Dynamic (4 to 20 cmH ₂ O)
±0.5cmH ₂ O	≤ 2cmH ₂ O

Bi-Level PAP

Static	Dynamic (4 to 30 cmH ₂ O)
±0.5cmH ₂ O	≤ 2cmH ₂ O

Maximum Flow:

According to the maximum flow of ISO 80601-2-70:2015 Standard

CPAP

	Test pressure (cmH ₂ O)				
	4	8	12	16	20
Average flow at the patient connection port (l/min)	100	120	120	120	120

Bi-Level PAP

	Test pressure (cmH ₂ O)				
	4	9	15	20	25
Average flow at the patient connection port (l/min)	100	120	120	120	120

Note: all test data were carried out under conditions with the humidifier and 22mm tube.

Ramp: The ramp time ranges from 0 to 60 minutes.

SpO2 Range: 0 100%.

The margin of error for SpO2 between 70% and 100% is $\pm 3\%$. No strict accuracy requirements for SpO2 below 70%.

Pulse Rate

Range: 40 240 BPM

Margin of Error: $\pm 3\%$

Tubing

Length: 1.8m

The Form and the Dimensions of the Patient Connection Port

The 22 mm conical air outlet complies with ISO 5356-1.

Filter


Dimension 45mm*33mm*10mm

Humidifier

Tested according to EN ISO 8185-2009 standard or equivalent methods.

Output air flow temperature:	40°C
Humidity output	Not less than 10mg H ₂ O/L Measured conditions: Max flow, 95°F, 15% relative humidity.
Pressure drop caused by humidifier	1cmH ₂ O (with the flow rate of 60 LPM)
Leaking under maximum working pressure:	<25mL/min (Together with the tubing)
Adaptability	20mL/kPa (Together with the tubing)
Capacity	300ml

Pneumatic Diagram



Traveling with the Device

1. Only use the BEYOND Medical carrying case when transporting the device and accessories. The device should never be placed in checked baggage.
2. This device operates on power supplies of between 100 – 240 V and 50 / 60 Hz. It can be used in any country in the world, no special adjustments are necessary. However, you will need to find out what type of power sockets are standard in your destination. Power socket adaptors are available separately anywhere you can purchase general electronics.
3. **Security Check Points:** For convenience at security check points, there is a note on the bottom of the device stating that it is medical equipment. It may be helpful to bring this manual along with you to help security personnel better understand the device.

Empty the water tank of the humidifier and allow it to dry completely before packing the device for your trip. This will prevent any remaining water from entering into the device.

Be sure to use the device at the proper elevation setting, failure to do so could result in airflow pressures different than the prescribed setting. Always verify both your actual elevation and the device's elevation setting when using the device in a different location.

If the device is used when the atmospheric pressure is outside of the stated range (See Section 9), the accuracy of the leakage alert will be negatively affected.

Service

The device does not require any routine servicing.

If you notice any unexplained changes in the performance of the device, if it is making unusual or harsh sounds, if it has been dropped or mishandled, if the enclosure is broken, or if water has entered the enclosure, you must discontinue use and contact your home healthcare provider.

If the device malfunctions, contact your home healthcare provider immediately. Never attempt to open the enclosure of the device. Repairs and adjustments must only be performed by service personnel that have been authorized by BEYOND Medical. Unauthorized service could cause injury, invalidate the warranty, and/or result in costly damage.

If necessary, contact your local authorized dealer or BEYOND Medical for technical support and documents.

Technical Support

Please contact BEYOND Medical directly if you need the circuit diagram of the device or the list of components for certain purposes (such as, maintenance or connection to other equipment). BEYOND Medical will provide the circuit diagram and/or other technical documents in whole or in part according to your needs.

Disposal

When the device reaches the end of its service life, please dispose of the device and its packaging in accordance with local laws and regulations.

Warranty

From the date of purchase, the manufacturer provides a one-year limited warranty for the host unit and a 3-month limited warranty for the tubing, mask and humidifier. This limited warranty is only available to the initial consumer. It is not transferable. For one year (or 3-month) from the original date of purchase, BEYOND Medical will repair or replace any parts of this appliance that prove to be defective in materials or workmanship when such appliance is installed, used and maintained in accordance with the provided instructions.

Exclusions This warranty does not cover:

1. Product with original serial numbers that have been removed, altered or cannot be readily determined;
2. Any damage caused as a result of improper use, abuse, excessive usage, modification or alteration of the device;
3. Repairs carried out by any service organization that has not been expressly authorized by BEYOND Medical to perform such repairs;
4. Any damage or contamination due to cigarette, pipe, cigar or other smoke;
5. Any damage caused by exposure to ozone, activated oxygen or other gases;
6. Any damage or contamination due to insect infestation; and
7. Any damages caused by external causes such as accidents, fires, or acts of God. Considering the life of components and safety, medical devices should not be used longer than 5 years. Expired products should be discarded according to corresponding local laws and regulations.

DISCLAIMER OF IMPLIED WARRANTIES; LIMITATION OF REMEDIES

CUSTOMER'S SOLE AND EXCLUSIVE REMEDY UNDER THIS LIMITED WARRANTY SHALL BE PRODUCT REPAIR OR REPLACEMENT AS PROVIDED HEREIN. CLAIMS BASED ON IMPLIED WARRANTIES, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, ARE LIMITED TO ONE YEAR OR THE SHORTEST PERIOD ALLOWED BY LAW, BUT NOT LESS THAN ONE YEAR. BEYOND MEDICAL SHALL NOT BE LIABLE FOR CONSEQUENTIAL OR INCIDENTAL DAMAGES SUCH AS PROPERTY DAMAGE AND INCIDENTAL EXPENSES RESULTING FROM ANY BREACH OF THIS WRITTEN LIMITED WARRANTY OR ANY IMPLIED WARRANTY. SOME STATES DO NOT ALLOW THE EXCLUSION OR LIMITATION OF INCIDENTAL OR CONSEQUENTIAL DAMAGES, OR LIMITATIONS ON THE DURATION OF IMPLIED WARRANTIES, SO THESE LIMITATIONS OR EXCLUSIONS MAY NOT APPLY TO YOU. THIS WRITTEN WARRANTY GIVES YOU SPECIFIC LEGAL RIGHTS. YOU MAY ALSO HAVE OTHER RIGHTS THAT VARY FROM STATE TO STATE.

EMC Requirements

Guidance and Manufacturer's Declaration: Electromagnetic Emissions

This device is intended for use in the electromagnetic environment specified below. The user of the device should ensure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic environment-guidance
RF emissions CISPR11	Group 1	The device uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR11	Class B	This device is suitable for use in all establishments including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuation /flicker emissions IEC61000-3-3	Complies	

Guidance and Manufacturer's Declaration – Electromagnetic Immunity

This device is intended for use in the electromagnetic environment specified below. The user of the device should make sure that it is used in such an environment.

Immunity test	IEC60601 test level	Compliance level	Electromagnetic environment-guidance
Electrostatic discharge (ESD) IEC61000-4-2	±6KV contact ±8KV air	±6KV contact ±8KV air	Floor should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient /pulse IEC61000-4-4	±2KV for power supply lines ±1KV for input/output line	±2KV for Main power line ±1KV for input/output lines	Mains power quality should be that of a typical home or hospital environment
Surge IEC61000-4-5	±1KV differential mode ±2KV common mode	±1KV differential mode ±2KV common mode	Mains power quality should be that of a typical home or hospital environment

<p>Voltage dips, short interruptions and voltage variations of the input power IEC61000-4-11</p>	<p><5% UT(>95% dip in UT),for 0.5 cycle 40% UT(60% dip in UT),for 5cycles 70% UT(30% dip in UT)for 25 cycles <5% UT(>95% dip in UT) for 5 s</p>	<p><5% UT(>95% dip in UT),for 0.5 cycle 40% UT(60% dip in UT),for 5 cycles 70%UT(30% dip in UT) for 25 cycles <5% UT(>95% dip in UT) for 5 s</p>	<p>Mains power quality should be that of a typical commercial or hospital environment. If the user of the device requires continued operation during power mains interruptions, it is recommended that the device be powered from an uninterruptible power supply or from a battery.</p>
<p>Power frequency (50/60Hz) magnetic field IEC61000-4-8</p>	<p>3A/m</p>	<p>3A/m</p>	<p>Power frequency magnetic fields should be at levels characteristic of a typical location in a typical hospital or home environment.</p>

Note: UT is the a.c mains voltage prior to application of the test level.

Guidance and Manufacturer’s Declaration – Electromagnetic Immunity

This device is intended for use in the electromagnetic environment specified below. The user of the device should make sure that it is used in such an environment.

Immunity test	IEC60601 test level	Compliance level	Electromagnetic environment- guidance
----------------------	--------------------------------	-------------------------	--

Conducted RF IEC 61000-4-6	3Vrms 150kHz to 80MHz	3Vrms 3V/m	Portable and mobile RF communications equipment should be used further away from any part of the device (including cables) than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Radiated RF IEC61000-4-3	3V/m 80MHz to 2.5GHz		Recommended separation distance $d = 1.2 \sqrt{P}$ 80MHz to 800MHz $d = 2.3 \sqrt{P}$ 800MHz to 2.5GHz is the maximum normal output power of the transmitter, its unit is Watt (W) and d is the recommended separation distance, its unit is meter (m).

			<p>Measured magnetic field strengths from a fixed RF transmitter should be less than the compliance level in each frequency range b.</p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p>
--	--	--	---

- Note 1:** At 80 MHz and 800 MHz, the higher frequency range applied.
- Note 2:** These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection of the structures, objects and people.
- Field strengths from fixed transmitters, such as base stations for radio (cellular / cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the device is used exceeds the applicable RF compliance level above, the device should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the device.
 - Over the frequency range 150 kHz to 80 MHz, the field strengths should be less than 3 V/m.

Recommended separation distances between portable and mobile RF communications equipment and the device:

The device is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the device as

recommended below, according to the maximum output power of the communications equipment.

Rate d	Separation Distances According to Frequency of Transmitter			
	maximum			
output	150 kHz 80 MHz	80 MHz 800 MHz	800 MHz 2.5 GHz	
of transmitter (W)	d = 1.2	d = 1.2	d = 2.3	
0.01	0.12	0.12	0.23	
0.1	0.38	0.38	0.73	
1	1.2	1.2	2.3	
10	3.8	3.8	7.3	
100	12	12	23	
<p>For transmitters rated at a maximum output power not listed above, the recommended separation distance (d) in meters(m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts(W) according to the transmitter manufacturer.</p> <p>Note 1: At 80 MHz and 800 MHz, the higher frequency range applied.</p> <p>Note 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection of the structures, objects and people.</p>				

Manufacturer:

Hunan Beyond Medical Technology Co.,Ltd. Contact Number: +86-731-82564299
Address: BEYOND ZONE, LIJIACUN RD, XUESHI STREET, CHANGSHA, HUNAN 410208, CHINA.

U.S Agent:

United Source LLC
Email: BEYOND@united-source.com
Address: 1521 Concord Pike, Suite 301, Wilmington, DE, 19803, USA

Version No.: U.S./20201218/A1

Documents / Resources

