



## euromex RF.6190 Handheld Analog Refracto Meter User Manual

[Home](#) » [euromex](#) » euromex RF.6190 Handheld Analog Refracto Meter User Manual 

*euromex RF.6190 Handheld Analog Refracto  
Meter User Manual*

# Analog refractometer

RF.6190 - RF.6510 - RF.6520 - RF.6532 - RF.6562 - RF.6580 - RF.6582  
RF.6592 - RF.6635 - RF.6642 - RF.6644 - RF.6627 - RF.6610 - RF.6628



## Introduction

With your purchase of a Euronext hand refractometer you have chosen for a quality product. The Euronext hand refractometers are developed for use in laboratories and in the food industry.

The maintenance requirement is limited when using the refractometer in a decent manner.


This manual describes the construction of the refractometer, how to use the refractometer and maintenance of the refractometer.

## Contents

- 1 General safety instructions
- 2 Construction of the refractometer
- 3 Functions of the refractometer
  - 3.1 Standard accessories
- 4 Working with the refractometer
  - 4.1 Light entrance
  - 4.2 Actual measurement
  - 4.3 Temperature compensation
- 5 Maintenance and cleaning
- 6 Documents / Resources
- 7 Related Posts

## General safety instructions

- This product is a high quality optical instrument. Delicate handling is required
- Impacts, even small ones, can affect the precision of the device
- Keep the device and its optics clean for maximum performance
- Precautions should be taken with the samples; substances under observation may be a risk to the health of humans and other living organisms or the environment.



## Construction of the refractometer

The names of the parts are listed below and are indicated in the picture above.

|   |       |   |  |
|---|-------|---|--|
| A | Prism | C | Adjustment screw (underneath protective cap) |
| B | Cover | D | Adjustable eyepiece                          |

## Functions of the refractometer

The Euronext hand refractometers are widely used for measuring sugar concentrations, in the table on the next page the different models are shown with their specific way of calibrating them.

| Model   | Type          | Range        | Accuracy    | Calibration     |
|---------|---------------|--------------|-------------|-----------------|
| RF.6190 | Universal     | 0 - 90 Brix  | 0.2         | testpiece 78,8% |
| RF.6510 | High contrast | 0 - 10 Brix  | 0.1         | distilled water |
| RF.6520 | High contrast | 0 - 20 Brix  | 0.1         | distilled water |
| RF.6532 | High contrast | 0 - 32 Brix  | 0.2         | distilled water |
| RF.6562 | High contrast | 28 - 62 Brix | 0.2         | testpiece 78,8% |
| RF.6580 | High contrast | 0 - 80 Brix  | 0.5         | distilled water |
| RF.6582 | High contrast | 40 - 82 Brix | 0.5         | testpiece 78,8% |
| RF.6592 | High contrast | 58 - 92 Brix | 0.2         | testpiece 78,8% |
| RF.6635 | High contrast | multiple     | 0.2/1       | distilled water |
| RF.6642 | High contrast | multiple     | 0.1         | testpiece 19,6% |
| RF.6644 | High contrast | multiple     | 0.5         | testpiece 78,8% |
| RF.6627 | High contrast | multiple     | 0.2         | distilled water |
| RF.6610 | High contrast | multiple     | 0.005/0.1/1 | distilled water |
| RF.6628 | High contrast | multiple     | 0.2         | distilled water |

#### Standard accessories

- For all types: carrying case, screwdriver or alien key for scale adjustment, plastic pipette
- For RF 190 only: temperature correction thermometer
- For RF.6190, RF.6562, RF.6582, RF.6592, RF.6642 and RF.6644: test piece 19,6% or 78,8% and dispersion fluid.

### Working with the refractometer

Prior to the actual measurement, the scale should be checked if it is correct or not (calibrated). If not, the scale should be adjusted by using the supplied tool to turn the scale adjustment screw (C). For this check, two different standard specimens for each type of hand refractometer are used. These are distilled water and a test-piece, as suggested in the table above.

#### A) Distilled water


Open the prism cover (B). and put a drop of distilled water on the surface of prism (A). Close the prism cover and peep through the eyepiece (D). You will see the horizontal demarcation line as well as the scale in the field of view. If the scale is correctly calibrated the horizontal demarcation line should be exactly on the 0% position of the scale. If not, one can adjust the scale with the screwdriver until the demarcation line is at the scale's 0% position.

#### B) Test-piece


For RF.6190, RF.6562, RF.6582, RF.6592, RF.6642 and RF.6644 a standard test-piece is included to calibrate the scale. Put a small drop of the supplied dispersion fluid on the prism and put the test piece into this drop, close the cover gently. The value should read 19.6% or 78.8% depending on the model.

#### Light entrance

Usually daylight is good enough for all types of hand refractometers but depending of the kind of sample, one can also direct the refractometer to a bright light source.



*High contrast type*



*Universal type*

**Note:** Only for the RF.6190, a special daylight window for opaque specimens is provided. Usually it should be kept closed

### Actual measurement

Wipe off the distilled water, or dispersion fluid. Put a few drops of the specimen on the prism. Close the prism cover and the percentage of the measurement can be read at the position of the demarcation line

### Temperature compensation

All types have an Automatic Temperature Compensation system (ATC), except for the RF.6190 universal type. With this type, if the temperature at the measurement is not 20°C but higher or lower, the reading should be compensated in accordance with the compensating table below. For example, when the reading is 20 at a temperature of 28°C. the compensated percentage is  $20\% + 0.62\% = 20.62\%$ .

| %  | 5  | 10   | 15   | 20   | 25   | 30   | 35   | 40   | 45   | 50   | 55   | 60   | 65   | 70   |
|----|--|------|------|------|------|------|------|------|------|------|------|------|------|------|
| °C | (-) reduce figure from the read figure (%) |      |      |      |      |      |      |      |      |      |      |      |      |      |
| 15 | 0.29                                       | 0.31 | 0.33 | 0.34 | 0.34 | 0.35 | 0.36 | 0.37 | 0.37 | 0.38 | 0.39 | 0.39 | 0.40 | 0.40 |
| 16 | 0.24                                       | 0.25 | 0.26 | 0.27 | 0.28 | 0.28 | 0.29 | 0.30 | 0.30 | 0.30 | 0.31 | 0.31 | 0.32 | 0.32 |
| 17 | 0.18                                       | 0.19 | 0.20 | 0.21 | 0.21 | 0.21 | 0.22 | 0.22 | 0.23 | 0.23 | 0.23 | 0.23 | 0.24 | 0.24 |
| 19 | 0.06                                       | 0.06 | 0.07 | 0.07 | 0.07 | 0.07 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 |
|    | (+) add figure to the read figure (%)      |      |      |      |      |      |      |      |      |      |      |      |      |      |
| 21 | 0.07                                       | 0.07 | 0.07 | 0.07 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 | 0.08 |
| 22 | 0.13                                       | 0.14 | 0.14 | 0.15 | 0.15 | 0.15 | 0.15 | 0.15 | 0.16 | 0.16 | 0.16 | 0.16 | 0.16 | 0.16 |
| 23 | 0.20                                       | 0.21 | 0.22 | 0.22 | 0.23 | 0.23 | 0.23 | 0.23 | 0.24 | 0.24 | 0.24 | 0.24 | 0.24 | 0.24 |
| 24 | 0.27                                       | 0.28 | 0.29 | 0.30 | 0.30 | 0.31 | 0.31 | 0.31 | 0.31 | 0.31 | 0.32 | 0.32 | 0.32 | 0.32 |
| 25 | 0.35                                       | 0.36 | 0.36 | 0.38 | 0.38 | 0.38 | 0.40 | 0.40 | 0.40 | 0.40 | 0.40 | 0.40 | 0.40 | 0.40 |
| 26 | 0.42                                       | 0.43 | 0.44 | 0.45 | 0.46 | 0.47 | 0.48 | 0.48 | 0.48 | 0.48 | 0.48 | 0.48 | 0.48 | 0.48 |
| 27 | 0.50                                       | 0.52 | 0.53 | 0.54 | 0.55 | 0.55 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 | 0.56 |
| 28 | 0.57                                       | 0.60 | 0.61 | 0.62 | 0.63 | 0.63 | 0.64 | 0.64 | 0.64 | 0.64 | 0.64 | 0.64 | 0.64 | 0.64 |
| 29 | 0.66                                       | 0.68 | 0.69 | 0.71 | 0.72 | 0.72 | 0.72 | 0.73 | 0.73 | 0.73 | 0.73 | 0.73 | 0.73 | 0.73 |
| 30 | 0.74                                       | 0.77 | 0.78 | 0.79 | 0.80 | 0.80 | 0.81 | 0.81 | 0.81 | 0.81 | 0.81 | 0.81 | 0.81 | 0.81 |

### Maintenance and cleaning



Always clean the prism of the refractometer after use with a soft tissue



**Warning**

Cleaning cloths containing plastic fibres can damage the coating of the prism!

Euromex Microscopen bv • Papenkamp 20 • 6836 BD Arnhem • The Netherlands  
T +31 (0) 26 323 22 11 • [info@euromex.com](mailto:info@euromex.com) • [www.euromex.com](http://www.euromex.com)




*All information may be changed without prior notice v.223371*

[info@euromex.com](mailto:info@euromex.com)

[www.euromex.com](http://www.euromex.com)

## Documents / Resources

|  |  |
|--|--|
| <p>Analog refractometer<br/>RF.6190 RF.6510 RF.6520 RF.6530 RF.6540 RF.6550<br/>RF.6560 RF.6570 RF.6580 RF.6590 RF.6600 RF.6610</p>  <p>euromex user manual</p> | <p><a href="#">euromex RF.6190 Handheld Analog Refractometer</a> [pdf] User Manual<br/>RF.6190, RF.6510, Handheld Analog Refractometer</p> |
|--|--|

Manuals+.