



# LAUPER INSTRUMENTS Sensit HXG-3 Combustible Gas Leak Detector Instruction Manual

[Home](#) » [Lauper Instruments](#) » LAUPER INSTRUMENTS Sensit HXG-3 Combustible Gas Leak Detector Instruction Manual 

## Contents

- [1 LAUPER INSTRUMENTS Sensit HXG-3 Combustible Gas Leak Detector](#)
- [2 Product Information](#)
- [3 Product Usage Instructions](#)
- [4 QUICK-START INSTRUCTIONS](#)
- [5 Documents / Resources](#)
  - [5.1 References](#)
- [6 Related Posts](#)



**LAUPER INSTRUMENTS Sensit HXG-3 Combustible Gas Leak Detector**



## Product Information

The product is a gas detection instrument designed to detect gases in the environment. It has a gooseneck with a PPM/LEL sensor and cap at the tip. The instrument operates on batteries and has a warm-up sequence of 40 to 180 seconds. The product is intrinsically safe, and service must be performed by factory authorized technicians with approved replacement parts only. The product is made in the USA with globally sourced components.

## Product Usage Instructions

1. Install the batteries by pushing down the locking tab and sliding the handle away from the top of the instrument. When replacing handle be sure the tab is securely in place.
2. Carefully observe polarity when changing batteries. Incorrect installation can damage an internal factory service-only fuse.
3. Allow the unit to go through the warm-up sequence in clean air.
4. If FAIL is displayed for any of the sensor readings, make sure the instrument is in clean air; push and hold the C button until AUTO ZERO is displayed. If this process does not clear the fail on the display, this could indicate a problem with the instrument or sensor.
5. Extend the gooseneck (the PPM/LEL sensor and cap is at the tip).
6. You are now ready to use the instrument. Enter the area and detect gases.
7. If the source of an odor needs to be located, rotate the thumb wheel to be heard at a comfortable tick rate. As the instrument is moved closer to a combustible source, the tick rate will increase. Rotate the thumb wheel back to the comfortable rate, and continue the investigation until the source is located.
8. Press and hold the C button to zero the instrument (only in a gas-free area).
9. When your investigation is complete, roll the thumb wheel back down and click it to the off position. Push button A and hold for 5 seconds until the instrument displays POWER OFF then release to shut off.

## Warnings

- To prevent ignition of flammable or combustible atmospheres, disconnect power before servicing.
- To reduce the risk of ignition of a flammable atmosphere, batteries must only be changed in an area known to be nonflammable.
- Do not mix batteries of different age or type.
- ONLY zero instrument in a gas-free environment.

## Important

- When replacing the gooseneck into the clip on the right side, wrap the gooseneck in a wide circular manner counterclockwise around the back of the instrument. Bending in the oppositedirection may cause damage over the life of the product.

## Hint

- As you approach the source of the leak, keep dialing down the rate of the audible tick by dialing down the thumb wheel located on the right side of the instrument.

## QUICK-START INSTRUCTIONS

1. Install the batteries by pushing down the locking tab and sliding the handle away from the top of the instrument. When replacing handle be sure the tab is securely in place.
2. IMPORTANT: CAREFULLY OBSERVE POLARITY WHEN CHANGING BATTERIES. Incorrect installation can damage an internal “factory service only” fuse.
3. Locate button “A” – push & hold until the unit powers up, then release the power button.
4. Allow unit to go through the warm up sequence in clean air. At the end of warm up, the unit will auto zero and enter the working display. This requires from 40 up to 180 seconds.
5. If FAIL is displayed for any of the sensor readings, make sure the instrument is in clean air; push and hold the “C” button until AUTO ZERO is displayed. If this process does not clear the fail on the display, this could indicate a problem with the instrument or sensor.
6. Look at the display – PPM and LEL readings are displayed
7. Extend the goose neck (the PPM/LEL sensor and cap is at the tip).
8. You are now ready to use the instrument. You can now enter the area and detect gases.
9. Once the environment is determined to be safe to work in, if the source of an odor needs to be located, the thumb wheel can be rotated to be heard at a comfortable tick rate. Once an investigation is begun, as the instrument is moved closer to a combustible source, the tick rate will increase. The thumb wheel can then be rotated back to the comfortable rate and the investigation continued until the source is located.
10. Press and hold the “C” button to zero the instrument. (only in a gas free area).
11. When your investigation is complete, roll the thumb wheel back down and click it to the off position, push button “A” and hold for 5 seconds until the instrument displays “POWER OFF” then release to shut off



**HINT:** As you approach the source of the leak, keep dialing down the rate of the audible “tick” by dialing down the thumb wheel which is located on the right side of the instrument.

**IMPORTANT:** When replacing the gooseneck into the clip on the right side, wrap the gooseneck in a wide circular manner counter-clockwise around the back of the instrument. Bending in the opposite direction may cause damage over the life of the product.

**WARNINGS:**

To prevent ignition of flammable or combustible atmospheres, disconnect power before servicing. To reduce the risk of ignition of a flammable atmosphere, batteries must only be changed in an area known to be nonflammable. Do not mix batteries of different age or type. ONLY zero instrument in a gas free environment. To maintain intrinsic safety, service must be performed by factory authorized technicians with approved replacement parts only.

**Distributed by**

Lauper Instruments AG

Irisweg 16B

CH-3280 Murten

Tel. +41 26 672 30 50


[info@lauper-instruments.ch](mailto:info@lauper-instruments.ch)

[www.lauper-instruments.ch](http://www.lauper-instruments.ch)



This is not a substitute for the instruction manual. It is for reference only. Read and understand the instruction manual before use.

**Documents / Resources**

	<p><b><a href="#">LAUPER INSTRUMENTS Sensit HXG-3 Combustible Gas Leak Detector</a></b> [pdf] Instruction M anual</p> <p>Sensit HXG-3 Combustible Gas Leak Detector, Sensit HXG-3, Combustible Gas Leak Detector, Gas Leak Detector, Leak Detector, Detector</p>
---	--

References

-  [suche.ch](#)
-  [Home | Lauper Instruments](#)