



Adam Equipment

# LBX C



Easy Reference:

Model name of the scale:	
Serial number of the unit:	
Software revision number (Displayed when power is first turned on):	
Date of Purchase:	
Name of the supplier and place:	

## CONTENTS

P.N. 3.02.6.6.15555, Revision A2, September 2020

<b>1.0</b>	<b>INTRODUCTION .....</b>	<b>4</b>
<b>2.0</b>	<b>SPECIFICATIONS.....</b>	<b>5</b>
<b>3.0</b>	<b>SAFETY PRECAUTIONS.....</b>	<b>5</b>
<b>4.0</b>	<b>SET UP .....</b>	<b>6</b>
<b>4.1</b>	<b>ASSEMBLING OF SCALE.....</b>	<b>6</b>
<b>5.0</b>	<b>WEIGHING .....</b>	<b>6</b>
<b>5.1</b>	<b>BEFORE WEIGHING.....</b>	<b>6</b>
<b>5.2</b>	<b>WEIGHING .....</b>	<b>6</b>
<b>5.3</b>	<b>USER PARAMETERS .....</b>	<b>7</b>
<b>5.3.1</b>	<b><i>Setting backlight</i>.....</b>	<b>7</b>
<b>5.3.2</b>	<b><i>Auto off</i>.....</b>	<b>7</b>
<b>5.3.3</b>	<b><i>Unit</i>.....</b>	<b>7</b>
<b>5.4</b>	<b>PARTS COUNTING FUNCTION .....</b>	<b>7</b>
<b>5.5</b>	<b>MEMORY ACCUMULATION .....</b>	<b>7</b>
<b>6.0</b>	<b>CALIBRATION .....</b>	<b>8</b>
<b>6.1</b>	<b>USER CALIBRATION PARAMETER SETTING .....</b>	<b>8</b>
<b>7.0</b>	<b>ERROR MESSAGE .....</b>	<b>9</b>
	<b>WARRANTY INFORMATION .....</b>	<b>10</b>

### WARNING:



Please read carefully following instructions for installation and use before starting your work with a new balance. Failure to use the balance in accordance with the advice in the user manual could mean that the product gives incorrect readings.

## 1.0 INTRODUCTION

- The LBX C Models are supplied with a power supply adapter, to charge the internal rechargeable battery.
- Four models:
  - 3kg x 0.1g
  - 6kg x 0.2g
  - 15kg x 0.5g
  - 30kg x 1g
- 3 LCD displays with backlight
- 5 weighing units: kg, g, lb, oz, and lb:oz
- Check-weighing
- Indicators for Stable, zero, net weight, and low battery functions
- Stainless steel pan with ABS case and pan support
- Sealed Membrane keypad
- Overload Protection
- Spirit level and leveling feet
- Auto calibration, zero tracking and tare

## 2.0 SPECIFICATIONS


Model	LBX 3C	LBX 6C	LBX 15C	LBX 30C
Capacity	3Kg	6Kg	15Kg	30Kg
Readability	0.1g	0.2g	0.5g	1g
Repeatability	0.2g	0.4g	1g	2g
Linearity $\pm$	0.2g	0.4g	1g	2g
Units	kg / g / lb / oz / lb:oz			
Stabilization time	1 second	1 second	1 second	1 second
Operating temperature	0°C to 40°C 32°F to 104°F			
Power supply	5VDC @ 800mA adapter 4V4.5Ah rechargeable battery Typical battery life expectancy: 240 hours			
Calibration	External calibration			
Calibration Weights	1kg,2kg,3kg	2kg,4kg,6kg	5kg,10kg,15Kg	10kg,20kg,30kg
Display	3 x18mm LCD displays			

## 3.0 SAFETY PRECAUTIONS

### Caution:

Please use only the original DC adapter supplied with the scale. Other cords or adapters may damage the scale.

### Notice:

- Avoid extremes of temperature, do not place in direct sunlight or near air conditioning vents.
- Battery should be removed if the scale is not used for a long period of time.
- Do not stack material on the scale when it is not in use.
- Avoid high humidity that might cause condensation and keep away from direct contact with water.
- Static may influence the weighing result, to reduce the static wipe the pan and the scale with anti-static wipes.
- When the battery is lower than 3.7v, the battery symbol  will flash to indicate the low battery; when it is lower than 3.4v, the scale will turn off automatically.

## **4.0 SET UP**

### **4.1 ASSEMBLING OF SCALE**

- Put the scale on a stable and even table.
- Adjust the height of the feet to level the scale; the spirit level bubble should be in the centre of the spirit level.
- Put the pan support in place, inserting the pan support pins into the holes on top of the scale. Place the stainless steel pan on the support.
- Avoid excess force to weighing pan or the load cell could be damaged.

## **5.0 WEIGHING**

### **5.1 BEFORE WEIGHING**

Check the setting of the auto shut off function, see section 6.2. If the auto shut off function is on, the scale will shut off after a period of time if the scale is not used and it has a stable reading.

### **5.2 WEIGHING**

- Press [On/Off] to turn on the scale. The display will show the software revision then show the model "CAP XX". Then it will begin its self-checking test from 0 to 9. When reading is stable, the display will show "0.0", stable symbol, zero symbol, and the weighing unit used the last time the scale was used.
- Press [UNITS/→] to select units from "g" / "kg" / "lb" / "oz"/"lb:oz". Only the weighing units enabled will be displayed.
- Check if the reading is "0.0", if not, press [Zero]. "0.0" and the zero symbol will be displayed.
- Put the item to be weighed on the pan. When the reading is stable, the gross weight will be shown.

If a container is used, put the container on the pan and press the [Tare/↵] key when the reading is stable. The display will show “0.0”, the zero symbol and ‘NET’. Load the container or remove material from the container, the scale will show the net weight of the material added or removed.

## 5.3 USER PARAMETERS

The parts counting, percent weighing and calibration functions can be accessed using the [CE] key. In normal weighing press and hold [CE] key to enter user settings.

### 5.3.1 *Setting backlight*

BK backlight can be set as “auto”, “on”, “off”; default is “auto”. To select the available option press keys 1 and 6 and then press Tare key to confirm your selection.

### 5.3.2 *Auto off*

Auto off can be set as “off”, “1 min”, “5 min”, “10 min”; default is “5 min”. To select the available option press keys 1 and 6 and then press Tare key to confirm your selection.

### 5.3.3 *Unit*

Apart from kg, all other weighing units can be enabled or disabled.

## 5.4 PARTS COUNTING FUNCTION

- In normal weighing mode, press hand count the number of pieces to be counted and put them on the scale. Using the numerical keys enter the count and press [SMPL] key.
- If the weight of the item is known, enter this using the keypad and press [unit weight] key to confirm. The weight entered will be displayed in the middle display window.

## 5.5 MEMORY ACCUMULATION

- When parts counting, put the number of parts to be counted on the top pan, press [M+] key, the display will show ACC and the number of the memory entry, so for example ACC 1 will be the first memory entry.
- Ensure there is no weight on the top pan and the display is showing zero;

- Add another sample to the top pan and press the **[M+]** key; the memory entry will be incremented by 1, the middle display will show total weight of all samples counted in memory and the right display will show the total count of all samples in memory.
- To recall the total of the memory function, press the **[M+]** key when there is no weight on the top pan.
- To clear the memory, press **[M+]** key when there is no weight on the top pan and then press the **[CE]** key; the display will briefly show ACC 0 to indicate the memory has been cleared.

## 6.0 CALIBRATION

- Calibration before use is recommended for best accuracy. It is suggested that the scale be calibrated when first installed and any time the scale is moved to a different environment.
- Ensure weights are ready before calibrating, the weight should be of a known accuracy, for example OIML class M1 or Class 4 of ASTM E617.
- Do not calibrate the scale if a proper calibration weight is not available.

### 6.1 USER CALIBRATION PARAMETER SETTING

When power on, press TARE during self-checking, enter pass code 1000 to access user calibration parameter.

Use keys 1 and 6 to select available options and press TARE button to conform the selection.


#### ***U-CAL (user calibration)***

This parameter sets the calibration of the scale. Press [TARE]; display will show unload; with no weight on the top pan and the stable indicator On, press [TARE] key to set the zero point. The display will then indicate the load (SET) amount. Use Keys 1 and 6 and CE to enter the desired load amount and then press the [TARE] key. The display will then show the pre-selected (LOAD) amount for calibration; put this weight on the top pan, wait for the stable indicator to show on the display and then press the [TARE] key.

**AdCnt** – this will show the internal A/D counts.



## 7.0 ERROR MESSAGE

--AdO	A/D counts are too high
--AdL	A/D counts are too low
BEEP 15	Wrong operation
- -OL- -	Overload 9d
 and "BATTLO"	Low battery indication
-- Err -	Over zero key range

## **WARRANTY INFORMATION**

Adam Equipment offers Limited Warranty (Parts and Labour) for the components failed due to defects in materials or workmanship. Warranty starts from the date of delivery.

During the warranty period, should any repairs be necessary, the purchaser must inform its supplier or Adam Equipment Company. The company or its authorized Technician reserves the right to repair or replace the components at any of its workshops depending on the severity of the problems. However, any freight involved in sending the faulty units or parts to the service center should be borne by the purchaser.


The warranty will cease to operate if the equipment is not returned in the original packaging and with correct documentation for a claim to be processed. All claims are at the sole discretion of Adam Equipment.

This warranty does not cover equipment where defects or poor performance is due to misuse, accidental damage, exposure to radioactive or corrosive materials, negligence, faulty installation, unauthorized modifications or attempted repair or failure to observe the requirements and recommendations as given in this User Manual. Additionally rechargeable batteries (where supplied) are not covered under warranty.

Repairs carried out under the warranty does not extend the warranty period. Components removed during the warranty repairs become the company property.

The statutory right of the purchaser is not affected by this warranty. The terms of this warranty is governed by the UK law. For complete details on Warranty Information, see the terms and conditions of sale available on our web-site.

## WEEE 2012/19/EU



This device may not be disposed of in domestic waste. This also applies to countries outside the EU, per their specific requirements. Disposal of batteries (if fitted) must conform to local laws and restrictions.

Cet appareil ne peut être éliminé avec les déchets ménagers. L'élimination de la batterie doit être effectuée conformément aux lois et restrictions locales.

Dieses Gerät nicht mit dem Hausmüll entsorgt.

Dispositivo no puede ser desechado junto con los residuos domésticos

Dispositivo non può essere smaltito nei rifiuti domestici.

## FCC / IC CLASS A DIGITAL DEVICE EMC VERIFICATION STATEMENT

NOTE: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to Part 15 of the FCC rules and Canadian ICES-003/NMB-003 regulation. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.

## CALIFORNIA PROPOSITION 65 - MANDATORY STATEMENT

WARNING: This product includes a sealed lead-acid battery which contains chemicals known to the State of California to cause cancer and birth defects or other reproductive harm



Adam Equipment products have been tested with, and are always supplied with mains power adaptors which meet all legal requirements for the intended country or region of operation, including electrical safety, interference and energy efficiency. As we often update adaptor products to meet changing legislation it is not possible to refer to the exact model in this manual. Please contact us if you need specifications or safety information for your particular item. Do not attempt to connect or use an adaptor not supplied by us.

**ADAM EQUIPMENT** is an ISO 9001:2015 certified global company with more than 50 years' experience in the production and sale of electronic weighing equipment.

Adam products are predominantly designed for the Laboratory, Educational, Health and Fitness, Retail and Industrial Segments. The product range can be described as follows:

- Analytical and Precision Laboratory Balances
- Compact and Portable Balances
- High Capacity Balances
- Moisture analysers / balances
- Mechanical Scales
- Counting Scales
- Digital Weighing/Check-weighing Scales
- High performance Platform Scales
- Crane scales
- Mechanical and Digital Electronic Health and Fitness Scales
- Retail Scales for Price computing

For a complete listing of all Adam products visit our website at [www.adamequipment.com](http://www.adamequipment.com)

<p><b>Adam Equipment Co. Ltd.</b> Maidstone Road, Kingston Milton Keynes MK10 0BD UK Phone:+44 (0)1908 274545 e-mail: <a href="mailto:sales@adamequipment.co.uk">sales@adamequipment.co.uk</a></p>	<p><b>Adam Equipment Inc.</b> 1, Fox Hollow Rd. 06478  USA Phone: +1 203 790 4774 e-mail: <a href="mailto:sales@adamequipment.com">sales@adamequipment.com</a></p>	<p><b>AE Adam GmbH.</b> Instenkamp 4 D-24242 Felde  Germany Phone +49 (0)4340 40300 0 e-mail: <a href="mailto:vertrieb@aeadam.de">vertrieb@aeadam.de</a></p>
<p><b>Adam Equipment S.A. (Pty) Ltd.</b> 7 Megawatt Road, Spartan EXT 22 Kempton Park, Johannesburg, Republic of South Africa  Phone +27 (0)11 974 9745 e-mail: <a href="mailto:sales@adamequipment.co.za">sales@adamequipment.co.za</a></p>	<p><b>Adam Equipment (S.E. ASIA) PTY Ltd</b> 70 Miguel Road Bibra Lake Perth WA 6163 Western Australia  Phone: +61 (0) 8 6461 6236 e-mail: <a href="mailto:sales@adamequipment.com.au">sales@adamequipment.com.au</a></p>	<p><b>Adam Equipment (Wuhan) Co. Ltd.</b> A Building East Jianhua Private Industrial Park Zhuanyang Avenue Wuhan Economic &amp; Technological Development Zone 430056 Wuhan P.R.China Phone: + 86 (27) 59420391  e-mail: <a href="mailto:info@adamequipment.com.cn">info@adamequipment.com.cn</a></p>

© Copyright by Adam Equipment Co. All rights reserved. No part of this publication may be reprinted or translated in any form or by any means without the prior permission of Adam Equipment.

Adam Equipment reserves the right to make changes to the technology, features, specifications and design of the equipment without notice.

All information contained within this publication is to the best of our knowledge timely, complete and accurate when issued. However, we are not responsible for misinterpretations which may result from the reading of this material.

The latest version of this publication can be found on our Website.

[www.adamequipment.com](http://www.adamequipment.com)