Entrant[™] HF CRT-D

CDHFA300B





Compatible with myMerlinPulse™ app

Product Highlights

- Bluetooth* Low Energy (LE) communication enabling Smartphone Connectivity through data encryption.
- SyncAV[™] CRT technology offers dynamic AV timing with customizable programming to ensure BiV pacing.
- Cold can programmability provides an additional RV-SVC shock configuration to decouple the can from the shocking vector parameters in cases of lead problems.
- DeFT Response[™] Technology offers noninvasive programming options to optimize rescue therapy to each patient's unique physiology and changing conditions.
- VF Therapy Assurance decreases time to treatment for arrhythmias in patients who are likely to be hemodynamically unstable.
- Antitachycardia pacing (ATP) while charging and prior to charging in the VF zone further extends the programming options for terminating tachyarrhythmias without a highvoltage shock.
- ShockGuard[™] technology with DecisionTx[™] programming designed to reduce inappropriate therapy and minimize the need for programming adjustments at implant.
 - SecureSense™ RV lead noise discrimination detects sustained lead noise and short bursts of oversensing that would otherwise go unnoticed or potentially lead to one or more inappropriate shocks.

- Far Field MD™ morphology discrimination and Chamber Onset discrimination enhance SVT and VT discrimination for reduced inappropriate therapies.
- SenseAbility[™] sensing algorithm feature provides the flexibility to fine-tune programming around T-wave oversensing without decreasing sensitivity.
- DynamicTx[™] over-current detection algorithm automatically changes shock configurations to ensure delivery of highvoltage therapy when high current is detected.
- MRI-Ready device tested in combination with MR Conditional leads for full-body scans using a 1.5T or 3T (Tesla) field strength MRI Scanner*.
- Premature Atrial Contraction (PAC) Response to avoid pacing the atrium in a vulnerable zone.
- Physiologic rate responsive AV Delay and PVARP.
- QuickOpt[™] timing cycle optimization provides quick and effective optimization at the push of a button.
- Dual patient notification: audio notification through the device and visual notification via myMerlinPulse™ app.
- The CorVue[™] thoracic impedance feature measures transthoracic impedance changes over time to provide additional insight into the patient's heart failure condition.

Ordering Information

Contents: Cardiac Pulse Generator

MODEL NUMBER	DIMENSIONS (L × W × H) (MM)	WEIGHT (G)			CONNECTOR SENSE/BACE	CONNECTOR PACE - LEFT VENTRICLE
CDHFA300B	79 × 51 × 12	79	37	DF-1	IS-1	IS-1

*See MRI Scan Parameters in MRI-Ready Systems Manual.





Model	CDHFA300B		
Telemetry	Bluetooth* LE Communication		
Delivered/Stored Energy	36/39 J		
Volume	37 cc		
Weight			
	79 g		
Size	79 × 51 × 12 mm		
Defibrillation Lead Connection	DF-1		
Atrial Sense/Pace Lead Connection	IS-1 in-line bipolar		
Ventricular Sense/Pace Lead Connection	IS-1 in-line bipolar		
Left Ventricular Pace Lead Connection	IS-1		
High Voltage Can	Electrically active titanium can		
Parameter	Settings		
Biventricular Pacing			
V. Triggering	On; Off		
V-V Timing	Simultaneous [†] ; RV First; LV First		
Interventricular Pace Delay	RV First 10-80/LV First 15-80 ms		
Ventricular Sensing	RV only (not programmable)		
Ventricular Pacing Chamber	RV only; Biventricular		
SyncAV™ CRT Technology Delta	-10 to -120 ms; Off		
Sensing/Detection			
Sense <i>Ability</i> ™ Sensing Algorithm	Automatic sensitivity control adjustment for atrial and ventricular events		
Low Frequency Attenuation	On; Off		
Threshold Start	Post-Sensed: 50; 62.5; 75; 100%; Post-Paced; Atrial: 0.2-3.0 mV Post-Paced; Ventricular: Auto: 0.2-3.0 mV		
Decay Delay	Post-Sensed: 0-220 ms Post-Paced; Atrial: 0-220 ms Post-Paced; Ventricular: Auto; 0-220 ms		
Ventricular Sense Refractory	125; 157 ms		
Detection Zones	3 zone programming - 1 zone; 2 zones or 3 zones (VT-1; VT-2; VF)		
SVT Discriminators	AV Rate Branch; Arrhythmia Onset (Chamber Onset or Sudden Onset); Interval Stability; AV Association Morphology; Discrimination (Far Field MD™ Morphology Discrimination or Original MD) with Automatic Template Update		
Monitor Mode	Detection; discrimination and diagnostics; no therapy delivery (VT or VT-1 zone)		
Discrimination Modes	On; Passive; Off		

Parameter	Settings				
	Settings				
Sensing/Detection	150 240 hours				
SVT Upper Limit	150-240 bpm				
SVT Discrimination Timeout	20s-60 min; Off Continuous sonsing during charging				
Reconfirmation	Continuous sensing during charging				
SecureSense™ RV Lead Noise Discrimination Algorithm	On; On with Timeout; Passive; Off				
VF Therapy Assurance	On; Off				
Antitachycardia Pacing Therapy					
ATP Configurations	Ramp; Burst; Scan; 1 or 2 schemes per VT zone				
ATP in VF Zone	ATP While Charging; ATP Prior to Charging; Off				
ATP Upper Rate Cutoff	150-300 bpm				
Burst Cycle Length	Adaptive (50%-100%); Fixed (200-550 ms)				
Min. Burst Cycle Length	150-400 in increments of 5 ms				
Readaptive	On; Off				
Number of Bursts/Stimuli	1-15 with 2-20 Stimuli				
Add Stimuli per Burst	On; Off				
ATP Pulse Amplitude	7.5 V independent from Bradycardia and Post-Therapy Pacing				
ATP Pulse Width	1.0 or 1.5 ms independently programmable from Bradycardia and Post-Therapy Pacing				
High-Voltage Therapy					
DynamicTx [™] Over-Current Detection Algorithm	On; Off				
DeFT Response™ Technology	Programmable pulse width for P1/P2 and tilt				
High-Voltage Output Mode	Fixed Pulse Width; Fixed Tilt				
Waveform	Biphasic; Monophasic				
RV Polarity	Cathode (-); Anode (+)				
Electrode Configuration	RV to Can; RV to SVC/Can; RV to SVC				
Bradycardia Pacing					
Permanent Modes	DDD(R); DDT(R); DDI(R); VVT(R); VVI(R); AAI(R); Off				
Temporary Modes	DDD; DDT; DDI; VVT; VVI; AAI; AAT; DOO; VOO; AOO; Off				
Rate-Adaptive Sensor	On; Off; Passive				
Programmable Rate and Delay Parameters	Off; Base Rate (bpm); Rest Rate (bpm); Maximum Tracking Rate (bpm); Max Trigger Rate (bpm) Maximum Sensor Rate (bpm); Paced AV Delay (ms); Sensed AV Delay (ms); Rate Responsive AV Delay; Hysteresis Rate (bpm); Rate Hysteresis with Search				
Pulse Amplitude	0.25-7.5 V				
Pulse Width	0.05; 0.1-1.5 ms				
LVCap [™] Confirm Feature RVCap [™] Confirm Feature	Setup; On; Monitor; Off Setup; On; Monitor; Off				
ACap™ Confirm Feature	On; Monitor; Off				
Auto Mode Switch (AMS)	DDI(R); DDT(R); VVI(R); VVT(R); Off				
Atrial Tachycardia					
Detection Rate	110-300 bpm				
AMS Base Rate	40; 45; 135 bpm				
Auto PMT Detection/Termination	Atrial Pace; Passive; Off				
Rate Responsive PVARP	Low; Medium; High; Off				
Rate Responsive V Pace Refractory	On; Off				
PAC Response	On; Off				
PAC Response Interval	200-400 ms				
Shortest AV Delay	25-120 ms				



Parameter	Settings	
Post-Therapy Pacing (Independently	programmable from Bradycardia and ATP)	
Post-Shock Pacing Mode	AAI; VVI; DDI; or DDD; Off	
Post-Shock Base Rate	30-100 bpm	
Post-Shock Pacing Duration	0.5; 1; 2.5; 5; 7.5; or 10 min; Off	
Device Testing/Induction Methods	0.0, 1, 2.0, 0, 7.0, 01 10 min, On	
DC Fibber™ Induction Method		
Pulse Duration	0.5-5.0 sec	
BurstFibberCycle Length	20-100 ms	
Noninvasive Programmed Stimulation (NIPS)	2-25 stimuli with up to three extra stimuli	
Patient Notifiers		
Programmable Notifiers (On; Off)	BatteryAssurance [™] alert; Possible HV circuit damage; HV charge timeout; Long charge time for Capacitor Maintenance; Device at ERI; Right ventricular pacing lead impedance out of range; Left ventricular lead impedance out of range; High-voltage lead impedance out of range; AT/AF episode duration; AT/AF Burden; High ventricular rate during AT/AF; SecureSense [™] lead noise detection; Non-sustained ventricular oversensing; Biventricular pacing percentage lower than limit	
Device Parameter Reset	On	
Entry into Backup VVI Mode	On	
Auditory Duration	2; 4; 6; 8; 10; 12; 14; 16 sec	
Number of Audio Alerts per Notification	2	
Number of Notifications	1-16	
Time Between Notifications	10; 22 hours	
Electrograms and Diagnostics		
Stored Electrograms	Up to 15 minutes (2 user programmable + discrimination channel); up to one minute programmable pretrigger data per VT/VF electrograms; additional triggers include lead noise detection; non-sustained ventricular oversensing; morphology template updates; atrial episode; PMT termination; PAC response; magnet reversion; noise reversion	
Therapy Summary	Diagram of therapies delivered	
Episodes Summary	Directory listing of up to 60 episodes with access to more details including stored electrograms	
Lifetime Diagnostics	History of bradycardia events and device-initiated charging	
AT/AF Burden Trend	Trend data and counts	
Ventricular HV Lead Impedance Trend	Multi-Vector Trend Data	
Histograms and Trends	Event Histogram; AV Interval Histogram; Mode Switch or AT/AF Duration Histogram; Peak Filtered Atrial Rate Histogram; Atrial Heart Rate Histogram; Ventricular Heart Rate Histogram; AT/AF Burden; Exercise and Activity Trending; V Rates During AMS; DirectTrend™ reports up to 1 year	
PMT Data	Information regarding PMT detections	
Real-Time Measurements (RTM)	Pacing lead impedances; high-voltage lead impedances; and signal amplitudes	
CorVue Thoracic Impedance CorVue Thoracic Impedance	On; Off Threshold 8-18 days	
MRI Settings		
Tachy Therapy	Disabled	
MRI Mode	DOO; VOO; AOO; Pacing Off	
MRI Base Rate	30-100 bpm	
MRI Paced AV Delay	25-110 ms	
MRI RA and RV Pulse Amplitude	5.0 or 7.5 V	
MRI RA and RV Pulse Width	1.0 ms	
MRI RA and RV Pulse Configuration	Bipolar	
MRI V Pacing Chamber	RV Only	
MRI Timeout	3; 6; 9; 12; 24 hours; Off	



MRI SCAN PARAMETERS ⁵					
Lead Model	Magnet (Tesla)	RF Transmit Conditions	Scan Region		
Durata™ Defibrillation Lead			Full-body		
7120 (Lead lengths: 65 cm)	1.5 T / 3 T				
7122 (Lead lengths: 60, 65 cm)		_			
Optisure [™] Lead	1.5 m / 2 m				
LDA210 (Lead lengths: 65 cm)	1.5 T / 3 T	Normal			
Tendril™ STS Pacing Lead		Operating Mode			
2088TC (Lead lengths: 46, 52, 58 cm)	1.5 T / 3 T				
UltiPace™ Pacing Lead					
LPA1231 (Lead lengths: 46, 52, 58, 65 cm)	1.5 T / 3 T				

[†]LV first with 10 ms interventricular delay.

Brief Summary: This product is intended for use by or under the direction of a Physician. Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use.

15900 Valley View Court, Sylmar, CA 91342 Tel: +1 818 362 6822 Abbott.com



For additional information about specific MR Conditional CRT-Ds and leads, including scan parameters, warnings, precautions, adverse conditions to MRI scanning, and potential adverse events, please refer to the Abbott MRI-Ready Systems Manual at manuals.eifu.abbott.

 $^{^{\}circ}$ Indicates a trademark of the Abbott group of companies. $^{\circ}$ Indicates a third party trademark, which is property of its respective owner.