

Certificate of analysis

Product name: **Starrsed Rinse solution**
Article number: **QRR 010934**
LOT number: 2323509818
Expiry date: 2025-08

LOT release date: 2023-09-08

Test	Specification	Measured value
Conductivity (mS)	1020 - 1060	1035
Performance test*	passes test	Passes test
pH at 25°C	7.10 - 7.60	7.50
Total Bacterial Growth (CFU/100ml)	Max. 5	0

*) All parameters except mid cells: Test/Ref ration > 95% and < 105%. Correlation coefficient: > 0.95

This lot of product has been controlled and approved for distribution.

Country of origin: The Netherlands

For technical information regarding this product, refer to the Package Insert (PI).

For safety related information regarding this product, refer to the Safety Data Sheet (SDS).

For questions about this certificate of analysis, please contact the Quality Assurance & Regulatory Affairs department at gara@rrmechatronics.com

Signed: 

Date: 13-nov-2023

QARA Department

