



全漢企業股份有限公司

安規證書

料號 :
Model Name : FSP250-60MPC,FSP220-60MPC,SPI250F4BMB,SPI220F4BMB
版次 : 04
文件編號 : SFD11008985-R4
作者 : chow/周懿慧(FSP)
Model No/Type : FSP250-60MPC,FSP220-60MPC,SPI250F4BMB,SPI220F4BMB
SFA NO. : SFA23041121
Trade mark : FSP & SPI
安規種類 : EMC/醫療類
發行日期 : 2023/05/29 11:40:25
備註 : EMC 醫療類 4.1 for FSPxxx-60MPC series



Technical Compliance Statement

EMC Test Report

For the following information

Ref. File No.: C1M2305071

Product : Switching Power Supply

Model Name : (1)FSP250-60MPC (2)FSP220-60MPC (3)SPI250F4BMB
(4)SPI220F4BMB

Brand Name : (1)~(2)FSP (3)~(4)SPI

Applicant : FSP Group Inc.

Standards :

EN 60601-1-2:2015 +A1:2021 (IEC 60601-1-2:2014 +AMD:2020)(Edition 4.1)
BS EN 60601-1-2-2015 +A1:2021

Emission: CISPR 11:2015 +A1:2016 +A2:2019, Class B
EN 55011:2016 +A11:2020
EN IEC 61000-3-2:2019+A1:2021 (IEC 61000-3-2:2018 +A1:2020),
BS EN IEC 61000-3-2:2019+A1:2021 and
EN 61000-3-3:2013 +A1:2019 (IEC 61000-3-3:2013 +A1:2017),
EN 61000-3-3:2013 +A1:2019 +A2:2021
(IEC 61000-3-3:2013 +A1:2019 +A2:2021)
BS EN 61000-3-3:2013 +A1:2019,
BS EN 61000-3-3:2013 +A1:2019 +A2:2021

Immunity: IEC 61000-4-2:2008, IEC 61000-4-3:2020
IEC 61000-4-4:2012, IEC 61000-4-5:2014+A1:2017,
IEC 61000-4-6:2013, IEC 61000-4-8:2009,
IEC 61000-4-11:2020, IEC 61000-4-39:2017

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the Regulation (EU) 2017/745 concerning medical device. The test data and results are included in the test report **EM-E160240B**.

Signature



Tony Lee/Deputy Manager
Date: 2023. 05. 24

Test Laboratory:
Audix Technology Corporation, EMC Department
TAF Accreditation No.: 1724
Web Site: www.audixtech.com

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.