



REF CDC-12802-1A

8 Fr. 2 Lumen 16 cm catheter length .032 inch dia. spring-wire guide



Two-Lumen CVC

Contents:

- 1: Two-Lumen Catheter: 8 Fr. (2.8 mm OD) x 16 cm

1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer

1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle

1: Injection Needle: SafetyGlide™¹ 25 Ga. x 1" (2.54 cm)

1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)

1: Pressure Transduction Probe

1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe

1: Injection Needle: SafetyGlide™¹ 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe

1: Syringe: 3 mL Luer-Lock

1: Tissue Dilator: 9 Fr. (3.0 mm) x 10.2 cm

1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep

1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®² One-Step Solution with Hi-Lite Orange™ Tint

1: SecondSite™ Adjustable Hub: Fastener

1: SecondSite™ Adjustable Hub: Catheter Clamp

1: SharpsAway® Disposal Cup

1: SharpsAway® II Locking Disposal Cup

1: Maximal Barrier Drape™ with 4" fenestration

1: Needle Holder

1: Safety Scalpel: #11

1: Checklist/CLIP Sheet

1: Sterile Procedure Sign

1: Medication Label: 1% Lidocaine
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)

5: Gauze Pad: 4" x 4" (10 cm x 10 cm)

1: Surgical Apparel: Impervious Gown

1: Surgical Apparel: Bouffant Cap

1: Surgical Apparel: Mask with Eye Shield

1: Suture: 3-0 Silk with Curved Needle

1: HemoHopper® Fluid Receptacle
- ¹A trademark of Becton, Dickinson and Company.

²A registered trademark of CareFusion or one of its subsidiaries.

Not made with natural rubber latex.
Store between 20 – 25°C (68 – 77°F).
Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (14 Ga.)	0.66	5279
Proximal (14 Ga.)	0.71	4764

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

