

In association with

St John
Ambulance





kinetik
WELLBEING

-Quick Start Guide-

JPD-500G

Finger Pulse
Oximeter

1



Open battery cover,
insert batteries
and replace cover.

2




Warm your hands.
Remove nail polish
and false nails
if necessary.

3




Rest for 5 minutes
before taking
your measurement.

4



Insert finger
into device.

5



Press button to
take measurement.
Readings can take
up to 1 minute
to stabilise.





for more information visit :
kinetikwellbeing.com