

Make/Manufacturer:	
Model Number:	
Serial Number:	
Gas Detected:	
Detector Location:	
Purchase Date:	

## Calibration Best Practices

- Ensure calibration takes place at least every **12 months.**
- Follow the manufacturer's calibration step-by-step instructions for your detector model.
- Before calibrating, ensure Zero Calibration takes place and the correct exposure time is used.
- Calibrate with a certified NIST reference gas that is within its expiration date.
- Never calibrate in confined spaces.
- Verify that gas detector alarm setpoints and alarms are functioning properly.

## Calibration History

[illegible]