



全漢企業股份有限公司

安規證書

料號 :
Model Name : FSP065-DHBM1,FSP065-DGBM1,FSP065-DDBM1,FSP065-DBBM1,FSP065-DCBM1,FSP065-DABM1,PMP65S-12,PMP65S-13,PMP65S-13-1,PMP65S-13-2,PMP65S-13-3,PMP65S-14
版次 : 03
文件編號 : SFD14021383-R4
作者 : zoedai/戴淑慧(FSP)
Model No/Type : FSP065-DHBM1,FSP065-DGBM1,FSP065-DDBM1,FSP065-DBBM1,FSP065-DCBM1,FSP065-DABM1,PMP65S-12,PMP65S-13,PMP65S-13-1,PMP65S-13-2,PMP65S-13-3,PMP65S-14
SFA NO. : SFA22011353
Trade mark : -
安規種類 : EMC/醫療類
發行日期 : 2022/02/17 12:16:01
備註 : EMC醫療類指令中的EN 55011:2016/A11:2020 法規升版

Ref. No. VXD0341A/2022

Page: 1 of 1

VERIFICATION OF COMPLIANCE

Issue Date: February 15, 2022
Applicant: FSP Group Inc.
Address: No 22, Jianguo East Road, Taoyuan City,
 Taiwan 330, R.O.C.
Manufacturer: N/A
Address: N/A
Contact Information: Web: N/A
 TEL#: N/A
 E-mail#: N/A
Product: Switching Power Adapter
Brand Name/Trade Mark: N/A
Model/Type: FSP065-DHBM1
Added Model(s): FSP065-DGBM1; FSP065-DDBM1; FSP065-DBBM1; FSP065-
 DCBM1; FSP065-DABM1; PMP65S-12yyyyy; PMP65S-13yyyyy;
 PMP65S-13-1yyyyy; PMP65S-13-2yyyyy; PMP65S-13-3yyyyy;
 PMP65S-14yyyyy (y=0-9, A-Z, hyphen or blank.)
Applicable Standards: IEC 60601-1-2: 2014
 EN 60601-1-2: 2015
 CISPR 11: 2015 + A1: 2016 (Ed 6.1)
 EN 55011: 2016 / A11: 2020 (Group 1, Class B)
 IEC 61000-4-2: 2008; EN 61000-4-2: 2009
 IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010;
 EN 61000-4-3: 2006 + A1: 2008 + A2: 2010
 IEC 61000-4-4: 2012; EN 61000-4-4: 2012
 IEC 61000-4-5: 2014 + A1: 2017; EN 61000-4-5: 2014 + A1: 2017
 IEC 61000-4-6: 2013 + COR1: 2015; EN 61000-4-6: 2014 + AC: 2015
 IEC 61000-4-8: 2009; EN 61000-4-8: 2010
 IEC 61000-4-11: 2004 + A1: 2017; EN 61000-4-11: 2004 + A1: 2017
 IEC 61000-3-2: 2018 / A1: 2020; EN IEC 61000-3-2: 2019 / A1: 2021
 IEC 61000-3-3: 2013 + AMD1: 2017; EN 61000-3-3: 2013 / A1: 2019
Test Laboratory: Compliance Certification Services Inc.
 Xindian Lab
 No. 163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, Taiwan
Test Report No.: TMXD2202000527DE, dated on February 15, 2022
Conclusion: Based upon a review of the Test Report(s), the tested sample of the product
 mentioned above is deemed to comply with the requirements of the above
 standards
Note: This verification is only valid for the product and configuration described and in conjunction
 with the test report as detailed above.

Authorised Signatory:


Compliance Certification Services Inc.
 Jason Lee
 Section Manager

FM-023A-R01