



MICHELL Instruments XTP 601 SIL Oxygen Analyzer User Manual

[Home](#) » [MICHELL Instruments](#) » MICHELL Instruments XTP 601 SIL Oxygen Analyzer User Manual 

Contents

- 1 MICHELL Instruments XTP 601 SIL Oxygen Analyzer
- 2 Product Information
- 3 Product Usage Instructions
- 4 Safety Guidelines
- 5 INTRODUCTION
- 6 SAFETY INSTRUCTIONS
- 7 DEVICE-SPECIFIC SAFETY INSTRUCTIONS
- 8 Appendix A
- 9 Documents / Resources
 - 9.1 References



MICHELL Instruments XTP 601 SIL Oxygen Analyzer



Product Information

Specifications:

- **Product Models:** XTP, XTC, XPM 601
- **Safety Manual:** XTP/XTC/XPM 601 SIL Safety Manual
- **Issue Date:** March 2024

Product Usage Instructions

Introduction

Ensure you have the necessary documentation for reference.

Safety Instructions

Understand the Safety Integrity Level (SIL) of the product for safe operation.

Device-Specific Safety Instructions

- **Applications:** Understand the intended use of the product.
- **Safety Function:** Familiarize yourself with the safety features.
- **Settings:** Adjust settings as required for your application.
- **In case of faults:** Know how to respond to any malfunctions.
- **Maintenance/Calibration:** Follow recommended maintenance and calibration procedures.
- **Safety Characteristics:** Be aware of the safety aspects of the product.

Frequently Asked Questions (FAQ):

- **Q: Can I modify the product?**

A: No, unauthorized modifications can compromise safety and void warranties. Contact Michell Instruments Ltd for any queries.

- **Q: What should I do in case of faults?**

A: Refer to the manual for troubleshooting steps or contact Michell Instruments Ltd for assistance.

For Michell Instruments' contact information please go to www.ProcessSensing.com

© 2024 Michell Instruments

This document is the property of Michell Instruments Ltd and may not be copied or otherwise reproduced, communicated in any way to third parties, nor stored in any Data Processing System without the express written authorization of Michell Instruments Ltd.

The contents of this safety manual shall not become part of or modify any prior or existing agreement, commitment or legal relationship. All obligations on the part of Michell Instruments are contained in the respective sales contract which also contains the complete and solely applicable warranty conditions. Any statements contained herein do not create new warranties or modify the existing warranty.

NOTE: This product must not be modified or altered in any way. Unauthorised change is not permitted and to do so would cause the Functional Safety, as confirmed by the IEC 61508 assessment, to be null-and-void. This products design is strictly controlled and to do so would invalidate all approvals, certification and warranties this product holds. Please consultant Michell Instruments Ltd directly for any functionality or service queries you may have.

Safety Guidelines

This manual relates only to the SIL aspects of this product. For all other operation, installation & maintenance information refer to the product manual. The user must not use this equipment for any other purpose than that stated. Do not apply values greater than the maximum value stated. This manual contains information relating to the SIL aspects of operating this product. Use competent personnel using good engineering practice for all procedures in this manual.

Qualified Personnel

This product should only be set up and used in conjunction with this documentation. Commissioning and operation of this product should only be performed by qualified personnel.

Abbreviations

The following abbreviations are used in this manual:

- λ Failure Rate
- λD Dangerous Failure Rate
- λDD Dangerous Detected Failure Rate
- λDU Dangerous Undetected Failure Rate
- λs Safe Failure Rate
- /hr Per Hour
- **ADC** Analogue-To-Digital Converter
- **DAC** Digital-To-Analogue Converter
- **DC** Diagnostic Coverage
- **E/E/PE** Electrical/Electronic/Programmable Electronic
- **EMF** Electromotive Force

- **ESC** Engineering Safety Consultants
- **EUC** Equipment Under Control
- **FIT** Failure in time
- **FMEDA** Failure Mode Effect and Diagnostics Analysis
- **FMR** Failure Mode Ratio
- **FS** Functional Safety
- **FSM** Functional Safety Management
- **HFT** Hardware Fault Tolerance
- **MDT** Mean Down Time
- **MTTR** Mean Time To Restoration
- **NPRD** Non-Electronic Parts Reliability Data
- **O2** Oxygen
- **O/C** Open Circuit
- **PFD** Probability of Failure on Demand
- **PFH** Average Frequency of a Dangerous Failure per Hour
- **PLC** Programmable Logic Controller
- **PTI** Proof Test Interval
- **QA** Quality Assurance
- **RBD** Reliability Block Diagram
- **S/C** Short Circuit
- **SFF** Safe Failure Fraction
- **SIF** Safety Instrumented Function
- **SIL** Safety Integrity Level
- **SR** Safety Related
- **Tp** Proof Test Interval

INTRODUCTION

General

This manual refers only to:

- XTP601 Oxygen Transmitter
- XTP601 Oxygen Analyzer
- XTC601 Binary Gas Analyzer
- XTC601 Binary Gas Transmitter
- XPM601 Paramagnetic Oxygen Analyzer

There are derivatives of each model as shown in the below table

Analyzer Name	Type
XTP601-GP1	General purpose analyzer with display
XTP601-GP2	General purpose analyzer with flame arrestors
XTP601-EX1	Hazardous area analyzer with display
XTC601-GP1	General purpose analyzer with display
XTC601-GP2	General purpose analyzer with flame arrestors
XTC601-EX1	Hazardous area analyzer with display
XPM601-EX	Hazardous area analyzer with display

Required documentation

This document only applies in conjunction with the following documentation:

Analyzer Name	Type	Document No.
XTP601	Process Oxygen Analyzer User's Manual (UK)	97313
XTC601	Binary Gas Analyzer User's Manual (UK)	97400
XPM601	Paramagnetic Oxygen Analyzer	97632

NOTE: For each type, there are manuals with the same content translated into other languages.

This document contains SIL-related data that will be required when using the XTP601, XTC601 and XPM601 products in safety-instrumented systems.

It is aimed at system planners, constructors, service and maintenance engineers and personnel who will commission the device.

SAFETY INSTRUCTIONS

These products are intended for use in safety applications.

All safety instructions relate exclusively to the analog output signal (4...20 mA). The products meet the requirements of IEC 61508 (SIL2 Capable). The product software meets the requirements of IEC 61508 (SIL2 Capable). The use of these products integrated in to safety-related systems is therefore possible.

Definition: Safety-instrumented system

A safety-instrumented system executes the safety functions that are required to achieve or maintain a safe status in a system. It consists of a sensor, logic unit/control system and final controlling element. A Safety Instrumented System (SIS) could be made of an analyzer (e.g. XTP O2 Concentration), a Safety rated logic Solver (e.g. safety relay or safety rated PLC) and a final element (e.g. valve, or alarm with defined response).

Definition: Safety function

Defined function executed by a safety-instrumented system to achieve or maintain a safe system considering a defined dangerous occurrence.

Example: XTP O2 concentration above or below a defined threshold.

Safety Integrity Level (SIL)

The international standard IEC 61508 defines four discrete Safety Integrity Levels (SIL) from SIL 1 to SIL 4. Each level corresponds to the probability range for the failure in a safety function. The higher the SIL of the safety-instrumented system, the higher the probability that the required safety function will work.

The achievable SIL is determined by the following safety characteristics:


- Average probability of dangerous failure of a safety function in case of demand (PFDavg)
- Hardware fault tolerance (HFT)
- Safe failure fraction (SFF)

Description:

The following table shows the dependency of the SIL on the average probability of dangerous failures of a safety function of the entire safety-instrumented system (PFDavg). The table deals with “Low demand mode”, i.e. the safety function is required a maximum of once per year on average.

SIL level	PFDavg
SIL 4	$10^{-4} > \text{PFDavg} \geq 10^{-5}$
SIL 3	$10^{-3} > \text{PFDavg} \geq 10^{-4}$
SIL 2	$10^{-2} > \text{PFDavg} \geq 10^{-3}$
SIL 1	$10^{-1} > \text{PFDavg} \geq 10^{-2}$

The “average probability of dangerous failures of the entire safety-instrumented system” (PFDavg) is normally split between the entire SIL system.



The following table shows the achievable Safety Integrity Level (SIL) for the entire safety-instrumented system for type B systems depending on the proportion of safe failures (SFF) and the hardware fault tolerance (HFT). XTP, XTC and XPM units are considered Type B due to their complexity. Type B systems also include sensors and positioners actuators with complex components, e.g. microprocessors (see also IEC 61508, Section 2).

SFF	HFT		
	0	1	2
<60%	Not allowed	SIL1	SIL2
60 to 90%	SIL1	SIL2	SIL3
90 to 99%	SIL2	SIL3	SIL4
>99%	SIL3	SIL4	SIL4

DEVICE-SPECIFIC SAFETY INSTRUCTIONS

Applications

The Hardware assessment of the XTP601, XTC601 and XPM601 shall provide the safety instrumentation engineer with the required failure data as per IEC 61508. The hardware of XTP601, XTC601 and XPM601 satisfies the requirements in terms of functional safety in accordance with IEC 61508 (SIL Capable). The XTP601, XTC601 and XPM601 are usable in safety applications to monitor limits.

Safety Function

The XTP601, XTC601 and XPM601 are mainly used for user-defined threshold monitoring.

The XTP601 Process Oxygen Analyzer and the XPM601 Paramagnetic Analyzer were assessed against the following safety function:

- Ability to detect oxygen presence within another gas stream and generate a 4...20 mA output.

The XTC601 Binary Gas Analyzer was assessed against the following safety function:

- Ability to detect target gas in another gas stream and generate a 4...20 mA output.

Warning

See the “Settings” and “Safety characteristics” sections for the binding settings and conditions. These conditions must be met to fulfil the safety function. When the safety function has been executed, safety-instrumented systems with no self-locking function should be brought to a monitored or otherwise safe status within the Mean Time To Repair (MTTR). The MTTR is 168 hours. For full product information refer to User Manuals 97313, 97400 and 97632.

Settings

After installation and commissioning (refer to User Manuals), the following parameter settings should be made for the safety function:

- **Safety parameters**

Function	
Analog Output	Select 4...20 mA (NAMUR)

- **Protection against configuration changes**

After configuration, the menu access codes of XTP601, XTC601 and XPM601 shall be changed so that the device is protected against unauthorized changes and operation.

- **Checking the safety function after installation**

After installation a safety function test must be carried out. Using reference gas, i.e. N₂, 4mA must be measured at the analog output. For the test of the safety function it is fundamental to use a second reference gas with a defined proportion of oxygen. The results of the measurement must be within a range of ±5% (full span) of the expected result.

In case of faults

- **Fault**

The procedure in case of faults is described in the user manuals.

- **Repair**

The defective product should be sent to a Michell Instruments Service Department with details of the fault and

the cause. When ordering a replacement product, please specify the serial number of the original product. The serial number can be found on the nameplate. Information regarding the location of Michell Instruments Service centers can be found at the following web address: www.ProcessSensing.com.

Maintenance/Calibration

We recommend that the functioning of the XTP601, XTC601 and XPM601 is checked at one-year intervals. Check at least the following: Test the basic functionality of the XTP601, XTC601 and XPM601 as described in the user manual.

Checking safety

You should regularly check the safety function of the entire safety circuit in line with IEC 61508/61511.

The testing intervals are determined during the circulation of each individual safety circuit in a system. The recommended prove interval depends on the application but it should be at least once a year.

To detect dangerous undetected faults, the XTP601, XTC601 and XPM601 analog output shall be checked with the following test:

To execute the safety proof test both tests (1 and 2) must be performed.

Proof test 1 consists of the steps described in the table below.

Step	Action
1	Bypass the safety PLC or take other appropriate action to avoid a false trip.
2	Generate or simulate an alarm condition to force the product to go to the high alarm current output and verify that the analog current reaches that value.
3	Generate or simulate an alarm condition to force the product to go to the low alarm current output and verify that the analog current reaches that value.
4	Restore the loop to full operation.
5	Remove the bypass from the safety PLC or otherwise restore normal operation.

Proof test 2 consists of the steps described in the table below

Step	Action
1	Bypass the safety PLC or take other appropriate action to avoid a false trip.
2	Perform Proof Test 1.
3	Perform a 2-point calibration of the product.
4	Perform a reference measuring with at least one measuring point between min and max concentration. You must use a calibration gas with a well-known gas concentration. The expected result must have a tolerance of not more than 5%.
5	Restore the loop to full operation.
6	Remove the bypass from the safety PLC or otherwise restore normal operation.

This test will detect more than 90% of possible “du” failures in the product. Should faults be detected, the product should not be used until completely rectified.

Safety Characteristics

The safety characteristics necessary for use of the system are listed in the SIL declaration of conformity (see

Appendix A.1). These values apply under the following conditions:

- The XTP601, XTC601 and XPM601 are only used in safety-related systems with a low-demand mode for the safety function.
- The safety-related parameters/settings (see “Settings” section) have been entered by local operation and checked before commencing safety-instrumented operation.
- The XTP601, XTC601 and XPM601 are blocked against unwanted and unauthorized changes/operations.
- The maximum operating temperature for the XTP601, XTC601 and XPM601 is +60 °C, but guidance in the user manual must be followed.
- All used materials are compatible with process conditions.
- The MTTR after a device fault is 168 hours.
- The logic solver (PLC) must be configured to detect over-range (>21mA) and under-range (<3.6mA) failure of the XTP601, XTC601 and XPM601 (Fail High and Fail Low) and will recognize these as internal failures of the products and not cause a spurious trip.

Also see the Settings section of this manual and Appendix below.

Appendix A

A.1 SIL Declaration of Conformity

IEC 61508 Safety Integrity Level Capability Certificate

Functional Safety of Safety-Related Programmable Electronic Systems

The Michell Instruments UK Ltd, XTP601 Process Oxygen Analyser, XTC601 Binary Gas Analyser & XPM601 Paramagnetic Gas Analyser have been assessed and are considered capable for use in a low demand Safety Function up to (and including) SIL 2 capability with regards to systematic, random hardware failures and architectural constraints.

The assessment was based on the assumptions, data provided, and recommendations given in:

- Environmental Resources Management Ltd Report: H215_FM001 rev. 5.

The products were assessed against the following failure modes:

- XTP601: Ability to detect oxygen presence within another gas stream and generate a 4-20mA output;
- XTC601 & XPM601: Ability to detect target gas in another gas stream and generate a 4-20mA output.

The assessment was carried out to determine compliance with IEC 61508 (2010 Edition) with regards to:

- SIL 2 with a HFT = 0 via Route 1H;
- Architectural Constraint (Type B, SFF >90%, <99%), HFT = 0;
- Systematic Capability of SIL 2 capability against IEC 61508 (2010 Edition) via Route 2S.

Note 1: The SIL of a complete SIF (sensor, logic solver and final element subsystems) must be verified to calculate the required PFD / PFH, considering any redundancy, Proof Test Interval (PTI), Proof Test Coverage (PTC), Mission Time and Mean Time To Restoration (MTTR) for all elements included in the SIF. Each subsystem should be verified to ensure compliance with the minimum HFT requirements.

Device	λS (/hr)	λDD (/hr)	λDU (/hr)	SFF	Type	Estimated SIL Capability (Arch. Constraints)
XTP601	1.6E-07	7.4E-07	5.4E-08	94%	B	2
XTC601	1.6E-07	7.0E-07	3.9E-08	96%	B	2
XPM601	1.6E-07	6.8E-07	3.9E-08	96%	B	2

IMPORTANT: It should be noted that this assessment does not include confirmation of the response time of the devices. For response times (along with any relevant assumptions) reference should be made to the Safety Manual of each device and the total SIF response time MUST be compared against the process safety time for the specific application.



Partner: Simon Burwood

Assessment Date: February 2020

Renewal Date: September 2023, valid to September 2025

Certificate: H215_CT001 rev. 4

ENVIRONMENTAL RESOURCES MANAGEMENT LTD 2nd Floor, Exchequer Court, 33 St. Mary Axe, London, EC3A 8AA UK

Telephone: +44 (0)20 8542 2807

Registered in England and Wales: 7006868

Engineering Safety Consultants Limited. London, UK Test Report extract

General

This report provides a Prior Use Assessment of the Michell Instruments UK Ltd, XTP601 Process Oxygen Analyzer, XTC601 Binary Gas Analyzer and XPM601 Paramagnetic Oxygen Analyzer, as defined in the Prior Use requirements in IEC 61511 (2nd Edition) Clause 11.5.3 and 11.5.4 [2] including an estimation of Probability of Failure on Demand (PFD), Safe Failure Fraction (SFF) and a review of the systematic capability as supporting evidence for avoidance and minimisation of systematic failures.

A Failure Mode Effects and Diagnostics Analysis (FMEDA) was conducted on the XTP601, XTC601 and XPM601 to estimate the random hardware failure rate to assess suitability for use in a safety function with regards to the PFD and the architectural requirements in terms of Hardware Fault Tolerance (HFT) and SFF, using the approach detailed in Route 1H in IEC 61508-2 [1].

Hardware Reliability Verification

These devices will form part of the sensor element sub-system of a Safety Instrumented Function (SIF) and thus an assessment was conducted to demonstrate its capabilities in terms of PFD. The remaining sensing, logic solver and final element sub-systems were excluded from the assessment, to allow for their PFD contributions, the devices were assessed against 20% of Safety Integrity Level (SIL) 2 PFD band (e.g. SIL 2 band modified to 2.0E-03). The analysis was based on the assumption that repairs would be carried out with a Mean downtime (MDT) of 168 hours, a Proof Test Interval (PTI) of one year (8760 hours) and capable of revealing 100% of undetected failures.

The XTP601 Process Oxygen Analyzer and XPM601 Paramagnetic Oxygen Analyzer were assessed against the following safety function:

- Ability to detect oxygen presence within another gas stream and generate a 4...20 mA output.

The XTC601 Binary Gas Analyzer was assessed against the following safety function:


- Ability to detect target gas in another gas stream and generate a 4...20 mA output.

Table 3 shows a summary of the results of the XTP601, XTC601 and XPM601 based on the data provided and the assumptions given in this report. The full set of results for the hardware reliability verification is presented in Table 4.

Device	PFD Target (20% of SIL2 band)	PFD achieved	PFD achieved (SIL)	SFF	Type	Achieved SIL (Architecture HFT =0)	Overall achieved SIL
XTP601	2.0E-03	3.6E-04	2	94%	B	2	2
XTC601	2.0E-03	2.9E-04	2	96%	B	2	2
XPM601	2.0E-03	2.9E-04	2	96%	B	2	2

Device Reference		XTP601, XTC601 & XPM601
Function Specification		XTP601 Oxygen Transmitter XTC601 Binary Gas Analyzer XPM601 Paramagnetic Oxygen Analyzer
Software Configuration/Settings		As per customer order
Software Version		Firmware for XTP601: 36217 V1.09 Firmware for XTC601: 37701 V1.06 Firmware for XPM601: 36268 V1.01
Hardware Diagram version		XTP601: 80895/C V2.0 XTC601: 81003/C V1.0 XPM601: 83322 V1.0
Hardware Configuration/Settings		As per customer order
Failure Mode(s) Definition	Dangerous detected	dangerous detected failure rate per hour
	Dangerous undetected	dangerous undetected failure rate per hour
	Safe	safe (or spurious) failure rate per hour
Dangerous Undetected Failures (?DU)		XTP601 5.4E-08 , XTC601 3.9E-08 XPM601 3.9E-08 (<i>FIT/hr</i>)
Dangerous Detected Failures (?DD)		XTP601 7.4E-07 , XTC601 7.0E-07 XPM601 6.8E-07 (<i>FIT/hr</i>)
Safe Failures (?S)		XTP601, XTC601 & XPM601 1.6E-07 (<i>FIT/hr</i>)
Safe Failure Fraction (SFF)		XTP601 94% XTC601 96% XPM601 96%
Hardware Fault Tolerance (HFT)		0
Classification (Type A or Type B)		B
Demand (Low demand or High Demand)		Low
Proof Testing Procedures		See section 3.5
Installation		Refer to user manual 97313 (XTP), 97400 (XTC) & 97632 (XPM)
Average lifetime of device (yrs)		5
Environmental Profile		Operating Temperature: +5...+60 °C
Systematic/Proven in Use Safety Integrity Level		2
Assumptions		Refer to user manual
General Notes and applicable regulations		This product complies with applicable standards and clauses of EU ATEX, EMC, PED Directives. Refer to the EU Declaration supplied with each product for full details of the latest versions.
Testing requirements		See section 3.5

Documents / Resources

	<p>MICHELL Instruments XTP 601 SIL Oxygen Analyzer [pdf] User Manual XTP 601 SIL Oxygen Analyzer, XTP 601, SIL Oxygen Analyzer, Oxygen Analyzer, Analyzer</p>
-----------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

References

- [User Manual](#)

[Manuals+](#), [Privacy Policy](#)

This website is an independent publication and is neither affiliated with nor endorsed by any of the trademark owners. The "Bluetooth®" word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. The "Wi-Fi®" word mark and logos are registered trademarks owned by the Wi-Fi Alliance. Any use of these marks on this website does not imply any affiliation with or endorsement.