



MOKO SMART Commure Tag Wearable Bluetooth Device Owner's Manual

[Home](#) » [MOKO SMART](#) » MOKO SMART Commure Tag Wearable Bluetooth Device Owner's Manual 

Contents

- [1 MOKO SMART Commure Tag Wearable Bluetooth Device](#)
- [2 Product Information](#)
- [3 Product Usage Instructions](#)
- [4 Commure Tag](#)
- [5 Features](#)
- [6 Specifications](#)
- [7 FCC STATEMENT](#)
- [8 Documents / Resources](#)



MOKO SMART Commure Tag Wearable Bluetooth Device



Product Information

The Commure Tag is a battery-powered, wearable Bluetooth device designed for medical situations. It features BLE technology and positioning technology, allowing for immediate and effective response to emergency situations involving patients and staff. The device aims to create a safer workplace environment.

Features:

- Durability

Specifications

- Wireless BLE 5.0
- Power Consumption
- **Battery Type:** CR2450
- Dimensions & Material: 45.6*38*12.2mm ABS
- IP66 Waterproof
- FCC Certification

FCC Statement

The device has been evaluated and meets general RF exposure requirements. It can be used in portable exposure conditions without any restrictions or limitations.

Product Usage Instructions

To use the Commure Tag, please follow the steps below:

1. Ensure that the device is charged or the battery is properly inserted.
2. Attach the Commure Tag securely to your clothing or any suitable location using the provided attachment

mechanism.

3. To initiate a medical emergency signal, press and hold the single button on the Commure Tag for a few seconds until the device vibrates or indicates activation.
4. The signal will be sent via Bluetooth to the designated recipients or connected devices, enabling them to respond immediately and effectively.
5. In case of accidental activation or false alarm, press the button again to cancel the signal.

Note: Ensure that the Commure Tag is within the Bluetooth range of the recipient devices for proper signal transmission.

Commure Tag

The Commure Tag is a battery-powered, wearable Bluetooth device containing a single button used for medical situations. With BLE tech and positioning technology, the beacon helps ensure immediate and effective response to your patients and staff's emergency situations. So you can create a safer workplace environment for your patients and staff.

Features

- Raised, protective button edge to minimize accidental button presses
- Soft carbon- pill-based, ergonomic button designed for comfort and durability
- Easily mounted
- Operational self-check feature
- The piezoelectric speaker is used to provide audible feedback on alert trigger
- Flashing red light to indicate alert receipt by Cloud Services
- Minimal radio spectrum footprint

Specifications

- **Wireless** BLE 5.0
- **Power Consumption** < 15 μ A in normal mode < 200mA during alert mode
- **Operating temperature** -20° C to 60° C
- **Battery Type** CR2450 10 – 12 years
- **Battery Life** 45.6*38*12.2mm
- **Dimensions & Material** ABS
- **Waterproof** IP66
- **Certification** FCC

FCC STATEMENT

1. This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:
 - (1) This device may not cause harmful interference.
 - (2) This device must accept any interference received, including interference that may cause undesired operation.

2. Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.


NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation.

This equipment generates uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures: Reorient or relocate the receiving antenna. Increase the separation between the equipment and the receiver. Connect the equipment to an outlet on a circuit different from that to which the receiver is connected. Consult the dealer or an experienced radio/TV technician for help.

RF warning statement:

The device has been evaluated to meet general RF exposure requirements. The device can be used in portable exposure conditions without restriction.

Documents / Resources

	MOKO SMART Commure Tag Wearable Bluetooth Device [pdf] Owner's Manual 2AO94-COMMURETAG, 2AO94COMMURETAG, Commure Tag Wearable Bluetooth Device, W earable Bluetooth Device, Bluetooth Device
---	--