

NATIONAL INSTRUMENTS PXIe-6544 Digital Waveform Instrument User Guide

<u>Home</u> » <u>NATIONAL INSTRUMENTS</u> » NATIONAL INSTRUMENTS PXIe-6544 Digital Waveform Instrument User Guide [™]

NATIONAL INSTRUMENTS PXIe-6544 Digital Waveform Instrument



Contents

- 1 COMPREHENSIVE SERVICE
- **2 SELL YOUR SURPLUS**
- 3 OBSOLETE NI HARDWARE IN STOCK & READY TO

SHIP

- 4 Manufacturer
- **5 Board Assembly Part Numbers**
- **6 Volatile Memory**
- 7 Non-Volatile Memory (incl. Media Storage)
- **8 Procedures**
- **9 Terms and Definitions**
- **10 Customers Support**
- 11 Documents / Resources
 - 11.1 References
- **12 Related Posts**

COMPREHENSIVE SERVICE

We offer competitive repair and calibration service, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.



OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock New, New Surplus, Refurbished, and Reconditioned NI Hardware.

Bridging the gap between the manufacturer and your legacy test system.



Manufacturer

National Instruments

Board Assembly Part Numbers

(Refer to Procedure 1 for identification procedure):

Part Number and Revision	Description
196524A-0xL or later	PXIe-6544
196525A-0xL or later	PXIe-6545
199952A-0xL or later	PXIe-6547
199953A-0xL or later	PXIe-6548

Volatile Memory

Target Data	Туре	Size	Battery Backup	User Access ible	System Acc essible	Sanitization Procedure
Waveform data	Synchronous DRAM	512 MB	No	Yes	Yes	Cycle Power
PCI communication	MITE FIFO Re gister	208 bytes	No	Yes	Yes	Cycle Power
Buffering data	FPGA Block R AM	720 KB	No	Yes	Yes	Cycle Power
Memory control	FPGA Block R AM (x2)	112 KB	No	Yes	Yes	Cycle Power
Transfer data to/fr om daughter card	FPGA Block R AM	864 KB	No	Yes	Yes	Cycle Power

Non-Volatile Memory (incl. Media Storage)

Target Data	Type	Size	Battery Ba ckup	User Acces sible	System Acc essible	Sanitization Procedure
Board configuration	EEPRO M	16 KB	No	No	Yes	None
Device configuration Device information FPGA bitstream Calibration metadata Calibration data	Flash	256 MB	No	No No Yes No	Yes Yes Ye s Yes	None None Procedure 2 None
FPGA Configuration	CPLD	64 Macro Cells	No	No	No	None

- 1. Refer to Terms and Definitions section for clarification of User and System Accessible
- 2. Calibration constants that are stored on the device include information for the device's full operating range. Any implications resulting from partial self-calibration can be eliminated by running the full self-calibration procedure.

Procedures

Procedure 1 – Board Assembly Part Number Identification:

To determine the Board Assembly Part Number and Revision, refer to the label applied to the surface of your product. The Assembly Part Number should be formatted as "P/N: 19####A-0#L"

Procedure 2 – Device Information Flash (Calibration Metadata):

The user-accessible areas of the Device Configuration Flash are exposed through a calibration Applications Programming Interface (API) in LabVIEW. To clear the calibration meta-data area, complete the following steps:

- 1. Use function niHSDIO_ Init Ext Cal to create and initialize a special NI-HSDIO external calibration session. The Vi Session returned is an NI-HSDIO session that can be used during the calibration session.
- 2. Call function niHSDIO_ Change Ext Cal Password to clear the calibration password that is required to initialize an external calibration session.
- 3. Call function niHSDIO_ Cal Adjust Channel Voltage to clear the user-defined information of the selected channel(s).
- 4. Finally, call the function niHSDIO_ Close Ext Cal with the calibration action set to 'Commit', in order to store the new calibration constants and data in the onboard Flash and to close the calibration session.

Terms and Definitions

Cycle Power:

The process of completely removing power from the device and its components and allowing for adequate discharge. This process includes a complete shutdown of the PC and/or chassis containing the device; a reboot is not sufficient for the completion of this process.

Volatile Memory:

Requires power to maintain the stored information. When power is removed from this memory, its contents are lost.

This type of memory typically contains application specific data such as capture waveforms.

Non-Volatile Memory:

Power is not required to maintain the stored information. Device retains its contents when power is removed. This type of memory typically contains information necessary to boot, configure, or calibrate the product or may include device power up states.

User Accessible:

The component is read and/or write addressable such that a user can store arbitrary information to the component from the host using a publicly distributed NI tool, such as a Driver API, the System Configuration API, or MAX.

System Accessible:

The component is read and/or write addressable from the host without the need to physically alter the product.

Clearing:

Per NIST Special Publication 800-88 Revision 1, "clearing" is a logical technique to sanitize data in all User Accessible storage locations for protection against simple non-invasive data recovery techniques using the same interface available to the user; typically applied through the standard read and write commands to the storage device.

Sanitization:

Per NIST Special Publication 800-88 Revision 1, "sanitization" is a process to render access to "Target Data" on the media infeasible for a given level of effort. In this document, clearing is the degree of sanitization described.

This document is subject to change without notice For the most recent version, visit ni.com/manuals.

Customers Support

Contact: 866-275-6964 support@ni.com

Contact: 866-275-6964 support@ni.com



1-800-915-6216





All trademarks, brands, and names are the property of their respective owners.



Documents / Resources



NATIONAL INSTRUMENTS PXIe-6544 Digital Waveform Instrument [pdf] User Guide PXIe-6544, PXIe-6545, PXIe-6547, PXIe-6548, PXIe-6544 Digital Waveform Instrument, Digital Waveform Instrument, Waveform Instrument

References

- <u>Product Documentation NI</u>
- /W PXIe-6548 National Instruments PXI Digital Waveform Instrument | Apex Waves

Manuals+,