



# **senseca LPUVI02 UV Index Radiometer Instruction Manual**

[Home](#) » [senseca](#) » **senseca LPUVI02 UV Index Radiometer Instruction Manual** 

## **Contents**

- [1 senseca LPUVI02 UV Index Radiometer Instruction Manual](#)
- [2 Product Information](#)
- [3 INTRODUCTION](#)
- [4 WORKING PRINCIPLE](#)
- [5 INSTALLATION](#)
- [6 ELECTRICAL CONNECTIONS](#)
- [7 MEASUREMENT](#)
- [8 MAINTENANCE](#)
- [9 TECHNICAL SPECIFICATIONS](#)
- [10 SAFETY INSTRUCTIONS](#)
- [11 ACCESSORIES ORDERING CODES](#)
- [12 DECLARATION OF CONFORMITY](#)
- [13 WARRANTY](#)
- [14 TECHNICAL INFORMATION](#)
- [15 DISPOSAL INFORMATION](#)
- [16 Read More About This Manual & Download PDF:](#)
- [17 Documents / Resources](#)
  - [17.1 References](#)

## senseca LPUVI02 UV Index Radiometer Instruction Manual



### Product Information

<b>Product Name</b>	UVA Radiometer
<b>Model</b>	LPUVA02
<b>Manufacturer</b>	Delta OHM
<b>Website</b>	<a href="http://www.deltaohm.com">www.deltaohm.com</a>

### Working Principle

The LPUVA02 radiometer measures the global irradiance in the UVA spectral range on a flat surface ( $\text{W/m}^2$ ). It calculates the measured irradiance by summing up the direct solar irradiance and diffuse irradiance. The radiometer uses a solid state sensor with a special filter to achieve the desired spectral match. It has a 50 mm external diameter dome to protect the sensor from weather agents. The response of the radiometer follows the cosine law, ensuring accurate measurements even when the sun is at a low raising angle.

### Installation

To install the LPUVA02 radiometer, follow these steps:

1. Remove the perforated cap from the silica-gel cartridge.
2. Fill the cartridge with the silica-gel crystals from the sealed sachet.
3. Close the cartridge with the cap.
4. Attach the cartridge to the radiometer body using a coin.
5. Ensure the cartridge is tightly screwed to maintain the life of the silica gel.
6. Position the shade disk and secure it with the screws.
10. The radiometer is now ready for use.

## INTRODUCTION

The LPUVA02 radiometer measures the global irradiance in the UVA spectral range on a flat surface (W/m<sup>2</sup>). Measured irradiance is the result of the sum of direct solar irradiance and of diffuse irradiance. The radiometer can also be used for monitoring indoor UVA emissions. Available in the following versions:


- **LPUVA02:** PASSIVE
- **LPUVA02AC:** ACTIVE with 4..20 mA CURRENT output (0...200 W/m<sup>2</sup>)
- **LPUVA02AV:** ACTIVE with 0..1 or 0..5 or 0..10 V VOLTAGE output to be de-fined when ordering (0...200 W/m<sup>2</sup>).

Each radiometer is individually calibrated at the factory and is distinguished by its own calibration factor. The calibration is performed by comparison with the primary stand-ard of the Delta OHM metrology laboratory, using the emission line at 365 nm of a suitably filtered Xe-Hg lamp.

Note: there is no international standard for the calibration of UVA radiometers; there-fore, when considering the value of the calibration coefficient, the method by which it was obtained must be taken into account; the same radiometer calibrated with differ-ent procedures can have different calibration factors.

## WORKING PRINCIPLE


LPUVA02 radiometer is based on a solid state sensor, the spectral match with the de-sired curve is obtained using a special filter. The relative spectral response is reported on figure 2.1.



**Fig. 2.1**

The radiometer is provided with a 50 mm external diameter dome in order to supply a suitable protection of the sensor to the weather agents.

The response in accordance with the cosine law has been obtained thanks to the par-ticular shape of the diffuser and of the housing. The deviation between the theoretical response and the measured one is shown in the Fig. 2.2.



**Fig. 2.2**

The excellent relation between the response of the radiometer and the cosine law allows using the instrument also when the sun has a very low raising (the UVA diffuse radiation increases as the sun moves away from the zenith, therefore the error on the direct radiation, due to the imperfect response according to the cosine law, becomes negligible on the measurement of the global radiation).


## INSTALLATION

Before installing the radiometer, refill the cartridge containing silica-gel crystals. Silica gel absorbs humidity in the dome chamber and prevents, in particular climatic conditions, condensation on the internal walls of the domes and measurement alteration.

Do not touch the silica gel crystals with your hands while refilling the cartridge. Carry out the following instructions in an environment as drier as possible:


1. Loosen the three screws that fix the white shade disk.
2. Unscrew the silica gel cartridge using a coin.
3. Remove the cartridge perforated cap.
4. Open the sachet containing silica gel (supplied with the radiometer).
5. Fill the cartridge with the silica gel crystals.
6. Close the cartridge with its own cap, paying attention that the sealing O-ring be properly positioned.
7. Screw the cartridge to the radiometer body using a coin.
8. Check that the cartridge is screwed tightly (if not, silica gel life will be reduced).
9. Position the shade disk and screw it with the screws.
10. The radiometer is ready for use.

The figure below shows the operations necessary to fill the cartridge with the silica gel crystals.




**Fig. 3.1: filling the silica-gel cartridge**

- The radiometer must be mounted in an easy-to-reach location in order to clean the dome regularly and carry out maintenance. At the same time, make sure that no buildings, constructions, trees or obstructions exceed the horizontal plane where the radiometer lies. If this is not possible, select a site where obstructions in the path of the sun from sunrise to sunset do not exceed 5 degrees of elevation.
- The radiometer must be located far from any kind of obstruction, which might re-reflect sunlight (or sun shadow) onto the radiometer itself.
- For fixing, use the holes on the radiometer body (remove the shade disk to access the holes and reposition it after mounting) or the suitable accessories (see the figures below). In order to allow an accurate horizontal positioning, the radiometer is equipped with a levelling device: the adjustment is made by means of the two levelling screws that allow adjusting the radiometer inclination. The mast height does not exceed the radiometer plane to avoid measurement errors caused by any re-reflection or shadow of the mast itself.
- It is preferably to thermally insulate the radiometer from its mounting bracket ensuring, at the same time, a good electrical contact to ground.



**Fig. 3.2: fixing holes and levelling device**



**Fig. 3.3: fixing accessories**

## ELECTRICAL CONNECTIONS

LPUVA02... has a 4-pole connector and uses the CPM12AA4... optional cables.


The metallic housing of the radiometer should preferably be grounded (  $\frac{1}{\equiv}$  ) locally. In this case, do not connect the wire of the cable corresponding to the housing to prevent ground loops.

Only if it is not possible to ground locally the metallic case of the radiometer, connect the wire of the cable corresponding to the housing to ground. Note: in LPUVA02AV the housing is not connected to the connector.

## LPUVA02 CONNECTIONS

The radiometer LPUVA02 is passive and does not require power supply. It is to be connected either to a millivoltmeter or to a data acquisition system. Typically, the radiometer output signal does not exceed 50 mV. In order to better exploit the radiometer features, the readout instrument should have 1  $\mu$ V resolution.

Connector	Function	Color
1	+Vout	Red
2	-Vout	Blue
3	Housing (C)	White
4	Cable shield (SH)	Black




**Fig. 4.1: LPUVA02 connections**

## LPUVA02AC CONNECTIONS

The radiometer LPUVA02AC has 4...20 mA output and requires 10...30 Vdc external power supply. It is to be connected to a power supply and an instrument with 4...20 mA input as shown in fig. 4.2. The load resistance of the instrument reading the signal must be  $\leq 500 \Omega$ .

Connector	Function	Color
1	Positive (Iin)	Red
2	Negative (Iout)	Blue
3	Housing (C)	White
4	Cable shield (SH)	Black




**Fig. 4.2: LPUVA02AC connections**

## LPUVA02AV CONNECTIONS

The radiometer LPUVA02AV has 0...1 V, 0...5 V or 0...10 V output (depending on the ordered output) and requires external power supply: 10...30 Vdc for 0...1 V and 0...5 V outputs, 15...30 Vdc for 0...10 V output. It is to be connected to a power supply and an instrument with voltage input as shown in fig. 4.3. The load resistance of the instrument reading the signal must be  $\geq 100 \text{ k}\Omega$ .

Connector	Function	Color
1	Output positive (+Vout)	Red
2	Output negative Power supply negative (GND)	Blue
3	Power supply positive (+Vdc)	White
4	Cable shield (SH)	Black



**Fig. 4.3: LPUVA02AV connections**

## MEASUREMENT

Below are the ways to calculate the irradiance.

### LPUVA02

Each radiometer is distinguished by its own sensitivity (or calibration factor)  $S$  expressed in  $\mu\text{V}/(\text{Wm}^{-2})$  and shown in the label on the radiometer (and in the calibration report).

The irradiance  $E_e$  is obtained by measuring with a multimeter the difference of potential DDP at the ends of the sensor and applying the following formula:

$$E_e = \text{DDP} / S$$

#### where:

$E_e$  is the irradiance expressed in  $\text{W}/\text{m}^2$ ;

DDP is the difference of potential expressed in  $\mu\text{V}$  measured by the multimeter;  $S$  is the sensitivity of the radiometer expressed in  $\mu\text{V}/(\text{Wm}^{-2})$ .

### LPUVA02AC

The 4...20 mA output signal corresponds to 0...200  $\text{W}/\text{m}^2$  irradiance range.



The irradiance  $E_e$  is obtained by measuring with a multimeter the current  $I_{out}$  absorbed by the sensor and applying the following formula:

$$E_e = 12.5 \cdot (I_{out} - 4\text{mA})$$

**where:**

$E_e$  is the irradiance expressed in  $\text{W/m}^2$ ;

$I_{out}$  is the current expressed in mA absorbed by the radiometer.

## **LPUVA02AV**

The output signal (0...1 V, 0...5 V or 0...10 V depending on the version) corresponds to 0...200  $\text{W/m}^2$  irradiance range.

The irradiance  $E_e$  is obtained by measuring with a multimeter the output voltage  $V_{out}$  of the sensor and applying the following formula:

- $E_e = 200 \cdot V_{out}$  for the version 0...1 V
- $E_e = 40 \cdot V_{out}$  for the version 0...5 V
- $E_e = 20 \cdot V_{out}$  for the version 0...10 V

**where:**

$E_e$  is the irradiance expressed in  $\text{W/m}^2$ ;

$V_{out}$  is the output voltage expressed in V measured by the multimeter.

## **MAINTENANCE**

In order to grant measurements high accuracy, it is important to keep the glass dome clean. Consequently, the more the dome will be kept clean, the more measurements will be accurate.

You can wash it using water and standard papers for lens. If necessary, use pure ETHYL alcohol. After using alcohol, clean again the dome with water only.

Because of the high temperature changes between day and night, some condensation might appear on the radiometer dome. In this case the performed reading is highly over-estimated. To minimize the condensation, the radiometer is provided with a cartridge containing dessicant material (silica-gel). The efficiency of the silica-gel crystals decreases over time while absorbing humidity. Silica-gel crystals are efficient when their colour is yellow, while they turn white/translucent as soon as they lose their efficiency. Read instructions at chapter 3 about how to replace the silica-gel crystals. Silica-gel typical lifetime goes

To exploit all the radiometer features, it is highly recommended that the calibration be checked annually.

## **TECHNICAL SPECIFICATIONS**

<b>Measuring range</b>	0...200 W/m <sup>2</sup>
<b>Typical sensitivity</b>	70...200 $\mu\text{V/Wm}^{-2}$
<b>Viewing range</b>	2p sr
<b>Impedance</b>	3 k $\Omega$
<b>Spectral range</b>	342 nm...384 nm (1/2)330 nm...393 nm (1/10) 320 nm...400 nm (1/100) Peak: 365 nm
<b>Response time</b>	<0.5 s (95%)
<b>Output</b>	LPUVA02 = $\mu\text{V/Wm}^{-2}$ LPUVA02 <b>AC</b> = 4...20 mALPUVA02 <b>AV</b> = 0...1, 0...5, 0...10 V (depending on the model)
<b>Power supply</b>	10...30 Vdc (15...30 Vdc only for 0...10 V output)
<b>Operating temperature</b>	-40...+80 °C
<b>Response according to the cosine law</b>	< 8 % (0...80°)
<b>Long term instability (1 year)</b>	<  $\pm 3$   %
<b>Non linearity</b>	< $\frac{1}{2} \pm 1\frac{1}{2}$ %
<b>Temperature response</b>	< 0.1 %/°C
<b>Weight</b>	900 g approx.

## SAFETY INSTRUCTIONS

### General safety instructions

The instrument has been manufactured and tested in accordance with the safety standard EN61010-1:2010 "Safety requirements for electrical equipment for measurement, control and laboratory use" and has left the factory in perfect safety technical conditions.

The instrument proper operation and operating safety can be ensured only if all standard safety measures as well as the specific measures described in this manual are followed.

Do not use the instruments in places where there are:

- Corrosive or flammable gases.
- Direct vibrations or shocks to the instrument.
- High-intensity electromagnetic fields, static electricity.

### User obligations

The instrument operator shall follow the directives and regulations below that refer to the treatment of dangerous materials:

- EEC directives on workplace safety.
- National law regulations on workplace safety.

- Accident prevention regulations.

## ACCESSORIES ORDERING CODES

The radiometer is supplied with shade disk, silica-gel cartridge, 2 spare sachets, levelling device, M12 connector and Calibration Report.

### Accessories

- **LPSP1** UV-resistant shade disk (spare part).
- **LPS1** Only attachment bracket, suitable for mast with diameter 40...50 mm. Installation on horizontal or vertical mast.
- **LPRING02** Base with levelling device and adjustable holder for mounting the pyranometer in an inclined position (specify upon ordering on which radiometer model has to be mounted).
- **HD2003.79K** Kit to mount the radiometer on  $\varnothing$  40 mm clamping. To install the radiometer on a transverse mast.
- **LPS6** Kit for the installation of the radiometer. The kit includes: 750 mm mast, base fitting, graduated support plate, bracket for radiometer.
- **CPM12AA4...** 4-pole cable. 4-pole M12 connector on one end, open wires on the other end. Available length 2 m (CPM12AA4.2), 5 m (CPM12AA4.5) or 10 m (CPM12AA4.10).
- **LPSG** Cartridge to contain desiccant silica-gel crystals, complete with O-ring and cap (spare part).
- **LPG** Pack of 5 sachets of silica-gel crystals.

DELTA OHM metrology laboratories LAT N° 124 are ISO/IEC 17025 accredited by ACCREDIA for Temperature, Humidity, Pressure, Photometry / Radiometry, Acoustics and Air Velocity. They can supply calibration certificates for the accredited quantities.

## DECLARATION OF CONFORMITY

### WARRANTY

Delta OHM is required to respond to the “factory warranty” only in those cases provided by Legislative Decree 6 September 2005 – n. 206. Each instrument is sold after rigorous inspections; if any manufacturing defect is found, it is necessary to contact the distributor where the instrument was purchased from. During the warranty period (24 months from the date of invoice) any manufacturing defects found will be repaired free of charge. Misuse, wear, neglect, lack or inefficient maintenance as well as theft and damage during transport are excluded. Warranty does not apply if changes, tampering or unauthorized repairs are made on the product. Solutions, probes, electrodes and microphones are not guaranteed as the improper use, even for a few minutes, may cause irreparable damages.

Delta OHM repairs the products that show defects of construction in accordance with the terms and conditions of warranty included in the manual of the product. For any dispute, the competent court is the Court of Padua. The Italian law and the “Convention on Contracts for the International Sales of Goods” apply.

## TECHNICAL INFORMATION

The quality level of our instruments is the result of the continuous product development. This may lead to differences between the information reported in the manual and the instrument you have purchased. In case of discrepancies and/or inconsistencies, please write to [sales@deltaohm.com](mailto:sales@deltaohm.com). Delta OHM reserves the right to

change technical specifications and dimensions to fit the product requirements without prior notice.

## DISPOSAL INFORMATION

Electrical and electronic equipment marked with specific symbol in compliance with 2012/19/EU Directive must be disposed of separately from household waste. European users can hand them over to the dealer or to the manufacturer when purchasing a new electrical and electronic equipment, or to a WEEE collection point designated by local authorities. Illegal disposal is punished by law.

Disposing of electrical and electronic equipment separately from normal waste helps to preserve natural resources and allows materials to be recycled in an environmentally friendly way without risks to human health.

GHM GROUP – Delta OHM | Delta Ohm S.r.l. a socio unico  
Via Marconi 5 | 35030 Caselle di Selvazzano | Padova | ITALY

**Phone** +39 049 8977150


**Fax** +39 049 635596

[www.deltaohm.com](http://www.deltaohm.com)

[sales@deltaohm.com](mailto:sales@deltaohm.com)

***Read More About This Manual & Download PDF:***

## Documents / Resources

	<p><a href="#">senseca LPUVI02 UV Index Radiometer</a> [pdf] Instruction Manual LPUVI02AV, LPUVI02AV1, LPUVI02AV5, LPUVI02.1AV, LPUVI02.1AV1, LPUVI02.1AV5, LPUVI02 UV Index Radiometer, Index Radiometer, Radiometer</p>
---	---

## References

-  [Mess- und Regeltechnik | Senseca Germany GmbH](#)
- [User Manual](#)

### Manuals+ Privacy Policy

This website is an independent publication and is neither affiliated with nor endorsed by any of the trademark owners. The "Bluetooth®" word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. The "Wi-Fi®" word mark and logos are registered trademarks owned by the Wi-Fi Alliance. Any use of these marks on this website does not imply any affiliation with or endorsement.