



Solinst 3250 LevelVent 5 Logger User Manual

[Home](#) » [Solinst](#) » Solinst 3250 LevelVent 5 Logger User Manual 



More Info | Instructions | Get Quote
LevelVent: Hardware Compatibility
Recommended Firmware & Hardware Configurations

Contents

- [1 Software & Firmware](#)
- [2 Operating System Requirements](#)
- [3 LevelVent Accessories](#)
- [4 Documents / Resources](#)
 - [4.1 References](#)
- [5 Related Posts](#)

Software & Firmware

Solinst always recommends using the latest software and firmware release.
Current releases can be downloaded at any time, free of charge, from the Solinst web site at:
<https://downloads.solinst.com/>







Always use the latest Levellogger® PC Software version to install the latest Firmware versions.

Note: Solinst always recommends that our customers download the data from their LevelVent loggers, using the same Levellogger Software version they were originally programmed with.

Operating System Requirements

Levellogger PC Software Versions 4.6.3 and up are compatible with Windows 10 & 11.



Level Vent Model 3250	Levellogger PC Software	LevelVent Firmware	Solinst Levellogger App	SRU Firmware	DataGrabber Firmware
 LevelVent 5 Logger	4.6.3 (Includes Firmware Upgrade Utility 1.4.1)	1.006	3.5.1 iOS 13.0, or later  2.1.0 Android 9.0, or later 	1.001 and up 	DataGrabber 5: **Mk2 4.000 and up Mk1 3.002 
 LevelVent Logger		1.000			DataGrabber 5: Mk2: 4.00 and up Mk1: 3.002 DataGrabber: 2.000 to 3.002 (firmware version 2.000 and up is user-upgradable, below that, must be upgraded by Solinst)

**Mk2 DataGrabber 5 units sold starting October 2022; Mk1 units sold before then can only be upgraded to firmware 3.002.

LevelVent Accessories

Older style Vented Cables (sold before 4/16/21 with serial numbers lower than 506326) are not compatible with newer style LevelVent 5 Wellheads (sold after 4/14/21 with serial numbers 495959 and up) and Loggers (sold after 4/5/21) without a small user modification – contact Solinst for instructions. Newer style Vented Cables are compatible with all versions of the LevelVent Wellheads and Loggers – please note, after the new Vented Cable connector is fully tightened, the external o-ring on an old Logger or Wellhead is visible and it is normal not to seat against the connector. USB PC Interface Cables are compatible with all LevelVent versions.



LevelVent 5 Wellhead



Vented Cable



USB PC Interface Cable



* Levellogger 5
App Interface

*“Made for iPod,” “Made for iPhone,” and “Made for iPad” mean that an electronic accessory has been designed to connect specifically to iPod, iPhone, or iPad, respectively, and has been certified by the developer to meet Apple performance standards. Apple is not responsible for the operation of this device or its compliance with safety and regulatory standards. Please note that the use of this accessory with iPod, iPhone, or iPad may affect wireless performance.

® Apple, the Apple logo, iPhone, iPad, and iPod are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is a service mark of Apple Inc. iOS is a trademark or registered trademark of Cisco in the U.S. and other countries and is used under license.

Android and Google Play are trademarks of Google Inc.

®Solinst and Levellogger are registered trademarks of Solinst Canada Ltd.

Solinst Canada Ltd. 35 Todd Road, Georgetown, Ontario Canada L7G 4R8 www.solinst.com
E-mail: instruments@solinst.com Tel: +1 (905) 873-2255; (800) 661-2023 Fax: +1 (905) 873-1992
August 10, 2023

Solinst®

Documents / Resources



[Solinst 3250 LevelVent 5 Logger \[pdf\] User Manual](#)
3250 LevelVent 5 Logger, 3250, LevelVent 5 Logger, 5 Logger

References

- [Solinst Member Modal For CF Members](#)
- [Solinst on the App Store](#)
- [Solinst Readout Unit: SRU: Instant Access to Water Level Readings](#)
- [Solinst Groundwater and Surface Water Monitoring Instrumentation, Water Level Meters, Interface Meters, Dataloggers](#)
- [Solinst DataGrabber 5: Levellogger Data Transfer Device](#)
- [Solinst Levellogger 5 App Interface](#)
- [Solinst LevelVent 5 Vented Water Level Datalogger Quote Request Form](#)
- [Solinst LevelVent Instructions](#)
- [Solinst LevelVent 5: Vented Water Level Dataloggers](#)

-  [Solinst LevelVent 5: Vented Water Level Dataloggers](#)

Manuals+.