



## UNI-T UT334A Radiation Dose Tester User Manual

[Home](#) » [UNI-T](#) » UNI-T UT334A Radiation Dose Tester User Manual 

# UNI-T®

UT334A  
Radiation Dose Tester  
User Manual



Thank you for purchasing the new UT334A radiation dose tester. In order to use this product safely and correctly, please read this manual thoroughly, especially the Safety Instructions part.

After reading this manual, it is recommended to keep the manual in an easily accessible place, preferably close to the device, for future reference.

### **Limited Warranty and Liability**

Uni-Trend guarantees that the product is free from any defect in material and workmanship within one year from the purchase date. This warranty does not apply to damages caused by accident, negligence, misuse, modification, contamination, and improper handling. The dealer shall not be entitled to give any other warranty on behalf of Uni-Trend. If you need warranty service within the warranty period, please contact your seller directly.

This warranty is the only compensation you can obtain. Uni-Trend will not be responsible for any special, indirect, incidental, or subsequent damage or loss caused by any reason or speculation. As some areas or countries do not allow limitations on implied warranties and incidental or subsequent damage, the above limitation of liability and stipulation may not apply to you.

#### **Contents**

- [1 Overview](#)
- [2 Applications](#)
- [3 Total Dose Limit](#)
- [4 Features](#)
- [5 Specifications](#)
- [6 Buttons](#)
- [7 Operation](#)
- [8 Cautions](#)
- [9 Safety Instructions](#)
- [10 Packing List](#)
- [11 Documents / Resources](#)
- [12 Related Posts](#)

### **Overview**

UT334A is a compact radiation dose tester used to monitor X, f3, and y rays. The product has the characteristics of high sensitivity and accurate measurement and is equipped with a high-performance microprocessor and TFT 2.0 color screen, which makes the operation convenient and the anti-interference ability strong. It has audio and visual alarms. The alarm thresholds can be set. The product will raise alarms when the thresholds are reached. The main specifications conform to national and international standards.

### **Applications**

UT334A is widely used in home decoration, radiation processing enterprise, health and epidemic prevention, radiation therapy, nuclear laboratory, nuclear power plant, import and export inspection, building materials, petrochemical, geological survey, scrap steel, industrial nondestructive testing, and other environments existing ionizing radiation, personal radiation dose supervision and protection occasions.

### **Total Dose Limit**

**Professional staff:**

The annual average effective dose for 5 consecutive years	200mSv
The effective dose in any year	50mSv
Annual equivalent dose of the lens of the eye	150mSv
Annual equivalent dose for extremities (hands and feet) or skin	500mSv

#### Ordinary people:

Annual effective dose	1mSv
The effective dose for a single year if the average annual dose for 5 years does not exceed 1mSv	5mSv
Annual equivalent dose of the lens of the eye	15mSv
Annual equivalent dose for extremities (hands and feet) or skin	50mSv

#### Common sense of radioactivity:

International Standard-1990

Professional staff: 20mSv/year (10uSv/hour)

Ordinary people: 1mSv/year (0.52uSv/hour)

#### Features


- Measure dose rates in real-time and record total doses
- Calibration function
- Adjustable alarm thresholds
- The start time of current dose accumulation can be viewed.
- Scheduled measurement (00-999h)
- Over-limit alarm
- Save the records of the last ten measurements
- The clock runs normally after the product is turned off.
- Monitor battery status
- English/Chinese optional
- Sensor fault warning ( "Device damaged" will show on the display.)
- In the sleep mode, if the radiation dose exceeds the set alarm threshold, the product will automatically alarm and enter the measurement interface.


#### Specifications

Sensor	G-M counter tube
Range	0.05uSv-50mSy
Accuracy	-17%~+25% based on <sup>137</sup> CSy
Charging interface	Type-C interface (The green light is on when charging, and off when fully charged.)
Alarm response time	td, < 10s
Alarm threshold accuracy	-17%~+25%
Alarm modes	Audio and visual alarms
Detected rays	X, β, and γ rays
Battery	3.7V 1800mAh
Battery life	20 days (continuous measurement, particle sound off)
Product weight	127g
Product size	120*65*25mm

## Buttons

 : Power/select

 : Step through menus to the left/subtract value


 : Step through menus to the right/add value


## Operation

### 1. Power on

Long press  3s to turn on the product and enter the main interface.

### 2. Power off


Long press  for 3s to turn off the product.

**Note:** When  shown on the screen, the product cannot be turned off. If users need to turn it off at this time, enter the “Schedule” interface to stop timing.

### 3. Measure


Enter the “Measure” interface to view real-time dose rate (0.05uSv-50mSv), average dose rate (0.05uSv-50mSv) and total dose (0.1uSv-9999mSv).

### 4. Schedule

Enter the “Schedule” interface to set the start delay time (00-99h) and the measurement time (00-999h). Select “Start timing” and press 

5. and the product will start the countdown. When the start delay countdown returns to zero, the real-time dose rate, average dose rate, and total dose will also return to zero and the product will start measuring the radiation dose and total dose rate. When the measurement time countdown returns to zero, the product will automatically save the interval measurement data. Users can find these data in the “Records” interface.

**Note:** Start delay is the countdown to start the measurement, and the hour, minute, and second can be set.


Select “Start timing” and press  to start the countdown. When the countdown ends, the measurement begins.







## 6. Records

After each scheduled measurement, the product will automatically save one history record. A maximum of 10 records can be saved. After 10 records are saved, the new record automatically replaces the last one if users continue to store the new record.

## 7. Delete

Enter the “Delete” interface to delete current data or historical data. The current data is the average dose rate and total dose. The history data is all data in historical records.

**Note:** Historical records cannot be restored after being deleted. When  shown on the screen, the product cannot enter the “Delete” interface.

Select “Calibration factor” and press  twice, and “Password: 0000” will appear at the bottom of the screen. Press  to adjust “0000” to “0018”, and press  to confirm. Press  or  to adjust the calibration factor, and press  to save. It is used to calibrate the error in the range of 20uSv-1uSv.

**Note:** Cf=1.00 by default

Calibration factor Cf= Personal dose equivalent Hp (10) reference value/Device indicating value

Technical specification for calibration (code, name): JJG 1009-2016 Personal Dose Equivalent Hp (10) Monitors for X and y Radiations




## 8. Alarm

Dose rate: 0.1uSv-999uSv

Total dose: 1uSv-9999uSv

The product will raise alarms when the thresholds are reached.

## 9. Unit

In the unit setting interface, select “unit”, press , and then press  or  switch between uSv and uGy. The language can be English or Chinese.

## 10. Sound

In this interface, the alarm volume and alarm times can be set. Adjust the alarm volume to 0, and the sound will be turned off. If the particle sound is turned on, when ions enter the sensor, the product will make a particle sound.



## 11. Clock

In this interface, the date (year/month/day) and time (hour/minute/second) can be set.

**Note:** It is recommended to set the clock to the local date and time before using the product.

## 12. Sleep

Screen sleep: The unit of screen sleep time is a minute. When it is adjusted to , the screen will not sleep.

Auto shutdown: The unit of auto shutdown time is an hour. When it is adjusted to , the auto shutdown function is turned off. Users only can long press  to power off at this point.

Screen brightness: Adjust the screen brightness.

## Cautions

1. Please turn off the product after use. If not used for a long time, please fully charge the product and turn it off. Batteries need to be recharged every 12 months.

2. When the battery status is low, charge the battery in time to ensure measurement accuracy.
3. Turning on particle sounds reduces battery life.
4. The manual version is subject to update without prior notice.

## Safety Instructions

Accidental fall of product	Please confirm whether the measured values are normal and whether they will be updated. If any abnormality is found, do not use the product in places with strong radiation.
"Device damaged" shows on the display.	Please do not use the product in places with strong radiation.
The operating environment contains explosive and flammable gas or dust.	Do not use the product in this environment.
Waterproof	The waterproof grade of the product is IP40, and cannot get wet or splashed.
Energy response	The X-ray can only be used as a reference to test radiation intensity and cannot meet the energy response requirements.

## Packing List

Radiation dose tester	1PC
User manual	1PC
Carrying bag	1PC
Data cable	1PC
Hand strap	1PC



**UNI-TREND TECHNOLOGY (CHINA) CO., LTD.**

No.6, Gong Ye Bei 1st Road,  
Songshan Lake National High-Tech Industrial  
Development Zone, Dongguan City,  
Guangdong Province, China

Tel: (86-769) 8572 3888

[www.uni-trend.com](http://www.uni-trend.com)

P/N:110401111223X

2022/06/29 REV.0

## Documents / Resources



[UNI-T UT334A Radiation Dose Tester](#) [pdf] User Manual  
UT334A, Radiation Dose Tester, UT334A Radiation Dose Tester